



Overview of the IC2017n Results

Marie-Anne Chevallier¹, Elena Fantuzzi², Michael Hajek³, Marlies Luszik-Bhadra⁴,
Sabine Mayer⁵, Rick Tanner⁶, David Thomas⁷ & Filip Vanhavere⁸

¹Institut de Radioprotection et de Sûreté Nucléaire, Fontenay-aux-Roses, France

²Italian National Agency for New Technologies, Energy and Sustainable Economic Development, Bologna, Italy

³International Atomic Energy Agency, Vienna, Austria

⁴Physikalisch-Technische Bundesanstalt, Braunschweig, Germany

⁵Paul Scherrer Institut, Villigen, Switzerland

⁶Public Health England,

⁷National Physical Laboratory,

⁸SCK•CEN Belgian Nuclear Research Centre, Mol, Belgium



Outline


- Reminder of radiation qualities and spectra
- Categories of dosimeters
- Overview of participants' results
- Results for specific radiation qualities

Reminder — Radiation Qualities




No.	Radiation quality	$H_p(10)$ (mSv)		
1	Bare ^{252}Cf source at 0°	0.3	1.5	12
2	Bare ^{252}Cf & ^{137}Cs sources at 0° [$H_p(10)$ photons = 1 mSv]		1.5	
3	Bare ^{252}Cf source at 45°		1.5	
4	D_2O -moderated ^{252}Cf source at 0°		1.2	
5	D_2O -moderated ^{252}Cf source behind shadow block		1.0	
6	Bare $^{241}\text{Am-Be}$ at 0°		1.5	

Reminder — Fluence Spectra



Reminder — Personal Dose Equivalent Spectra



Categories of Dosemeters

33 dosemeter systems from 32 individual monitoring services

18 track systems

- 7 etched track detectors for fast neutrons with thermal neutron TLD
- 7 etched track detectors for fast neutrons with thermal neutron converters
- 3 etched track detectors for fast neutrons without evidence of thermal sensor
- 1 fission track detector

15 albedo systems

- 10 TLD with boron-loaded shield
- 3 TLD with cadmium shield
- 1 OSLD
- 1 TLD lacking information on shielding against direct thermal neutrons

To ease identification OG reassigned dosemeters registered under types “other” or “combination”

Partial Repeat of Irradiations

- Possibility of unexpected photon exposure for some of the participants' dosimeters could not be excluded
- OG requested submission of photon doses evaluated for all dosimeters without correction/subtraction due to issuing period or transportation
- Photon doses reported were not included in certificates but helped OG resolve the issues
- As results might have been impacted, three systems were offered a repeat for part of the irradiations at no additional cost

Dosemeter Response




$$R = \frac{H_m}{H_{ref}}$$

Irradiated dosimeters	924
Reported values	924
	R
Arithmetic mean	1.18
Median	1.02
Standard deviation	1.23
2.5 th -percentile	0.13
97.5 th -percentile	4.52

H_m ... measured $H_p(10)$ as provided by IMS

H_{ref} ... reference $H_p(10)$ as determined by irradiating laboratory

Distribution of Response




Mean and Standard Deviation of Response




Radiation quality	$H_p(10)$ (mSv)	All		Albedo		Track	
		Mean	σ	Mean	σ	Mean	σ
Bare ^{252}Cf at 0°	0.3	1.47	2.52	1.06	0.78	1.82	3.31
	1.5	1.11	0.63	0.99	0.40	1.21	0.75
	12	1.05	0.46	0.95	0.42	1.13	0.48
Bare ^{252}Cf & ^{137}Cs at 0°	1.5	1.07	0.45	0.94	0.40	1.17	0.47
Bare ^{252}Cf at 45°	1.5	0.86	0.59	0.88	0.34	0.85	0.74
D ₂ O-mod. ^{252}Cf at 0°	1.2	1.08	1.94	1.04	0.23	1.11	0.50
D ₂ O-mod. ^{252}Cf behind SB	1.0	2.19	1.94	3.22	1.96	1.33	1.47
Bare $^{241}\text{Am-Be}$ at 0°	1.5	0.94	0.78	0.70	0.32	1.14	0.97


Summary of Reported Responses




Bare ^{252}Cf Source at 0° – Summary




Bare ^{252}Cf Source at 0° — 0.3 mSv




Bare ^{252}Cf Source at 0° — 1.5 mSv




Bare ^{252}Cf Source at 0° — 12 mSv




Bare ^{252}Cf & ^{137}Cs Sources at 0° — 1.5 mSv (neutrons)




Bare ^{252}Cf Source at 45° — 1.5 mSv




D₂O-moderated ²⁵²Cf Source at 0° — 1.2 mSv



D₂O-mod. ²⁵²Cf Source, Shadow Block — 1 mSv



Bare $^{241}\text{Am-Be}$ Source at 0° — 1.5 mSv



Conclusions

- Applying approval criterion and performance limits of ISO 14146:2018, **9 (out of 15) albedo** and **12 (out of 18) track systems** passed with not more than two outliers
- Overresponse of albedo systems for D₂O-moderated ²⁵²Cf source behind shadow block due to nearly isotropic distribution and very soft field
 - *Some albedo systems responded within performance limits because of improved side shielding or correction based on ratio of readings behind front and albedo window*
- Track detectors tend to underestimate low-energy neutrons at high angles of incidence



Thank you!

Please let us know your suggestions or claims
by e-mail to coordinator@ic2017n.org