

EURADOS Intercomparison 2022 for Neutron Dosimeters Overview of Results

M-A Chevallier¹, E Fantuzzi², M Hajek³ & S Mayer⁴

¹ *Institut de Radioprotection et de Sûreté Nucléaire, Fontenay-aux-Roses, France*

² *Italian National Agency for New Technologies, Energy and Sustainable Economic Development, Bologna, Italy*

³ *International Atomic Energy Agency, Vienna, Austria*

⁴ *Paul Scherrer Institute, Villigen, Switzerland*

Outline



- Reminder of radiation qualities
- Categories of dosimeters
- Overview of participants' results
- Results for specific radiation qualities

Reminder — Radiation Qualities



No.	Radiation quality	$H_p(10)$ [mSv]	
1	Bare ^{252}Cf source at 0°	0.3	5.0
2	Bare ^{252}Cf source at 30°	0.5	
3	Bare ^{252}Cf source at 45°	0.5	
4	D_2O -moderated ^{252}Cf source at 0° & 1 mm Cd	0.8	
5	Bare ^{252}Cf source (0.45 mSv) & thermal neutron field (0.15 mSv)	0.6	
6	Bare $^{241}\text{Am-Be}$ at 0°	1.0	
7	Bare $^{241}\text{Am-Be}$ at 30°	0.5	

Categories of Dosemeters



31 dosimetry systems from 29 individual monitoring services in 19 countries

20 track systems

- 10 etched track detectors for fast neutrons with thermal neutron converters
- 8 etched track detectors for fast neutrons with thermal neutron sensor
- 2 etched track detectors for fast neutrons without evidence of thermal sensor

11 albedo systems

- 5 thermoluminescence detectors with boron-loaded shield
- 4 thermoluminescence detectors with cadmium shield
- 1 thermoluminescence detector lacking information on shielding
- 1 active personal dosemeter


To ease identification OG reassigned dosemeters registered under types “other” or “combination”

Overall Statistics




Irradiated dosimeters	744
Reported values	744
$R = G/H_{ref}$	
Arithmetic mean	1.34
Median	0.94
Standard deviation	3.04
2.5 th -percentile	0.00
97.5 th -percentile	8.26


Bare ^{252}Cf Source at 0° — 0.3 mSv




Bare ^{252}Cf Source at 0° — 0.3 mSv




Bare ^{252}Cf Source at 0° — 5.0 mSv




Bare ^{252}Cf Source at 0° – Summary




Bare ^{252}Cf Source at 30° – 0.5 mSv




Bare ^{252}Cf Source at 45° — 0.5 mSv




D₂O-moderated ²⁵²Cf Source at 0° — 0.8 mSv




Bare ^{252}Cf Source & Thermal Field – 0.6 mSv



Bare ^{241}Am -Be Source at 0° — 1.0 mSv



Bare ^{241}Am -Be Source at 30° – 0.5 mSv




Mean and Standard Deviation of Response




Radiation quality	$H_p(10)$ [mSv]	All		Albedo		Track	
		\bar{R}	σ	\bar{R}	σ	\bar{R}	σ
Bare ^{252}Cf at 0°	0.3	1.14	0.57	0.86	0.52	1.29	0.55
	5.0	0.92	0.29	0.85	0.37	0.95	0.23
Bare ^{252}Cf at 30°	0.5	0.79	0.33	0.75	0.39	0.82	0.29
Bare ^{252}Cf at 45°	0.5	0.66	0.28	0.72	0.35	0.63	0.24
D ₂ O-mod. ^{252}Cf at 0°	0.8	1.96	2.68	3.68	3.96	1.01	0.34
Bare ^{252}Cf & thermal field	0.6	4.06	8.79	8.18	13.86	1.79	1.56
Bare $^{241}\text{Am-Be}$ at 0°	1.0	0.79	0.30	0.72	0.37	0.82	0.25
Bare $^{241}\text{Am-Be}$ at 30°	0.5	0.75	0.32	0.72	0.36	0.76	0.30


Distribution of Response



Summary of Reported Responses



Test for False Positive Response – Unirradiated



Reporting Levels




20 track systems


- 0.1 mSv (11 IMSs), 0.15 mSv (1), 0.2 mSv (4), 0.3 mSv (2) and 0.5 mSv (2)

11 albedo systems


- 0.1 mSv (4 IMSs), 0.2 mSv (1), 0.3 mSv (1) and DL (1)



Test for False Positive Response – Unirradiated



Test for False Positive Response – Unirradiated



Conclusions



- Applying approval criterion and performance limits of ISO 14146:2018, **5 (out of 11) albedo** and **16 (out of 20) track systems** passed with not more than two outliers
- Significant overresponse of albedo systems to moderated and thermal neutrons due to very soft field and/or nearly isotropic distribution
 - *Some albedo systems responded within performance limits because of improved side shielding or correction based on ratio of readings behind front and albedo window*
- Track systems generally provided better estimate of neutron personal dose equivalent at lower uncertainty
 - *Track detectors tend to slightly underestimate neutron personal dose equivalent at higher angles of incidence*

Thank you!

Please let **us know your** suggestions or claims
by e-mail to coordinator@ic2022n.org