

**EURADOS Intercomparison 2019
for Extremity and Eye Lens Dosemeters -
Summary of Procedures, Results and Conclusions**

A. F. McWhan, W. Dobrzynska, H. Stadtmann,
T. W. M. Grimbergen, M. Figel, A. M. Romero, I. Clairand

ISSN 2226-8057
ISBN 978-3-943701-29-6
DOI: 10.12768/v11r-w898

EURADOS Report 2021-05
Neuherberg, September 2021

EURADOS Intercomparison 2019

for Extremity and Eye Lens Dosemeters - Summary of Procedures, Results and Conclusions

A. F. McWhan¹, W. Dobrzynska¹, H. Stadtmann²,
T. W. M. Grimbergen³, M. Figel⁴, A. M. Romero⁵,
I. Clairand⁶

¹ Cavendish Nuclear Limited, United Kingdom (Babcock International Group)

² Seibersdorf Labor GmbH, Austria

³ Mirion Technologies (Dosimetry Services) B.V., the Netherlands

⁴ Mirion Technologies (AWST) GmbH, Germany

⁵ CIEMAT, Spain

⁶ IRSN, Institut de Radioprotection et de Sûreté Nucléaire, France

ISSN 2226-8057

ISBN 978-3-943701-29-6

DOI: 10.12768/v11r-w898

Imprint

© EURADOS 2021

Issued by:

European Radiation Dosimetry Group e.V.
Postfach 1129
85758 Neuherberg
Germany
office@eurados.org
www.eurados.org

The European Radiation Dosimetry Group e.V. is a non-profit organisation promoting research and development and European cooperation in the field of the dosimetry of ionising radiation. It is registered in the Register of Associations (Amtsgericht München, registry number VR 207982) and certified to be of non-profit character (Finanzamt München, notification from 2021-04-23).

Liability Disclaimer

No liability will be undertaken for completeness, editorial or technical mistakes, omissions as well as for correctness of the contents.

.

Contents

Contents.....	i
Abstract	iii
1 Introduction.....	1
2 Outline of the EURADOS IC 2019_{exteye}.....	3
2.1 Organization Group	3
2.2 Scope.....	3
2.3 Coordination	3
2.4 Set-up.....	4
2.5 Time schedule.....	4
2.6 Irradiation plan	5
2.7 Participants and systems	6
2.8 Intercomparison procedure compared to ISO 14146	12
2.9 Background and transit dose control.....	14
2.10 Confidentiality of the data and the results	15
2.11 Certificates of participation and participants meeting.....	16
3 Results and Discussion	17
3.1 Review of the comments received from participants	17
3.2 Basic statistical results	17
3.3 Response values per radiation quality	19
3.4 Linearity.....	21
3.5 Response values as a function of reference doses.....	22
3.6 Outliers	24
3.7 Results for individual systems	24
4 Conclusions.....	27
References.....	29
Figures and tables	31
Appendix A: Previous whole body and extremity intercomparisons organised by EURADOS WG2.	33
Appendix B: List of participants	35
Appendix C: Example Irradiation Certificate - photon	39
Appendix D: Example Irradiation Certificate - beta.....	41
Appendix E: Example “Certificate of Participation”	45
Appendix F: Datasheets with results for individual participants	49

Abstract

EURADOS Working Group 2 has developed a system for a self-sustained programme of regular dosimeter intercomparisons (ICs), (Figel, M. Report to Council, WG02-SG2, 2007). ICs for whole body dosimeters were carried out in 2008, 2010, 2012, 2014, 2016 and 2018. In addition, four ICs with different scopes were carried out, two for extremity dosimeters in photon and beta fields (IC2009_{ext} and IC2015_{ext}) and two for whole-body dosimeters in neutron fields (IC2012_n and IC2017_n). This IC, ie. IC2019_{exteye}, for extremity and eye lens dosimeters (effectively two separate ICs), has built on the success of these previous ICs with a total of 113 participating systems (68 extremity and 45 eye lens) from 60 institutes with participants from 26 countries around the world.

The high number of participants with eye lens dosimeters demonstrates the recent increase in demand for these services which has arisen from the implementation in 2018 by EU Member States of the 20 mSv eye lens dose limit introduced by the European Council (European Council Directive 2013/59/EURATOM, 2014).

The systems tested during this exercise included 97 TLD, 12 OSL and 4 dosimeter systems based on other techniques (Other). A total of 3616 dosimeters were handled by the coordinator of which 2314 dosimeters were irradiated. Photon irradiations were carried out by GAEC and the beta irradiations were carried out by IRSN.

Out of the total of 113 systems, 57 reported results for photons only and 56 reported both photon and beta. In general, the participants showed a very satisfactory performance with the medians of all $H_p(0.07)$ and $H_p(3)$ response values very close to unity.

From the statistical overviews of the results, the reports provided by the irradiation laboratories and the remarks received from the participants, it can be concluded that the intercomparison exercise was successful and that there were no significant issues encountered by the intercomparison organization or the irradiation laboratories during the execution of the exercise.

The Participants Meeting had been scheduled to be held at IM2020 (23 April) in Budapest but unfortunately both of these events had to be cancelled as a result of Covid-19.

These IC results can assist the participants to show compliance with their quality management system, compare their results with those from other participants and develop action plans for improvement of their systems. The high number of participants confirms that there is strong demand for international IC exercises in Europe, and that these are of significant operational value for Individual Monitoring Services (IMS).

1 Introduction

EURADOS working groups on Harmonisation of Individual Monitoring in Europe (1997-2000 (Bartlett, et al., 2000) (Bartlett, et al., 2001), 2001-2004 (Van Dijk, et al., 2004)) have shown that intercomparison exercises (ICs) are fundamental for harmonisation of individual monitoring services (IMS). Consequently, these EURADOS working groups recommended periodic performance tests or IC exercises within the European Union (EU) to assist this objective. It was believed that ICs would stimulate IMS to improve the quality of their results, provide information on IMS quality throughout EU and assist harmonisation of IMS quality control standards. Further support was provided by the response to questionnaires sent to IMS in the EU and non-EU countries which showed very strong interest in participating in the proposed programme of periodic ICs.

Participation in regular ICs is now specifically recommended in the European Commission's *Technical Recommendations for Monitoring Individuals Occupationally Exposed to External Radiation* (European Commission, 2009). Participation is also being considered as an essential criterion for IMS approval by a number of national authorities. At the same time, a growing number of IMS are either working towards, or have already achieved, formal EN ISO/IEC 17025 (ISO 17025, 2017) accreditation which specifically requires participation in regular inter laboratory comparisons.

EURADOS WG2 has now successfully carried out five ICs for whole body dosimeters (IC2008, IC2010, IC2012, IC2014, IC2018) (Grimbergen, et al., 2012) (McWhan, et al., 2015) (McWhan, et al., 2015) (Stadtman, et al., 2018) and three for extremity dosimeters in photon and beta fields (IC2009, IC2015_{ext}, IC2019_{exteye}) – see (Grimbergen, et al., 2013). All of these ICs have been performed without any external funding with all costs being covered by the participants' fees. Details are summarised in the Appendix A and in further publications in scientific journals (Grimbergen, et al., 2016) (Romero, et al., 2016) (Figel, et al., 2016) (Stadtman, et al., 2017).

IC2019_{exteye} for extremity and eye lens dosimeters (effectively two separate ICs) has built on the success of these previous ICs and included eye lens dosimeters in response to the recent increase in demand for these services which has arisen from the implementation in 2018 by EU Member States of the 20 mSv eye lens dose limit introduced by the European Council (European Council Directive 2013/59/EURATOM, 2014).

This summary report provides an overview of the set-up and the results for IC2019_{exteye}.

Please note: *The tables and figures are presented as accurately as possible. However, there may appear to be a few inconsistencies in some of the data e.g. total number of dosimeters. This is a consequence of different data samples used for individual table and figures, e.g. where there have been wrongly irradiated dosimeters, missing reported numerical dose values etc.*

2 Outline of the EURADOS IC2019_{exteye}

2.1 Organization Group

Andrew McWhan/Wioletta Dobrzynska	Cavendish Nuclear Limited, UK (Coordinating Lab)
Hannes Stadtmann	Seibersdorf Labor GmbH, AT
Markus Figel	Helmholtz-Zentrum Muenchen, DE
Tom Grimbergen	Mirion Dosimetry Services, NL
Ana M. Romero	Ciemat, ES
Isabelle Clairand	IRSN, Institut de Radioprotection et de Sûreté Nucléaire, FR

Coordinating Laboratory:

Andrew McWhan & Wioletta Dobrzynska
Cavendish Nuclear Limited, Berkeley Approved Dosimetry Service
Building A11, Gloucestershire Science and Technology Park
Berkeley, Gloucestershire, GL13 9FB, UK

2.2 Scope

The scope was for:

- i) Extremity dosimeters designed to estimate $H_p(0.07)$. The dosimeters could be of type ring, stall or wrist, for wearing on fingers, wrist or ankle and are used routinely in individual monitoring of exposed workers.
- ii) Eye lens dosimeters designed to estimate $H_p(3)$ which are worn on the head (not on the chest).
- iii) Extremity and eye lens dosimeters designed to measure either photons and betas (indicated in this report by PhB) or photons only (indicated in this report by Ph).

Irradiations were carried out in two accredited European irradiation facilities (GAEC and IRSN) in terms $H_p(0.07)$ and $H_p(3)$ in the following ranges:

- Photon energy: 16 to 662 keV (carried out by GAEC)
- Beta mean energy 250 to 1000 keV (carried out by IRSN)
- Dose: 0.5 mSv to 1 Sv
- Angle of incidence range: $\pm 60^\circ$

32 dosimeters were required from each participant. 22 extremity and 18 eye dosimeters were irradiated with the remaining dosimeters available for use as transit and/or spares.

2.3 Coordination

Cavendish Nuclear Limited (Berkeley Approved Dosimetry Service) acted as the coordinating institute. The task of the coordinator is to receive, forward for irradiation and return all dosimeters back to the participants. The coordinator also carries out all communication between the participants, the irradiation laboratories and the OG. This includes the dose values reported by the participants and the irradiation laboratory. The on-line platform (OLP) employed in the previous IC was modified to facilitate handling both extremity and eye dosimeters. The entire process of

registration, communication and data exchange was managed by the OLP. The participants were also kept up to date with the current status of their documents, dosimeters etc. via the OLP.

2.4 Set-up

IC2019 set-up followed the same pattern as all previous EURADOS ICs.

All dosimeters were forwarded by the coordinator to the irradiation laboratories accompanied by electronic dosimeters to monitor the doses received in storage and transit. Road transport was used to avoid air travel to help reduce the risk of x-ray inspections in transit.

After irradiation, the coordinator received the dosimeters and returned them to the IC participants for evaluation along with instructions on reporting their results. The online platform was used to transfer the evaluated dose values to the coordinator.

After receiving the results, the coordinator calculated the value of the response, R , for each dosimeter by dividing the participant's result ($H_{p, \text{participant}}$) by the reference dose given by the irradiation laboratory ($H_{p, \text{reference}}$) in accordance with equation (1).

$$R = \frac{H_{p, \text{participant}}}{H_{p, \text{reference}}} \quad (1)$$

The calculated response values were downloaded from the OLP by each participant for the initial check and confirmation along with the opportunity to add any comments. Thus, each participant was informed of the radiation qualities and the doses given to their dosimeters. All comments were assessed by the OG to decide if any applications for result modification could be permitted (see Section 3.1). Note that as a general rule, the OG only accepts changes to the results when it is clear that there has been an error by the OG or by the irradiation laboratory.

The original "Certificates of Participation", signed by the coordinator and the chairperson of EURADOS, were sent by standard mail to the participants. Copies were also made available on the OLP.

The Participants Meeting had been scheduled to be held at IM2020 (23 April) in Budapest but unfortunately both of these events had to be cancelled as a result of Covid-19.

2.5 Time schedule

Announcement - Call for participants	April 2019
*Registration of participants and systems	May 2019
Deadline for IMS sending application forms	31 May 2019
Deadline for IMS sending dosimeters to Coordinator	30 June 2019
Irradiations	July - September 2019
Coordinator sending dosimeters for readout	October 2019
Deadline for IMS sending dosimeters results to Coordinator	13 December 2019
Final results available	February 2020
IMS receiving certificates of participation	March 2020
Participants Meeting	Cancelled

2.6 Irradiation plan

Photon irradiation qualities were chosen from ISO 4037-1 (ISO 4037-1, 1996), ISO 4037-2 (ISO 4037-2, 1997) and ISO 4037-3 (ISO 4037-3, 1999). Beta irradiations were chosen from ISO 6980 (ISO 6980, 2015). The plan details were confidential and only known by the OG and the irradiation laboratory. The exact dose value (which was varied from the nominal value) was selected by the irradiation laboratory for each irradiation. These exact dose values were known only by the irradiation laboratory until all measured values had been reported to the coordinator.

Table 1: Outline irradiation plan for the EURADOS 2019 ICs. Irradiation qualities are reported according to ISO 4037 and ISO 6980

Quality	Radiation type	Dose range	Number of Extremity dosemeters to be irradiated	Number of Eye lens dosemeters to be irradiated	Dose range [mSv]
W-80 0°	Photon	Low	4	4	2 - 10
W-80 0°	Photon	Medium	2	2	40 - 60
W-80 0°	Photon	High	2	2	300 - 400
Cs-137 0°	Photon	Low	2	2	2 - 10
N-30 0°	Photon	Low	2	2	2 - 10
N-30 60°	Photon	Low	2	2	2 - 10
W-110 0°	Photon	Low	2	2	2 - 10
Sr-90 0°	Beta	Low	2	2	2 - 10
Sr-90 45°	Beta	Low	2	0	2 - 10
Kr-85 0°	Beta	Low	2	0	2 - 10
		TOTAL	22	18	

Table 2: Detailed irradiation plans for extremity for the EURADOS 2019 IC s. Irradiation qualities are reported according to ISO 4037 and ISO 6980

radiation type		quality	min (mSv)	mean (mSv)	max (mSv)
photon	X-ray	N-30	4.3	5.0	5.7
		N-30/60°	2.8	4.9	5.7
		W-80-L	4.3	5.0	5.7
		W-80-M	43	50	58
		W-80-H	315	349	382
		W-110	4.3	4.9	5.8
	gamma	Cs-137	4.3	5.0	5.8
beta		Kr-85	4.5	5.3	8.0
		Sr-90	3.0	3.5	5.0
		Sr-90/45°	2.8	2.8	3.0
All			2.8	40.0	382.0

Table 3: Detailed irradiation plans for eye lens dosimeters.

radiation type		quality	min (mSv)	mean (mSv)	max (mSv)
photon	X-ray	N-30	4.3	4.9	5.7
		N-30/60°	4.3	4.9	5.7
		W-80-L	4.3	5.0	8.5
		W-80-M	43	50	57
		W-80-H	319	343	384
		W-110	4.3	5.0	5.8
	gamma	Cs-137	4.3	5.0	5.7
beta		Sr-90	2.2	2.2	2.2
All			2.2	47.5	384.0

2.7 Participants and systems

60 IMS from 26 countries participated with a total of 113 dosimetry systems:

Table 4 provides a summary of all participating countries and the number of participating services. All information was provided by the participants on the OLP and application forms.

- 60 Participants
- 113 Systems
- 26 Countries

Table 4: Number of participants and IMS per country

Individual Monitoring Service:	60																																																																																																																
Countries:	26: Austria; Bosnia and Herzegovina; Belgium; Bulgaria; Switzerland; Chile; Czech Republic; Germany; Denmark; Spain; France; United Kingdom; Greece; Croatia; Israel; Italy; Japan; Lithuania; Netherlands; Portugal; Romania; Serbia; Sweden; Slovenia; Turkey; United States;																																																																																																																
EURADOS Sponsor:	<table border="1"><thead><tr><th>Country</th><th>no</th><th>yes</th><th>total</th></tr></thead><tbody><tr><td>Austria ^{EU}</td><td>0</td><td>2</td><td>2</td></tr><tr><td>Bosnia and Herzegovina</td><td>1</td><td>0</td><td>1</td></tr><tr><td>Belgium ^{EU}</td><td>2</td><td>2</td><td>4</td></tr><tr><td>Bulgaria ^{EU}</td><td>1</td><td>0</td><td>1</td></tr><tr><td>Switzerland</td><td>0</td><td>2</td><td>2</td></tr><tr><td>Chile</td><td>1</td><td>0</td><td>1</td></tr><tr><td>Czech Republic ^{EU}</td><td>2</td><td>1</td><td>3</td></tr><tr><td>Germany ^{EU}</td><td>0</td><td>2</td><td>2</td></tr><tr><td>Denmark ^{EU}</td><td>0</td><td>1</td><td>1</td></tr><tr><td>Spain ^{EU}</td><td>2</td><td>1</td><td>3</td></tr><tr><td>France ^{EU}</td><td>1</td><td>2</td><td>3</td></tr><tr><td>United Kingdom ^{EU}</td><td>0</td><td>2</td><td>2</td></tr><tr><td>Greece ^{EU}</td><td>1</td><td>1</td><td>2</td></tr><tr><td>Croatia ^{EU}</td><td>2</td><td>1</td><td>3</td></tr><tr><td>Israel</td><td>2</td><td>0</td><td>2</td></tr><tr><td>Italy ^{EU}</td><td>11</td><td>0</td><td>11</td></tr><tr><td>Japan</td><td>1</td><td>0</td><td>1</td></tr><tr><td>Lithuania ^{EU}</td><td>2</td><td>0</td><td>2</td></tr><tr><td>Netherlands ^{EU}</td><td>0</td><td>1</td><td>1</td></tr><tr><td>Portugal ^{EU}</td><td>0</td><td>1</td><td>1</td></tr><tr><td>Romania ^{EU}</td><td>1</td><td>1</td><td>2</td></tr><tr><td>Serbia</td><td>1</td><td>1</td><td>2</td></tr><tr><td>Sweden ^{EU}</td><td>2</td><td>0</td><td>2</td></tr><tr><td>Slovenia ^{EU}</td><td>1</td><td>0</td><td>1</td></tr><tr><td>Turkey</td><td>3</td><td>0</td><td>3</td></tr><tr><td>United States</td><td>2</td><td>0</td><td>2</td></tr><tr><td></td><td>39</td><td>21</td><td>60</td></tr></tbody></table>	Country	no	yes	total	Austria ^{EU}	0	2	2	Bosnia and Herzegovina	1	0	1	Belgium ^{EU}	2	2	4	Bulgaria ^{EU}	1	0	1	Switzerland	0	2	2	Chile	1	0	1	Czech Republic ^{EU}	2	1	3	Germany ^{EU}	0	2	2	Denmark ^{EU}	0	1	1	Spain ^{EU}	2	1	3	France ^{EU}	1	2	3	United Kingdom ^{EU}	0	2	2	Greece ^{EU}	1	1	2	Croatia ^{EU}	2	1	3	Israel	2	0	2	Italy ^{EU}	11	0	11	Japan	1	0	1	Lithuania ^{EU}	2	0	2	Netherlands ^{EU}	0	1	1	Portugal ^{EU}	0	1	1	Romania ^{EU}	1	1	2	Serbia	1	1	2	Sweden ^{EU}	2	0	2	Slovenia ^{EU}	1	0	1	Turkey	3	0	3	United States	2	0	2		39	21	60
Country	no	yes	total																																																																																																														
Austria ^{EU}	0	2	2																																																																																																														
Bosnia and Herzegovina	1	0	1																																																																																																														
Belgium ^{EU}	2	2	4																																																																																																														
Bulgaria ^{EU}	1	0	1																																																																																																														
Switzerland	0	2	2																																																																																																														
Chile	1	0	1																																																																																																														
Czech Republic ^{EU}	2	1	3																																																																																																														
Germany ^{EU}	0	2	2																																																																																																														
Denmark ^{EU}	0	1	1																																																																																																														
Spain ^{EU}	2	1	3																																																																																																														
France ^{EU}	1	2	3																																																																																																														
United Kingdom ^{EU}	0	2	2																																																																																																														
Greece ^{EU}	1	1	2																																																																																																														
Croatia ^{EU}	2	1	3																																																																																																														
Israel	2	0	2																																																																																																														
Italy ^{EU}	11	0	11																																																																																																														
Japan	1	0	1																																																																																																														
Lithuania ^{EU}	2	0	2																																																																																																														
Netherlands ^{EU}	0	1	1																																																																																																														
Portugal ^{EU}	0	1	1																																																																																																														
Romania ^{EU}	1	1	2																																																																																																														
Serbia	1	1	2																																																																																																														
Sweden ^{EU}	2	0	2																																																																																																														
Slovenia ^{EU}	1	0	1																																																																																																														
Turkey	3	0	3																																																																																																														
United States	2	0	2																																																																																																														
	39	21	60																																																																																																														

Table 5: Dosimetry System Analysis – Quantity, Position and Materials

Dosimetry System

Dosimetry System:	113																				
Measurement Quantity:	<table border="1"> <thead> <tr> <th>Measurement Quantity</th> <th>number</th> </tr> </thead> <tbody> <tr> <td>H_p(0.07)</td> <td>70</td> </tr> <tr> <td>H_p(3)</td> <td>43</td> </tr> <tr> <td></td> <td>113</td> </tr> </tbody> </table>	Measurement Quantity	number	H _p (0.07)	70	H _p (3)	43		113												
Measurement Quantity	number																				
H _p (0.07)	70																				
H _p (3)	43																				
	113																				
Dosemeter wearing position:	<table border="1"> <thead> <tr> <th>Dosemeter wearing position</th> <th>number</th> </tr> </thead> <tbody> <tr> <td>Extremity, Finger ... H_p(0.07)</td> <td>48</td> </tr> <tr> <td>Extremity, Fingertip ... H_p(0.07)</td> <td>4</td> </tr> <tr> <td>Extremity, Ankle/Wrist ... H_p(0.07)</td> <td>16</td> </tr> <tr> <td>Eye lens, Head ... H_p(3)</td> <td>45</td> </tr> <tr> <td></td> <td>113</td> </tr> </tbody> </table>	Dosemeter wearing position	number	Extremity, Finger ... H _p (0.07)	48	Extremity, Fingertip ... H _p (0.07)	4	Extremity, Ankle/Wrist ... H _p (0.07)	16	Eye lens, Head ... H _p (3)	45		113								
Dosemeter wearing position	number																				
Extremity, Finger ... H _p (0.07)	48																				
Extremity, Fingertip ... H _p (0.07)	4																				
Extremity, Ankle/Wrist ... H _p (0.07)	16																				
Eye lens, Head ... H _p (3)	45																				
	113																				
Radiation type in certificate:	<table border="1"> <thead> <tr> <th>Radiation type in certificate</th> <th>number</th> </tr> </thead> <tbody> <tr> <td>Photon results only</td> <td>57</td> </tr> <tr> <td>Photon and Beta results</td> <td>56</td> </tr> <tr> <td></td> <td>113</td> </tr> </tbody> </table>	Radiation type in certificate	number	Photon results only	57	Photon and Beta results	56		113												
Radiation type in certificate	number																				
Photon results only	57																				
Photon and Beta results	56																				
	113																				
Detector Type / Material:	<table border="1"> <thead> <tr> <th>Detector Type / Material</th> <th>number</th> </tr> </thead> <tbody> <tr> <td>Other</td> <td>1</td> </tr> <tr> <td>OSL - Al₂O₃:C</td> <td>3</td> </tr> <tr> <td>OSL - BeO</td> <td>9</td> </tr> <tr> <td>RPL</td> <td>3</td> </tr> <tr> <td>TL - Other</td> <td>5</td> </tr> <tr> <td>TL - Li₂B₄O₇/CaSO₄</td> <td>8</td> </tr> <tr> <td>TL - LiF:Mg,Cu,P</td> <td>48</td> </tr> <tr> <td>TL - LiF:Mg,Ti</td> <td>36</td> </tr> <tr> <td></td> <td>113</td> </tr> </tbody> </table>	Detector Type / Material	number	Other	1	OSL - Al ₂ O ₃ :C	3	OSL - BeO	9	RPL	3	TL - Other	5	TL - Li ₂ B ₄ O ₇ /CaSO ₄	8	TL - LiF:Mg,Cu,P	48	TL - LiF:Mg,Ti	36		113
Detector Type / Material	number																				
Other	1																				
OSL - Al ₂ O ₃ :C	3																				
OSL - BeO	9																				
RPL	3																				
TL - Other	5																				
TL - Li ₂ B ₄ O ₇ /CaSO ₄	8																				
TL - LiF:Mg,Cu,P	48																				
TL - LiF:Mg,Ti	36																				
	113																				
Number of detectors / filters:	<table border="1"> <thead> <tr> <th>Number of detectors / filters</th> <th>number</th> </tr> </thead> <tbody> <tr> <td>1</td> <td>91</td> </tr> <tr> <td>2</td> <td>14</td> </tr> <tr> <td>3</td> <td>1</td> </tr> <tr> <td>4</td> <td>6</td> </tr> <tr> <td>> 4</td> <td>1</td> </tr> <tr> <td></td> <td>113</td> </tr> </tbody> </table>	Number of detectors / filters	number	1	91	2	14	3	1	4	6	> 4	1		113						
Number of detectors / filters	number																				
1	91																				
2	14																				
3	1																				
4	6																				
> 4	1																				
	113																				

Note: This table is taken directly from the data supplied to the OLP by the participants. Two of the participants selected to report eye lens dose as $H_p(0.07)$ instead of $H_p(3)$. For the analysis in this report, all results for eye lens dosimeters have been taken as $H_p(3)$. Consequently, there are 45 reported system results for eye lens and 68 for extremity.



Figure 1: Representative selection of finger ring doseometers




Figure 2: Representative selection of wrist and finger tip doseometers



Figure 3: Representative selection of eye lens dosimeters



Figure 4: Representative selection of eye lens dosimeters with headbands

Table 6: Number⁻¹ of systems per dosimeter type and type of detector and dose quantity

	systems	% of all	% of type
Extremity	68	60%	
LiF:Mg,Cu,P	28	25%	41%
LiF:Mg,Ti	21	19%	31%
Li2B4O7/CaSO4	6	5%	9%
BeO	5	4%	7%
Al2O3:C	3	3%	4%
RPL	2	2%	3%
TL-Li2B4O7	1	1%	1%
Aluminophosphate glass	1	1%	1%
nLi2B4O7:Cu	1	1%	1%
Eye	45	40%	
LiF:Mg,Cu,P	20	18%	44%
LiF:Mg,Ti	15	13%	33%
BeO	4	4%	9%
Li2B4O7/CaSO4	2	2%	4%
Foma dosimetric film	1	1%	2%
Li2B4O7:Mn,Si	1	1%	2%
RPL	1	1%	2%
CaSO4:Dy	1	1%	2%
All	113	100%	100%

⁻¹ see note below Table 5 for details of total numbers of extremity and eye systems




Figure 5: Summary of all systems extremity & eye (Ph for photons only, PhB for photons and betas as indicated by the participants)

Table 7: Reference radiation qualities provided by the participants

Extremity	Cs-137	66%
	Mixed	12%
	Single x-ray	22%
Eye Lens	Cs-137	56%
	Co-60	2%
	Mixed	13%
	Single x-ray	29%

2.8 Intercomparison procedure compared to ISO 14146

IC2019 was set up to meet the ISO standard (ISO 14146, 2018) "Criteria and performance limits for the periodic evaluation of dosimetry services". The performance limits set by ISO 14146 were also adopted for the analysis of the global results:

$$0.71 \cdot \left(1 - \frac{2 \cdot \frac{H_0}{1.33}}{\frac{H_0}{1.33} + H_c} \right) \leq R \leq 1.67 \cdot \left(1 + \frac{H_0}{4 \cdot H_0 + H_c} \right)$$

Where R is the response, H_c is the conventional true value and H_0 is the "lower dose limit below which irradiations should not be performed":

- $H_0 = 0.3$ mSv for eye lens dosimeters measuring $H_p(3)$
- $H_0 = 1$ mSv for extremity dosimeters measuring $H_p(0.07)$

For the analysis of the global results in this report, any result exceeding the above performance limits was considered as an outlier. The standard ISO 14146:2018 allows a maximum of one-tenth of the dosimeters irradiated to exceed these limits.

In total approximately 2314 dosimeters were irradiated with GAEC carrying out 1540 photon irradiations and IRSN carrying out 774 beta irradiations.

2.8.1 Photon irradiations at GAEC

Ionizing Radiation Calibration Laboratory of the Greek Atomic Energy Commission performed the photon irradiations during summer 2019. The calibrations were performed in terms of $H_p(0.07)$ for ring and wrist dosimeters and in terms of $H_p(3)$ for eye dosimeters. For gamma irradiations the OB6 S-Cs-137(STS GMBH OB6) irradiator was used. For x-ray irradiations a PANTAK HF225 x-ray generator was used, producing 3 different qualities: W-80, W-110, N-30. Dosimeters were irradiated with angles of incidence 0° and 60°. The characteristics of the x-ray beams are in accordance with ISO 4037/1-2-3 (ISO 4037-1, 1996) (ISO 4037-2, 1997) (ISO 4037-3, 1999).

Three different phantoms were used for the irradiation of the three different extremity dosimeters.

- Finger dosimeters: ISO rod phantom (cylinder diameter 19 mm, length 300 mm, PMMA)
- Wrist dosimeters: Cylindrical water wrist phantom (cylinder diameter 7 cm, length 30 cm)
- Eye lens dosimeters: Cylindrical water head phantom (height 20 cm and diameter 20 cm).

For gamma irradiations buildup plate with dimensions (0.2 x 30 x 30) cm³ was used during irradiations. The K_{air} reference values were obtained using the reference/transfer ionization chambers PTW W-32002-LS01 (S/N 69) or FC65-G and the electrometer PTW UNIDOS 10002 (S/N 20314). The LS01 chamber was calibrated in PTB during 06-07/03/2019 (PTB, Cal. Cert. No PTB-6.3-4094018). The irradiation conditions are in accordance with ISO 4037/1-2-3-4. The irradiation uncertainty was calculated from 4.9% up to 5.1%.



Figure 6: Finger, wrist and head phantoms used by GAEC for photon irradiations

2.8.2 Beta irradiations at IRSN

The laboratory of dosimetry of ionizing radiation (LDRI) at the Institute for Radiological and Nuclear Safety in France (IRSN) was in charge of the beta irradiations in this intercomparison. The irradiations took place in 2019, during summer, using the equipment shown in Figure 7. This is an ISOTRACK™–BSS2 beta irradiator placed on a table in order to obtain the reference point at a height of 1.50 m. The irradiation distance is defined by a standard gauge of 30 cm length. About 30 wrist dosimeters and 100 ring dosimeters were irradiated in each configuration in terms of $H_p(0.07)$: $^{90}\text{Sr}/\text{Y}$ at 0° and 45°; ^{85}Kr at 90°. About 90 eye lens dosimeters were irradiated for $^{90}\text{Sr}/\text{Y}$ at 0° configuration in terms of $H_p(3)$. The irradiation uncertainty was estimated to be between 5.7 and 6.2 % at k=2, depending on the configuration. The sources used for the irradiations are traceable to the PTB (IPFCO 15277-AE-4147 for ^{85}Kr and IPFCO 15277 –AG-4291 for $^{90}\text{Sr}/\text{Y}$) and the phantoms designed in the ISO standard (pillar, rod and head) were used to simulate the appropriate quantity in the standard beta field described in ISO 6980-1 (ISO 6980, 2015).



Figure 7: Beta irradiator (ISOTRACK – BSS2) used in this intercomparison

2.9 Background and transit dose control

For this intercomparison a total of 32 dosimeters were required from each participant. 22 extremity and 18 eye lens dosimeters were irradiated which allowed for 10 extremity dosimeters and 14 eye lens dosimeters to be designated for “background and transit dose control” for IMS to make corrections for background and transfer dose. These dosimeters were also available for use by the irradiation laboratory in case of damage or errors during the irradiations although only a few of these were required. The coordinator did not make any adjustment for transit or background dose and all associated correction was the responsibility of the participants.

The results of the electronic dosimeters (see Section 2.4) which accompanied the participants’ dosimeters for the land travel between the coordinator and the irradiation laboratories are shown in the table below. The participants were not informed of these dose values.

Table 8: EPD Transit Control Doses – road transport between coordinator and Irradiation Laboratories

Depart	Arrive	Transit Dose mSv	Depart	Arrive	Transit Dose mSv
Berkeley 11-07-2019	GAEC 26-07-2019	0.022	GAEC 14-09-2019	Berkeley 18-09-2019	0.016
Berkeley 11-07-2019	IRSN 17-07-2019	0.010	IRSN 17-09-2019	Berkeley 18-09-2019	0.003

The coordinator provided the participants with the identification codes of the unused “background and transit dose control” dosimeters. The mean values per system for all of the non-irradiated “background and transit control” dosimeters are shown in Figure 8. It can be seen that few significant transit doses were reported and that some of the participants reported zero doses.




Figure 8: Background/transport doses: each bar represents the mean reported background dose value for each individual system

2.10 Confidentiality of the data and the results

The data processed by the OG is always treated confidentially for two specific reasons.

Firstly, the IC programme was designed as a blind test for all the participants. This means that participants report their results without knowing any specific details of the irradiation plan, in particular the dose values.

The dose values are reported to the participants only after the coordinator has received the dose values evaluated by the participant. At the time of application for the IC, only the ranges of dose, energies and angles are known to the participants (section: 2.4 Set-up). Direct communication between participants and the irradiation laboratory is not allowed. It is acknowledged that some IMS participate with more than one dosimetry system and that some IMS might have access to results of other participants. In order to prevent these participants guessing dose values by combining results, the irradiation plan is executed in a random order for each participant. In addition, the irradiation laboratory varies the dose values in the irradiation plan within specified ranges from participant to participant, rather than using fixed dose values for each radiation quality.

Secondly, the individual results are the property of the participants only and thus have to be kept confidential.

To assure this confidentiality the coordinator separates all information which could possibly lead to the identity of the participants from the published results. In the overviews of the results, the participating dosimetry systems are only referenced by a randomized code. All participants' certificates contained "Reporting number used in publications" known only to the participant and

the coordinator. However, once participants have received their certificates, OG has no influence on the use that participants make of their own results.

During the IC exercises significant quantities of data have to be exchanged. In order to assure data integrity, parallel data streams (paper & electronic) are used. All official results are reported on signed papers. However, the majority of the data is exchanged in electronic formats by means of the web based OLP. To maintain data security, all relevant electronic documents containing IC results (draft results or certificate) are password protected.

2.11 Certificates of participation and participants meeting

EURADOS is not accredited for the evaluation of IMS and consequently the results issued by EURADOS itself cannot be regarded as an official test report. As an alternative, the established protocol is to report the results to the individual participants in the form of a "Certificate of Participation", with the irradiation reports of the accredited irradiation laboratories as an annex.

These certificates consist of two pages. The front page shows the certificate number, the reporting number, the details of the participant, the description of the system as given by the participant and a summary of the IC procedure. The front page is signed by both the EURADOS Chairperson and the IC coordinator. The second page shows the actual results for each dosimeter, irradiation quality, value of $H_p(0.07)$ or $H_p(3)$ as reported by participant, value of $H_p(0.07)$ or $H_p(3)$ as reported by the irradiation laboratory, and the ratio of these two values. In the Certificates, no performance limits were indicated because these might differ from one participant to the other. All participants received their signed certificates by post.

3 Results and Discussion

3.1 Review of the comments received from participants

After the participants submitted the results, the coordinator processed the dose values and created a draft report for each system. These reports were made available to each participant followed by a two-week time frame to comment on their report. The participants were informed that if they requested corrections because of errors made by the OG organization, they would have to indicate the results they were requesting to be corrected and the justification. The OG would have the final decision on any amendments.

In total 6 requests were received relating to 9 separate systems:

1. Lower beta results than expected – one IMS requested for information on irradiation conditions
2. Higher x-ray results than expected – one IMS requested confirmation that irradiations were correct
3. Cs-137 response higher than expected – reported by one IMS
4. Three separate IMS reported the same possible problem with one specific quality

OG convened before the release of the final results to review and discuss these requests. In this process confidentiality is maintained as far as practicable and only the coordinator knows the identity of the participants who have sent these comments.

In accordance with the terms and conditions, OG concluded that the requests 1, 2 & 3 (see above) could not be allowed and an appropriate letter was sent to these participants to explain that the OG was satisfied that all irradiations had been carried out in accordance with the appropriate ISO standards. However, after further investigation, OG concluded that there had been a problem with the one specific quality as reported by the three separate IMS (see 4 above) but it was established that none of the other participants had been affected. OG sent appropriate letters of apology informing these participants that these irradiations would be excluded from the certificates of participation.

3.2 Basic statistical results

The response (R) was calculated for each dosimeter by dividing the participant's result, $H_{p, \text{participant}}$ by the reference dose value (given by irradiation laboratory) $H_{p, \text{reference}}$ in accordance with equation (1). The corresponding frequency and cumulative distributions as well as basic statistical results are given for these distributions in the next figures.




Figure 9: Frequency and cumulative distributions of extremity response values




Figure 10: Frequency and cumulative distributions of eye response values

3.3 Response values per radiation quality

The energy and the angular response is given in the next diagrams.




Figure 11: Distributions of all $H_p(0.07)$ response values for different radiation qualities. Diamond (Median), box (50% range), bar (90% range), dots (minimum, maximum).




Figure 12: Distributions of all $H_p(0.07)$ mean response values for single systems for Sr-90 (only for beta dosemeters) and N-30 and different radiation angles and the corresponding ratios. Diamond (Median), box (50% range), bar (90% range), dots (minimum, maximum).




Figure 13: Distributions of all $H_p(3)$ response values for different radiation qualities. Diamond (Median), box (50% range), bar (90% range), dots (minimum, maximum).




Figure 14: Distributions of all $H_p(3)$ mean response values for single systems for N-30 and different radiation angles and the corresponding ratio. Diamond (median), box (50% range), bar (90% range), dots (minimum, maximum).

3.4 Linearity

Linearity was tested (using W-80) through a range, from 2 mSv to 400 mSv, without varying any other parameter.




Figure 15: Distribution of the mean $H_p(0.07)$ response values for W-80 irradiations for different doses (L: low dose; M: medium dose; H: high dose) and the distribution for different mean response ratios




Figure 16: Distribution of the mean $H_p(3)$ response values for W-80 irradiations for different doses (L: low dose; M: medium dose; H: high dose) and the distribution for different mean response ratios

3.5 Response values as a function of reference doses




Figure 17: Response values (lin) for $H_p(0.07)$ as a function of reference dose. The dashed lines represent the trumpet curves




Figure 18: Response values (log) for $H_p(0.07)$ as a function of reference dose. The dashed lines represent the trumpet curves




Figure 19: Response values (lin) for $H_p(3)$ as a function of reference dose. The dashed lines represent the trumpet curves




Figure 20: Response values (log) for $H_p(3)$ as a function of reference dose. The dashed lines represent the trumpet curves

3.6 Outliers

Table 9: Outliers - Extremity (defining outliers as response values outside of the trumpet curves)

Extremity - number of Outliers					Extremity - Outliers (%)				
Radiation	Quality	Ph	PhB	All	Radiation	Quality	Ph	PhB	All
beta	Kr-85/0°		44	44	beta	Kr-85/0°		57.9%	57.9%
	Sr-90/0°		1	1		Sr-90/0°		1.3%	1.3%
	Sr-90/45°		1	1		Sr-90/45°		1.3%	1.3%
	all betas		46	46		all betas		20.2%	20.2%
photon	N-30/0°	1	1	2	photon	N-30/0°	1.7%	1.3%	1.5%
	N-30/60°	2	2	4		N-30/60°	3.3%	2.6%	2.9%
	W-80/0°	16	5	21		W-80/0°	6.7%	1.6%	3.9%
	W-110/0°	0	0	0		W-110/0°	0.0%	0.0%	0.0%
	Cs-137/0°	2	0	2		Cs-137/0°	3.3%	0.0%	1.5%
	all photons	21	8	29		all photons	4.4%	1.3%	2.7%
All		21	54	75	All		4.4%	6.5%	5.7%

Table 10: Outliers - Eye (defining outliers as response values outside of the trumpet curves)

Eye - number of Outliers					Eye - Outliers (%)				
Radiation	Quality	Ph	PhB	All	Radiation	Quality	Ph	PhB	All
beta	Sr-90/0°		16	16	beta	Sr-90/0°		44.4%	44.4%
	N-30/0°	4	2	6		N-30/0°	7.7%	6.3%	7.1%
	N-30/60°	2	6	8		N-30/60°	3.7%	16.7%	8.9%
	W-80/0°	20	8	28		W-80/0°	9.3%	5.6%	7.8%
photon	W-110/0°	4	0	4		W-110/0°	7.4%	0.0%	4.4%
	Cs-137/0°	7	0	7		Cs-137/0°	13.0%	0.0%	7.8%
	all photons	37	16	53		all photons	8.6%	5.6%	7.4%
	All	37	32	69		All	8.6%	10.0%	9.2%

3.7 Results for individual systems

This paragraph presents results for individual systems separately (but anonymously). Individual systems are represented with a “reporting number”. This number was randomly assigned and has no relation to the participant number as used by the organizer for keeping track of correspondence etc.

Response values for each individual system are shown separately (by anonymous “reporting number”) in Figure 16, Figure 17 and Figure 18. It shows that most outliers are associated with a relatively small number of systems. Some systems show a significant bias, others have a greater than normal spread of results.

It should be noted, as in previous ICs, that there are examples of excellent performances within each type of dosimetry system.




Figure 21: Response values (lin) for $H_p(0.07)$ and $H_p(3)$ for each individual participant system




Figure 22: Response values (log) for $H_p(0.07)$ and $H_p(3)$ for each individual participant system

For each participating system a separate datasheet was prepared summarizing all the results and the underlying data. Data for $H_p(0.07)$ and $H_p(3)$ (if any) is presented in separate sheets. Each sheet shows the data reported by the irradiation laboratory and by the participant, and the response value calculated from these values (separately for each irradiation). Data has been combined for the radiation qualities and some statistical quantities are also provided. Two figures have been included

to show the response values in the trumpet curve and the response values for the different radiation qualities.

These sheets have been prepared primarily to enable the participants to analyse their own results and to compare these with the results of the other participants. The individual results will not be analysed in further detail in this report. The datasheets for all participants can be inspected in Appendix F.

4 Conclusions

EURADOS WG2 carried out IC2019_{exteye} for extremity and eye lens dosimeters as part of the programme which EURADOS has developed for self-sustained ICs for IMS for external radiation. It has been established that the results from these ICs can assist all participants to demonstrate compliance within their own quality management system, compare their results with other participants and develop action plans for improving their own systems.

IC2019_{exteye} had 113 participating systems (68 extremity and 45 eye lens) from 60 institutes with participants from 39 countries from around the world. In general, the participants showed a very satisfactory performance with the median of all response values very close to unity. This finding confirms that, in general, the calibration procedures, especially the traceability to standard metrology laboratories, meet the required standards without any general bias. However, the results, in particular the outliers, also indicate that a small number of IMS could improve the quality of their systems by reviewing their calibration procedures.

From the statistical overviews of the results, the reports provided by the irradiation laboratories and the remarks received from the participants, it can be concluded that the intercomparison exercise was successful and that there were no significant issues encountered by the intercomparison organization or the irradiation laboratories during the execution of the exercise.

Specific additional information was supplied by the participants. This allowed a more detailed analysis of the results with respect to detector type. The observed characteristics are generally in agreement with results established in scientific literature.

The high number of participating systems confirms that there is a significant demand for regular, internationally organized, ICs and that these are of operational value for individual monitoring services. Therefore, it is planned to continue with this programme of EURADOS ICs.

References

- Bartlett, D., Ambrosi, P., Bordy, J. & Van Dijk., J., 2000. Harmonisation and Dosimetric Quality Assurance in Individual Monitoring for External Radiation. *Radiat. Prot. Dosim.*, 89 (Special Issue)((1-2)).
- Bartlett, D., Boehm, J. & Hyvonen, H., 2001. Individual Monitoring of External Radiation.. *Radiat. Prot. Dosim.*, 96 (Special Issue)((1-3)).
- European Commission, 2009. Technical Recommendations for Monitoring Individuals Occupationally Exposed to External Radiation. *Directorate-General for Energy and Transport RP 160*.
- European Council Directive 2013/59/EURATOM, 2014. European Council Directive 2013/59/EURATOM. *Official Journal of the European Union*.
- Figel, M. et al., 2016. EURADOS Intercomparisons on Whole-Body Dosemeters for Photons from 2008 To 2014. *Rad. Prot. Dosim.*, 170(1-4), p. 113–116.
- Grimbergen, T. et al., 2016. EURADOS Programme of Intercomparisons for Individual Monitoring Services: Seven Years of Development and Future Plans. *Rad. Prot. Dosim.*, 170(1-4), p. 90–94.
- Grimbergen, T. et al., 2012. EURADOS Intercomparison 2008 for Whole Body Dosemeters in Photon Fields. *EURADOS Report 2012-01*.
- Grimbergen, T. et al., 2013. EURADOS Intercomparison 2009 for Extremity Dosemeters in Photon and Beta Fields,. *EURADOS Report 2013-01*.
- ISO 14146, 2018. Criteria and performance limits for the periodic evaluation of dosimetry services. *International Organization for Standardization*.
- ISO 17025, 2017. General requirements for the competence of testing and calibration laboratories. *International Organization for Standardization*.
- ISO 4037-1, 1996. X and gamma reference radiation for calibrating dosemeters and doserate meters and for determining their response as a function of photon energy — Part 1: Radiation characteristics and production methods.. *International Organization for Standardization*.
- ISO 4037-2, 1997. X and gamma reference radiation for calibrating dosemeters and doserate meters and for determining their response as a function of photon energy — Part 2: Dosimetry for radiation protection over the energy ranges from 8 keV to 1,3 MeV.... *International Organization for Standardization*.
- ISO 4037-3, 1999. X and gamma reference radiations for calibrating dosemeters and doserate meters and for determining their response as a function of photon energy. Part 3: Calibration of area and personal dosemeters and the measurement of their response.... *International Organization for Standardization*.
- ISO 6980, 2015. Nuclear energy - Reference beta-particle radiation - Part 3: calibration of area and personal dosemeters as a function of radiation energy and angle of incidence. *International Organization for Standardization*.
- McWhan, A. et al., 2015. EURADOS Intercomparison 2010 for Whole Body Dosemeters in Photon Fields. *EURADOS Report 2015-01*.

McWhan, A. et al., 2015. EURADOS Intercomparison 2012 for Whole Body Dosemeters in Photon Fields. *EURADOS Report 2015-02*.

Romero, A. et al., 2016. EURADOS Intercomparisons in External Radiation Dosimetry: Similarities and Differences among Exercises for Whole-Body Photon, Whole-Body Neutron, Extremity, Eye-Lens and Passive Area Dosemeters. *Rad. Prot. Dosim.*, 170(1-4), pp. 82-85.

Stadtman, H. et al., 2016. The Results of the EURADOS Intercomparison IC2014 for Whole-Body Dosemeters in Photon Fieldsn Dosimetry. *Rad. Prot. Dosim.*, Volume 170, pp. 86-89.

Stadtman, H. et al., 2017. EURADOS intercomparisons for individual monitoring services: Results of the 2015 extremity dosimeter intercomparison for photon and beta radiations.. *Radiation Measurements DOI:10.1016/j.radmeas 2017.03.030*.

Stadtman, H. et al., 2018. EURADOS Intercomparison 2014 for Whole Body Dosemeters in Photon Fields. *EURADOS Report 2018-01*.

Van Dijk, J. et al., 2004. Harmonisation of Individual Monitoring in Europe. *Radiat. Prot. Dosim.*, 112 (Special Issue)(1).

Figures and tables

Figure 1: Representative selection of finger ring dosemeters	9
Figure 2: Representative selection of wrist and finger tip dosemeters	9
Figure 3: Representative selection of eye lens dosemeters.....	10
Figure 4: Representative selection of eye lens dosemeters with headbands.....	10
Figure 5 Summary of all systems extremity & eye (Ph for photons only, PhB for photons and betas as indicated by the participants).....	11
Figure 6: Finger, wrist and head phantoms used by GAEC for photon irradiations.....	13
Figure 7: Beta irradiator (ISOTRACK – BSS2) used in this intercomparison	14
Figure 8: Background/transport doses: each bar represents the mean reported background dose value for each individual system.....	15
Figure 9: Frequency and cumulative distributions of extremity response values.....	18
Figure 10: Frequency and cumulative distributions of eye response values.....	18
Figure 11: Distributions of all $H_p(0.07)$ response values for different radiation qualities. Diamond (Median), box (50% range), bar (90% range), dots (minimum, maximum).	19
Figure 12: Distributions of all $H_p(0.07)$ mean response values for single systems for Sr-90 (only for beta dosemeters) and N-30 and different radiation angles and the corresponding ratios. Diamond (Median), box (50% range), bar (90% range), dots (minimum, maximum).	19
Figure 13: Distributions of all $H_p(3)$ response values for different radiation qualities. Diamond (Median), box (50% range), bar (90% range), dots (minimum, maximum).	20
Figure 14: Distributions of all $H_p(3)$ mean response values for single systems for N-30 and different radiation angles and the corresponding ratio. Diamond (median), box (50% range), bar (90% range), dots (minimum, maximum).....	20
Figure 15: Distribution of the mean $H_p(0.07)$ response values for W-80 irradiations for different doses (L: low dose; M: medium dose; H: high dose) and the distribution for different mean response ratios	21
Figure 16: Distribution of the mean $H_p(3)$ response values for W-80 irradiations for different doses (L: low dose; M: medium dose; H: high dose) and the distribution for different mean response ratios	21
Figure 17: Response values (lin) for $H_p(0.07)$ as a function of reference dose. The dashed lines represent the trumpet curves.....	22
Figure 18: Response values (log) for $H_p(0.07)$ as a function of reference dose. The dashed lines represent the trumpet curves.....	22
Figure 19: Response values (lin) for $H_p(3)$ as a function of reference dose. The dashed lines represent the trumpet curves	23
Figure 20: Response values (log) for $H_p(3)$ as a function of reference dose. The dashed lines represent the trumpet curves	23

Figure 21: Response values (lin) for H_p (0.07) and H_p (3) for each individual participant system25

Figure 22: Response values (log) for H_p (0.07) and H_p (3) for each individual participant system.....25

Table 1: Outline irradiation plan for the EURADOS 2019 ICs. Irradiation qualities are reported according to ISO 4037 and ISO 69805

Table 2: Detailed irradiation plans for extremity for the EURADOS 2019 IC s. Irradiation qualities are reported according to ISO 4037 and ISO 6980.....6

Table 3: Detailed irradiation plans for eye lens dosimeters.....6

Table 4: Number of participants and IMS per country7

Table 5: Dosimetry System Analysis – Quantity, Position and Materials.....8

Table 6: Number⁻¹ of systems per dosimeter type and type of detector and dose quantity11

Table 7: Reference radiation qualities provided by the participants12

Table 8: EPD Transit Control Doses – road transport between coordinator and Irradiation Laboratories.....14

Table 9: Outliers - Extremity (defining outliers as response values outside of the trumpet curves) .24

Table 10: Outliers - Eye (defining outliers as response values outside of the trumpet curves).....24

Appendix A: Previous whole body and extremity intercomparisons organised by EURADOS WG2

Intercomparison Type of dosimeter	Systems / IMS / Countries	Dosemeters required / irradiated	Quantities compared	Radiation categories
IC2008 Whole body	62 / 52 / 21	26 / 20	$H_p(10)$ and $H_p(0.07)$	S-Cs, S-Co, N-60(0°,45°), N-150(±45°), N-60+S-Cs
IC2009 _{ext} Extremity	59 / 44 / 18	28 / 22	$H_p(0.07)$	S-Cs, N-20, W-80(0°, 60°), N-150, ⁸⁵ Kr, ⁹⁰ Sr (0°, 60°)
IC2010 Whole body	85 / 70 / 30	26 / 20	$H_p(10)$ and $H_p(0.07)$	S-Cs, S-Co, N-40(30°), W-110 (45°), N-40+S-Cs, W-250+S-Cs
IC2012 _{ph} Whole body	88 / 74 / 30	26 / 16	$H_p(10)$ and $H_p(0.07)$	S-Cs, S-Co, N-60(0°,60°)
IC2012 _n Whole body neutron	34 / 31 / 18	36 / 24	$H_p(10)$	250 keV neutrons, ²⁵² Cf(0°, 45°), ²⁵² Cf(D ₂ O), ²⁵² Cf+shadow cone
IC2014 Whole body	112 / 96 / 35	30 / 20	$H_p(10)$ and $H_p(0.07)$	S-Cs, S-Co, W-80(0°,60°), W-150, RQR7
IC2015 _{ext} Extremity	72 / 52 / 22	30 / 22	$H_p(0.07)$	S-Cs, RQR3(0°,60°) W-80, RQR9, ⁹⁰ Sr+S-Cs, Kr-85 ⁹⁰ Sr(0°,60°)
IC2016 Whole body	103 / 86 / 36	30 / 22	$H_p(10)$ and $H_p(0.07)$	S-Cs, S-Co, N-40(0°,60°), N-150 (0°,45°) S-Cs+ ⁹⁰ Sr
IC2017 _n Whole body neutron	33 / 32 / 17	40 / 28	$H_p(10)$	²⁵² Cf (0°,45°), ²⁵² Cf+S-Cs, ²⁴¹ Am-Be, ²⁵² Cf(D ₂ O), ²⁵² Cf+shadow block

Intercomparison Type of dosimeter	Systems / IMS / Countries	Dosemeters required / irradiated	Quantities compared	Radiation categories
IC2018 Whole Body	121 / 101 / 40	34 / 22	$H_p(10)$ and $H_p(0.07)$	S-Cs, S-Co, N-60(0°,60°), N-150 (60°) W-110
IC2019_{exteye} Extremity and Eye Lens	113 / 60 / 26 68: $H_p(0.07)$ 45: $H_p(3)$	$H_p(0.07)$ 32 / 22 $H_p(3)$ 32 / 18	$H_p(0.07)$ and $H_p(3)$	S-Cs W-80 N-30 N-30 60° W-110 Sr-90 Sr-90 45° Kr-85

Appendix B: List of participants

(Participants sorted alphabetically by country and institute)

Institute	Place	Country	Ext	Eye
International Atomic Energy Agency	Vienna	Austria	X	
Seibersdorf Labor GmbH	Seibersdorf	Austria	X	X
AV-CONTROLATOM	Vilvoorde	Belgium	X	X
Belgoprocess	Dessel	Belgium	X	X
Dosimetrie SCK-CEN	Mol	Belgium	X	X
Personendosimetrie UZ Leuven	Leuven	Belgium	X	
Ekoteh DOO Mostar	Mostar	Bosnia and Herzegovina	X	
Laboratories Protecta	Sofia	Bulgaria		X
OSL Control Chile SA	Santiago de Chile	Chile	X	X
Ekoteh Dosimetry Co Ltd	Zagreb	Croatia	X	
Ruđer Bošković Institute	Zagreb	Croatia	X	X
Institute for Medical Research and Occupational Health	Zagreb	Croatia	X	X
National Radiation Protection Institute	Praha	Czech Republic		X
Nuvia Dosimetry	Praha	Czech Republic	X	X
Sluzba osobni dozimetrie VF	Cerna Hora	Czech Republic	X	
Danish Health Authority	Herlev	Denmark	X	X
Laboratoire de Dosimétrie de l'IRSN	Croissy-sur-Seine	France	X	X
LANDAUER	Vélizy-Villacoublay Cedex	France	X	X
Orano Cycle La Hague Activite Dosimetrie	La Hague	France	X	X
AWST	München	Germany	X	X

Institute	Place	Country	Ext	Eye
Landesanstalt für Personendosimetrie und Strahlenschutzausbildung Mecklenburg-Vorpommern	Berlin	Germany	X	X
Mediray Personnel Dosimetry Laboratory	Athens	Greece	X	
Personal Dosimetry Departement of Greek Atomic Energy Commissiion	Agia Paraskevi, Attiki	Greece	X	X
NRCN	Beer-Sheva	Israel	X	
Soreq NRC	Yavne	Israel		X
AOU Careggi	Firenze	Italy		X
CIR sas	Rome	Italy	X	X
Dosimetrie-Service Bozen	Bozen	Italy	X	
ENEA Radiation Protection Institute Individual Monitoring Dosimetry Service -	Bologna	Italy	X	X
EUROPEAN COMMISSION - JOINT RESEARCH CENTRE - Dosimetry Service	Ispra (VA)	Italy	X	
L.B. Servizi per le Aziende Srl	Roma	Italy	X	X
Lavoro e Ambiente srl	Forlì (FC)	Italy	X	X
MCF Ambiente Srl	Pasian di Prato (Udine)	Italy	X	X
Nucleonova Srl	Palermo	Italy	X	
Servizio Dooosimetrico Azienda USL Della Romagna	Cesena	Italy	X	
Tecnorad s.u.r.l.	Verona	Italy	X	X
X-GAMMAGUARD	Saronno (VA)	Italy	X	X
Chiyoda Technol Corporation	Ibaraki	Japan	X	X
Radiation Protection Centre	Vilnius	Lithuania		X
Mirion	Arnhem	Netherlands	X	X
IST-LPSR	Bobadela LRS	Portugal	X	
DOZIMED S.R.L.	Magurele	Romania	X	X
Rodos Laboratories	Magurele	Romania	X	X

Institute	Place	Country	Ext	Eye
Vinca Institute of Nuclear Sciences, Radiation and Environmental Protection Department	Belgrade	Serbia	X	X
X-ray Kosutic – Ekoteh dozimetrija doo	Beograd	Serbia	X	
Slovak legal metrology, non-profit org.	Bratislava	Slovakia		
Jozef Stefan Institute	Ljubljana	Slovenia	X	
ZVD	Ljubljana Polje	Slovenia		
Centro De Dosimetria	Barcelona	Spain	X	X
Centro Nacional de Dosimetria	Valencia	Spain	X	X
CIEMAT External Dosimetry Service	Madrid	Spain	X	
Landauer Nordic Holdings AB	Uppsala	Sweden	X	X
Ringhals AB	Varobacka	Sweden		X
Dosilab	Köniz	Switzerland	X	X
Paul Scherrer Institut	Villigen PSI	Switzerland	X	
RADAT Dosimetry Laboratory Services Co.	Ankara	Turkey	X	X
RADKOR Training Teaching Medical Production Energy Electricity Electronic and Information Technology Ltd. Co.	Ankara	Turkey	X	X
Turkish Atomic Energy Authority Individual Monitoring Service	Ankara	Turkey	X	
Berkeley Approved Dosimetry Service	Berkeley	United Kingdom	X	X
Nuvia Harwell	Harwell	United Kingdom	X	
Landauer	Glenwood, IL	United States	X	X
Mirion Technologies (GDS), Inc.	Irvine	United States	X	X

Appendix C: Example Irradiation Certificate - photon

<p>IRRADIATION CERTIFICATE FOR PERSONNAL DOSEMETERS EEAE-IRCL-Cal. Cert. No: DOS /2353- █/19</p> <p></p> <p>HELLENIC REPUBLIC MINISTRY OF DEVELOPMENT AND INVESTMENTS GENERAL SECRETARIAT FOR RESEARCH & TECHNOLOGY</p> <p> Calibrations Cert. No 1164</p> <p>Ag. Paraskevi, 16/01/2020 Our Ref: B/428/842/2020</p> <p> ΕΕΑΕ ΕΛΛΗΝΙΚΗ ΕΠΙΤΡΟΠΗ ΑΤΟΜΙΚΗΣ ΕΝΕΡΓΕΙΑΣ GREEK ATOMIC ENERGY COMMISSION</p> <p>IONIZING RADIATION CALIBRATION LABORATORY Affiliated to the Hellenic Metrology Institute</p>	<p>E ERG(D)04 05 Page 1 of 2</p>																				
<p>IRRADIATION CERTIFICATE No: DOS /2353- █/19 Number of Pages: 2 Date of Issue: 20/01/2020</p>																					
<p>The following personnel dosimeters from: EURADOS INTERCOMPARISON PROGRAM System No: █ have been irradiated at the <i>Ionizing Radiation Calibration Laboratory of Greek Atomic Energy Commission</i>:</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 30%;">Personal Dosimeters (PD):</td> <td style="width: 70%;">Ring</td> </tr> <tr> <td>Dosimeter Identification:</td> <td>-</td> </tr> <tr> <td>Detection Principle:</td> <td>-</td> </tr> <tr> <td>Irradiation Period:</td> <td>See below</td> </tr> </table> <p>The Kair reference values have been obtained using the reference/transfer ionization chambers PTW W-32002-LS01 (S/N 69) or FC65-G and the electrometer PTW UNIDOS 10002 (S/N 20314). The LS01 chamber was calibrated in PTB during 06-07/03/2019 [PTB, Cal. Cert. No PTB-6.3-4094018]. The irradiation conditions are in accordance to ISO 4037/1-2-3-4 and IEC 62387.</p> <p>Irradiation conditions</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 20%;">Phantom:</td> <td>ISO rod phantom (cylinder diameter 19 mm, length 300 mm, PMMA)</td> </tr> <tr> <td>Source to PD Distance:</td> <td>100-200 cm, depending on required Kair rate</td> </tr> <tr> <td>Kair Rate:</td> <td>S-Cs: 217.1 µGy/min (at 150 cm) W-80: 2.64 mGy/min (at 200 cm) W-110: 1.96 mGy/min (at 200 cm) N-30: 0.36 mGy/min (at 150 cm)</td> </tr> <tr> <td>Field Size:</td> <td>S-Cs: Circular with diameter of 55.6 cm (at 200 cm) x-rays: Circular with diameter 26.8 cm (at 200 cm)</td> </tr> <tr> <td>Build up PMMA:</td> <td>S-Cs: (0.2 x 30x30) cm³</td> </tr> <tr> <td>Reference point of PD:</td> <td>Phantom surface</td> </tr> </table>		Personal Dosimeters (PD):	Ring	Dosimeter Identification:	-	Detection Principle:	-	Irradiation Period:	See below	Phantom:	ISO rod phantom (cylinder diameter 19 mm, length 300 mm, PMMA)	Source to PD Distance:	100-200 cm, depending on required Kair rate	Kair Rate:	S-Cs: 217.1 µGy/min (at 150 cm) W-80: 2.64 mGy/min (at 200 cm) W-110: 1.96 mGy/min (at 200 cm) N-30: 0.36 mGy/min (at 150 cm)	Field Size:	S-Cs: Circular with diameter of 55.6 cm (at 200 cm) x-rays: Circular with diameter 26.8 cm (at 200 cm)	Build up PMMA:	S-Cs: (0.2 x 30x30) cm ³	Reference point of PD:	Phantom surface
Personal Dosimeters (PD):	Ring																				
Dosimeter Identification:	-																				
Detection Principle:	-																				
Irradiation Period:	See below																				
Phantom:	ISO rod phantom (cylinder diameter 19 mm, length 300 mm, PMMA)																				
Source to PD Distance:	100-200 cm, depending on required Kair rate																				
Kair Rate:	S-Cs: 217.1 µGy/min (at 150 cm) W-80: 2.64 mGy/min (at 200 cm) W-110: 1.96 mGy/min (at 200 cm) N-30: 0.36 mGy/min (at 150 cm)																				
Field Size:	S-Cs: Circular with diameter of 55.6 cm (at 200 cm) x-rays: Circular with diameter 26.8 cm (at 200 cm)																				
Build up PMMA:	S-Cs: (0.2 x 30x30) cm ³																				
Reference point of PD:	Phantom surface																				
<p>T.B. 60092, Agia Paraskevi 153 10 Attiki, Tel.: +30 210 650 6765, Fax: +30 210 650 6748 e-mail : argiro.hoziai@eeae.gr, info@eeae.gr Website: http://www.eeaе.gr</p>																					

IRRADIATION CERTIFICATE FOR PERSONNAL DOSEMETERS
EEAE-IRCL-Cal. Cert. No: DOS /2353- █/19

E ERG(D)04 05
Page 2 of 2

Environmental conditions during irradiations:

Temperature	Pressure	Relative Humidity
20.0-22.5 °C	979.0-985.0 hPa	10 %

Irradiation Data

# Dosemeter	Date	Quality	<i>H_p (0.07) mSv</i>	<i>U %¹</i>
08	27/08/2019	S-Cs	5.70	4.9
12	27/08/2019	S-Cs	5.70	4.9
04	30/08/2019	W-80	4.70	4.9
05	30/08/2019	W-80	4.70	4.9
14	30/08/2019	W-80	4.70	4.9
16	30/08/2019	W-80	4.70	4.9
31	09/10/2019	W-80	53.0	4.9
32	09/10/2019	W-80	53.0	4.9
10	07/10/2019	W-80	370	4.9
11	07/10/2019	W-80	370	4.9
03	05/09/2019	W-110	4.30	4.9
07	05/09/2019	W-110	4.30	4.9
18	05/09/2019	N-30	4.80	4.9
23	05/09/2019	N-30	4.80	4.9
26	08/10/2019	N-30 60°	5.60	5.1
01	08/10/2019	N-30 60°	5.60	5.1

¹U= uncertainty 95% confidence level (k=2)

Background dosimeters: █, 13, █, 17, █, 28, █, 9.

Irradiations performed by:

Bozari A., Medical Physicist
Konstantinou P., Technician
Askounis P., Physicist

Digitally signed by ARGYRO BOZIARI
DN: o=Hellenic Public Administration
Certification Services,
serialNumber=ERMIS-25417930,
ou=ELLINIKI EPITROPI ATOMIKIS
ENERGEIAS, ou=GREEK ATOMIC
ENERGY COMMISSION, c=GR,
email=argiro.bozari@eeae.gr,
sn=BOZIARI, givenName=ARGYRO,
ou=APIYPO MTOZIAPH, cn=ARGYRO
BOZIARI
Date: 2020.02.06 10:38:55 +02'00'

Argiro Bozari
Scientific Personnel of the IRCL
Ionizing Radiation Calibration Laboratory



This certificate is issued in accordance with the requirements of ISO 17025. It provides traceability of measurements to recognized national standards laboratories. The HIRCL/GAEC is a member of the IAEA/WHO Secondary Standard Dosimetry Laboratory Network. This certificate may not be reproduced other than in full, except with the prior written approval of the HIRCL/GAEC

T.O. 60092, Agia Paraskevi 153 10 Attiki, Tel.: +30 210 650 6765, Fax: +30 210 650 6748
e-mail : argiro.bozari@eeae.gr, info@eeae.gr Website: http://www.eeae.gr

Appendix D: Example Irradiation Certificate - beta

<p>IRSN INSTITUT DE RADIOPROTECTION ET DE SÛRETÉ NUCLÉAIRE</p> <p>PSE-SANTE/SDOS/LDRI BP 17 92262 Fontenay-aux-Roses CEDEX France Tél : 01.58.35.87.62 Fax : 01.47.46.97.77</p>	<p>Test report</p> <p>N° 2019-[REDACTED]</p> <p>Subject : EURADOS comparison exercise 2019 - Beta irradiation of personnel extremity and eye lens dosimeters in terms of personnel dose equivalent</p> <p>Ordered by : EURADOS Working Group 2 "Harmonisation of individual Monitoring"</p> <p>Participant : EURADOS comparison dosimeter system Identification [REDACTED]</p> <p>Dosemeter : Personnal dosimeters</p> <p>This report includes 3 pages Date of issue: 23th February 2020 Technical Manager of calibration activities</p> <p> L. VAN-RYCKEGHEM</p> <p><i>La reproduction de ce certificat n'est autorisée que sous la forme de fac-similé photographique intégral</i></p>
--	---



INSTITUT DE RECHERCHES SUR LE NUCLEAIRE
ET LE RADIACTIVITE

1. METHOD

The irradiations are performed in terms of personal dose equivalent, Hp(0.07) or Hp(3) according to the Standard ISO 6980 [1-3]. The unity of this quantity is mSv.

The calibration equipment is a « beta Secondary Standard 2 » manufactured by ISOTRAK/PTB [4].

The radioactive sources used during this comparison exercise are $^{90}\text{Sr}/\text{Y}$ and ^{85}Kr . The traceability of these standards is established by the national standard laboratory, PTB (Physikalisch-Technische Bundesanstalt).

The reference quantity is established in terms of personal dose equivalent, Hp(0.07) or Hp(3), calculated by the BBS2 program [4] including the irradiation conditions. The angular conversion coefficients used in the program are those recommended by ISO Standard 6980-3 [3].

2. RADIATION FIELD AND IRRADIATION SET-UP

Each dosimeter irradiated is placed in the front face of the phantom according to the ISO-6980 standard. The phantom is positioned perpendicularly to the incident beta radiation field. Dosimeters are fixed on the phantom within ± 50 mm from the centre of the axis of the radiation field.

The irradiation configuration is as follows:

- Distance source - test point of the dosimeter: 30 cm. The test point corresponds to the reference point of the dosimeter defined by the participant;
- Diameter of the field and homogeneity: 15 cm in diameter with a factor of homogeneity of 1;
- Rotation axis: the rotation axis is parallel to the axis of the phantom and goes through the reference point of the dosimeter;
- Phantom for wrist dosimeter: ISO pillar phantom is a water-filled hollow cylinder with PMMA walls and an outer diameter of 73mm and 300 mm in length;
- Phantom for ring dosimeter: ISO rod phantom is cylindrical PMMA phantom with a diameter of 19 mm and 300 mm in length.
- Phantom for head dosimeter: ISO phantom is a water-filled hollow cylinder with PMMA walls with a diameter of 200 mm and a 200 mm in length.

Test report 2019- [REDACTED]



3. ENVIRONMENTAL CONDITIONS

- Temperature, T : 21 °C to 22.9°C
- Atmospheric pressure, P : 998 hPa to 1007.8 hPa
- Relative air humidity, H : 36 % to 63%

4. IRRADIATION RESULTS

The results are included in the following table:

Badge	Type	Source	Angle (")	Quantity	Dose (mSv)	Uncertainty % (k=2)	Date	Comments
2019-024	Wrist	$^{90}\text{Sr}/\text{Y}$	0	H _p (0,07)	5.02	5.7	22/08/2019	wrong irradiation
2019-022	Wrist	$^{90}\text{Sr}/\text{Y}$	0	H _p (0,07)	5.02	5.7	22/08/2019	wrong irradiation
2019-09	Ring	$^{90}\text{Sr}/\text{Y}$	0	H _p (0,07)	3	5.7	02/09/2019	
2019-19	Ring	$^{90}\text{Sr}/\text{Y}$	0	H _p (0,07)	3	5.7	02/09/2019	
2019-20	Ring	^{85}Kr	0	H _p (0,07)	4.511	5.9	05/09/2019	
2019-21	Ring	^{85}Kr	0	H _p (0,07)	4.511	5.9	05/09/2019	
2019-15	Ring	$^{90}\text{Sr}/\text{Y}$	45	H _p (0,07)	2.8	6.2	09/09/2019	
2019-25	Ring	$^{90}\text{Sr}/\text{Y}$	45	H _p (0,07)	2.8	6,2	09/09/2019	

Comments: the “wrong irradiation” corresponds to the case where a technical problem during the irradiation of the dosimeters happened. The dose is not well defined in this case.

Uncertainty:

According to the Guide to the expression of uncertainty in measurement [5], the budget uncertainty is calculated as the quadratic combination of each contribution of uncertainty with the uncertainty due to the determination of the reference quantity given by the BSS2 program. The expanded uncertainties for personal dose equivalent are calculated with a coverage factor of two.

5. REFERENCES

[1] ISO-6980-1. Reference beta-particle radiation - part 1: Methods of production (2006).

[2] ISO 6980-2. Reference beta particle radiation - part 2: Calibration fundamentals related to basic quantities characterizing the radiation fields (2005).

[3] ISO-6980-3. Reference beta particle radiation - part 3: Calibration of area and personal dosimeters and the determination of their response as a function of beta radiation energy and angle of incidence (2007).

[4] AEA Technology - ISOTRAK - Operation Manual Beta Secondary Standard 2 BBS2 (2000).

[5] JCGM 100:2008 - Evaluation of measurement data - Guide to the expression of uncertainty in measurement (2008).

Test report 2019-[REDACTED]

Appendix E: Example "Certificate of Participation"

 EURADOS European Radiation Dosimetry Group	
<i>Extremity and eye lens dosimeter Intercomparison IC2019_{exteye}</i>	
<h3>Certificate of Participation</h3>	
EURADOS Intercomparison 2019 for extremity and eye lens dosemeters (IC2019 _{exteye})	
Certificate Number:	EURADOS-2019-Sxxx/2019 for system Sxxx/2019
Number of pages:	3
Date of Issue:	February 26, 2020
Participating Institute:	xxxxxxxxxxxxxxxxxxxxxxxxxxxxxxxxxxxxxx
Dosimetry System:	xxxxxxxxxxxxxxxxxxxxxxxxxxxxxxxxxxxxxx
Reporting number:	xx (this anonymous number will be used in further publications)
Intercomparison procedure:	The EURADOS Intercomparison 2019 for extremity and eye lens dosemeters was managed and coordinated on behalf of EURADOS by the WG2 Intercomparison Organization Group (OG). The OG established the irradiation plan and announced the intercomparison, including the range limits of the doses and radiation qualities, in April 2019. The IC2019 _{exteye} on-line platform (IOP) was used by the participants for registration and for all data transfer between the participants and the OG Coordinator. Participants were required to provide details of their dosemeters (including the dosimeter reference point) on the IOP. The participants then sent their dosemeters to the Coordinator (June 2019). The Coordinator checked the correct labelling of the dosemeters and transferred all dosemeters, along with the technical details provided by the participants, to the two irradiation laboratories. The dosemeters were irradiated according to the irradiation plan and returned to the Coordinator (September and October 2019). The Coordinator then returned the dosemeters to the participants and indicated which dosemeters had not been irradiated. The participants were instructed to follow normal routine procedures as far as possible. The participants then sent the results of the dosimeter readings to the Coordinator (December 2019). After receipt of all the participants' results, the Coordinator sent the appropriate irradiation data to each participant.
Number of participants:	60 institutes from 26 countries participated in IC2019 _{exteye} with a total of 113 systems
Coordinator:	A.McWhan, W.Dobrzynska, S.Eliasik (Cavendish Nuclear Ltd., Berkeley ADS, Berkeley, GL13 9FB Gloucestershire, United Kingdom)
Intercomparison results:	See the table on page 2 of this certificate.
Irradiation data:	See the attached certificate of the irradiation laboratories No:DOS / 2353-xxx/19 and 2019-xxx.
Participant results:	See the table on page 2 of this certificate.
On behalf of the Intercomparison Organization Group:	On behalf of EURADOS:
	
Andrew McWhan Coordinator	Filip Vanhavere Chairperson



Extremity and eye lens dosimeter Intercomparison IC2019_{exteye}

Results of the Intercomparison IC2019_{exteye} (dosimetry system SXXX/2019):

EURADOS Dosemeter ID	Participant's Dosemeter ID	Radiation Quality	Quantity	Angle	Participant's Value [mSv]	Reference Value [mSv]	Response
SXXX/2019- 8	1049542	Cs-137	Hp(0.07)	0	7.27	5.7	1.275
SXXX/2019- 12	1050811	Cs-137	Hp(0.07)	0	7.28	5.7	1.277
SXXX/2019- 20	1052536	Kr-85	Hp(0.07)	0	4.34	4.511	0.962
SXXX/2019- 21	1052916	Kr-85	Hp(0.07)	0	4.04	4.511	0.896
SXXX/2019- 1	1042239	N-30	Hp(0.07)	60	7.1	5.6	1.268
SXXX/2019- 18	1051779	N-30	Hp(0.07)	0	6.33	4.8	1.319
SXXX/2019- 23	1056408	N-30	Hp(0.07)	0	5.92	4.8	1.233
SXXX/2019- 26	1056529	N-30	Hp(0.07)	60	6.86	5.6	1.225
SXXX/2019- 9	1049566	Sr-90	Hp(0.07)	0	4.49	3	1.497
SXXX/2019- 15	1051260	Sr-90	Hp(0.07)	45	3.97	2.8	1.418
SXXX/2019- 19	1052122	Sr-90	Hp(0.07)	0	4.04	3	1.347
SXXX/2019- 25	1056503	Sr-90	Hp(0.07)	45	3.98	2.8	1.421
SXXX/2019- 3	1048097	W-110	Hp(0.07)	0	5.46	4.3	1.27
SXXX/2019- 7	1049233	W-110	Hp(0.07)	0	5.74	4.3	1.335
SXXX/2019- 4	1048434	W-80	Hp(0.07)	0	7.3	4.7	1.553
SXXX/2019- 5	1048673	W-80	Hp(0.07)	0	7.31	4.7	1.555
SXXX/2019- 10	1049585	W-80	Hp(0.07)	0	580.6	370	1.569
SXXX/2019- 11	1049602	W-80	Hp(0.07)	0	617	370	1.668
SXXX/2019- 14	1051118	W-80	Hp(0.07)	0	7.16	4.7	1.523
SXXX/2019- 16	1051419	W-80	Hp(0.07)	0	7.6	4.7	1.617
SXXX/2019- 31	1057004	W-80	Hp(0.07)	0	82.17	53	1.55
SXXX/2019- 32	1057068	W-80	Hp(0.07)	0	81.42	53	1.536
SXXX/2019- 2	1044347	not irradiated					
SXXX/2019- 6	1049209	not irradiated					
SXXX/2019- 13	1050906	not irradiated					
SXXX/2019- 17	1051589	not irradiated					
SXXX/2019- 28	1056726	not irradiated					



Extremity and eye lens dosimeter Intercomparison IC2019_{exteye}

SXXX/2019-29	1056789	not irradiated
SXXX/2019-30	1056898	not irradiated
SXXX/2019-22	1056316	wrongly irradiated
SXXX/2019-24	1056431	wrongly irradiated
SXXX/2019-27	1056633	wrongly irradiated



Extremity and eye lens dosimeter Intercomparison IC2019_{exteye}

Radiation Qualities and average photon energies or maximum beta energies:

Beta radiation:

- Kr-85: 0.69 MeV (Emax) (ISO 6980)
- Sr-90/ Y-90: 2.3 MeV (Emax) (ISO 6980)

Gamma Radiation & X-rays:

- N-30 25 keV (ISO 4037)
- W-80 57 keV(80 kV) (ISO 4037)
- W-110 79 keV (ISO 4037)
- S-Cs: 662 keV (ISO 4037)

Appendix F: Datasheets with results for individual participants

In this annex all individual results are given for all participating systems, identified by their reporting number (see section 2.4), for the dose quantity $H_p(0.07)$ and $H_p(3)$. Since some systems were not designed to measure $H_p(0.07)$ these systems are missing in this part of the annex.

For the non-irradiated and wrongly irradiated dosimeters the following terms were used:

NIR not irradiated dosimeter (available for background and transport dose correction by the monitoring service)

WIR wrongly irradiated dosimeter (wrongly irradiated by the irradiation laboratory).

These results were not included in the data sheets.

Reporting number 1: Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	12	8.00	0.86	0.11	(outlier)
		13	8.00	0.77	0.10	(outlier)
	Sr-90	21	5.02	2.75	0.55	(OK)
		28	5.02	2.80	0.56	(OK)
	Sr-90/45°	15	3.00	0.89	0.30	(outlier)
photon		17	3.00	0.86	0.29	(outlier)
	N-30	18	4.60	4.21	0.92	OK
		19	4.60	3.95	0.86	OK
	N-30/60°	24	4.90	3.83	0.78	OK
		26	4.90	4.02	0.82	OK
		14	5.00	4.57	0.91	OK
	W-80-L	16	5.00	4.91	0.98	OK
		29	5.00	4.80	0.96	OK
		31	5.00	4.39	0.88	OK
	W-80-M	6	50.0	48.22	0.96	OK
		7	50.0	45.70	0.91	OK
	W-80-H	1	358	327.15	0.91	OK
		2	358	323.24	0.90	OK
	W-110	4	5.40	3.36	0.62	OK
		5	5.40	3.48	0.64	OK
	Cs-137	9	4.30	3.68	0.86	OK
		10	4.30	3.83	0.89	OK
	NIR	3				
	NIR	8				
	NIR	11				
	NIR	20				
	NIR	22				
	NIR	23				
	NIR	25				
	NIR	27				
	NIR	32				
	WIR	30				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.10	0.10	0.11	0.10	8%	-
Sr-90	2	0.55	0.55	0.56	0.55	1%	Sr-90: 45°/0°
Sr-90/45°	2	0.29	0.29	0.30	0.29	2%	0.53
beta all	6	0.29	0.32	0.56	0.10	64%	-
N-30	2	0.89	0.89	0.92	0.86	5%	N-30: 60°/0°
N-30/60°	2	0.80	0.80	0.82	0.78	3%	0.90
W-80-L	4	0.94	0.93	0.98	0.88	5%	W-80: H/L
W-80-M	2	0.94	0.94	0.96	0.91	4%	0.97
W-80-H	2	0.91	0.91	0.91	0.90	1%	
W-110	2	0.63	0.63	0.64	0.62	2%	-
Cs-137	2	0.87	0.87	0.89	0.86	3%	-
photon all	16	0.90	0.86	0.98	0.62	12%	-
All	22	0.86	0.71	0.98	0.10	40%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 2 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)		
beta	Kr-85	9	8.00	0.19	0.02	(outlier)	
		23	8.00	0.19	0.02	(outlier)	
	Sr-90	7	5.02	3.54	0.71	(OK)	
		31	5.02	3.50	0.70	(OK)	
photon	Sr-90/45°	5	3.00	1.18	0.39	(outlier)	
		8	3.00	1.35	0.45	(OK)	
	N-30	16	5.20	3.80	0.73	OK	
		17	5.20	3.83	0.74	OK	
	N-30/60°	11	4.30	2.82	0.66	OK	
		12	4.30	2.77	0.64	OK	
	W-80-L	6	5.30	4.18	0.79	OK	
		10	5.30	4.16	0.78	OK	
		22	5.30	4.38	0.83	OK	
		24	5.30	4.02	0.76	OK	
	W-80-M	28	56.0	44.57	0.80	OK	
		29	56.0	48.64	0.87	OK	
	W-80-H	2	366	235.25	0.64	outlier	
		3	366	263.82	0.72	OK	
	W-110	14	4.50	2.65	0.59	OK	
		15	4.50	2.83	0.63	OK	
	Cs-137	26	4.60	3.26	0.71	OK	
		27	4.60	3.17	0.69	OK	
	NIR	1				Legend for Quality	
	NIR	4				L low dose	
	NIR	13				M medium dose	
	NIR	18				H high dose	
	NIR	19				NIR not irradiated	
	NIR	20				WIR wrongly irradiated	
	NIR	21					
	NIR	25					
	NIR	30					
	NIR	32					

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.02	0.02	0.02	0.02	0%	-
Sr-90	2	0.70	0.70	0.71	0.70	1%	Sr-90: 45°/0°
Sr-90/45°	2	0.42	0.42	0.45	0.39	10%	0.60
beta all	6	0.42	0.38	0.71	0.02	80%	-
N-30	2	0.73	0.73	0.74	0.73	1%	N-30: 60°/0°
N-30/60°	2	0.65	0.65	0.66	0.64	1%	0.89
W-80-L	4	0.79	0.79	0.83	0.76	4%	W-80: H/L
W-80-M	2	0.83	0.83	0.87	0.80	6%	0.86
W-80-H	2	0.68	0.68	0.72	0.64	8%	
W-110	2	0.61	0.61	0.63	0.59	5%	-
Cs-137	2	0.70	0.70	0.71	0.69	2%	-
photon all	16	0.73	0.72	0.87	0.59	11%	-
All	22	0.70	0.63	0.87	0.02	36%	-

outliers: 1 of 16

fraction of outliers: 6%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 3 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	14	8.00	0.60	0.08	(outlier)
		18	8.00	1.04	0.13	(outlier)
	Sr-90	17	5.02	6.28	1.25	(OK)
		28	5.02	6.52	1.30	(OK)
	Sr-90/45°	6	3.00	1.62	0.54	(OK)
photon	N-30	31	4.30	4.98	1.16	OK
		32	4.30	4.87	1.13	OK
	N-30/60°	23	4.90	5.52	1.13	OK
		30	4.90	5.68	1.16	OK
	W-80-L	15	5.70	7.55	1.32	OK
		16	5.70	7.69	1.35	OK
		26	5.70	7.51	1.32	OK
		27	5.70	7.85	1.38	OK
	W-80-M	8	49.0	68.17	1.39	OK
		21	49.0	67.92	1.39	OK
	W-80-H	11	355	451.27	1.27	OK
		12	355	436.79	1.23	OK
	W-110	9	5.20	5.87	1.13	OK
		10	5.20	5.72	1.10	OK
	Cs-137	3	4.90	6.15	1.25	OK
		4	4.90	6.11	1.25	OK
	NIR	1				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	2				
	NIR	5				
	NIR	7				
	NIR	19				
	NIR	20				
	NIR	22				
	NIR	24				
	NIR	25				
	NIR	29				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.10	0.10	0.13	0.08	38%	-
Sr-90	2	1.28	1.28	1.30	1.25	3%	Sr-90: 45°/0°
Sr-90/45°	2	0.56	0.56	0.59	0.54	6%	0.44
beta all	6	0.56	0.65	1.30	0.08	82%	-
N-30	2	1.15	1.15	1.16	1.13	2%	N-30: 60°/0°
N-30/60°	2	1.14	1.14	1.16	1.13	2%	
W-80-L	4	1.34	1.34	1.38	1.32	2%	W-80: H/L
W-80-M	2	1.39	1.39	1.39	1.39	0%	
W-80-H	2	1.25	1.25	1.27	1.23	2%	-
W-110	2	1.11	1.11	1.13	1.10	2%	-
Cs-137	2	1.25	1.25	1.25	1.25	0%	-
photon all	16	1.25	1.25	1.39	1.10	8%	-
All	22	1.24	1.08	1.39	0.08	36%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 4 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	24	8.00	1.33	0.17	(outlier)
		32	8.00	1.47	0.18	(outlier)
	Sr-90	21	5.02	5.91	1.18	(OK)
		23	5.02	5.87	1.17	(OK)
	Sr-90/45°	17	3.00	2.59	0.86	(OK)
photon	N-30	28	4.30	5.25	1.22	OK
		29	4.30	5.22	1.21	OK
	N-30/60°	14	5.10	6.04	1.18	OK
		15	5.10	5.69	1.12	OK
	W-80-L	19	4.60	5.48	1.19	OK
		20	4.60	5.66	1.23	OK
		26	4.60	5.39	1.17	OK
		27	4.60	5.31	1.15	OK
	W-80-M	1	44.0	52.98	1.20	OK
		2	44.0	52.94	1.20	OK
	W-80-H	3	354	449.10	1.27	OK
		4	354	452.70	1.28	OK
	W-110	16	4.40	4.24	0.96	OK
		18	4.40	4.24	0.96	OK
	Cs-137	10	4.30	4.91	1.14	OK
		12	4.30	4.81	1.12	OK
	NIR	5				
	NIR	6				
	NIR	7				
	NIR	8				
	NIR	9				
	NIR	11				
	NIR	13				
	NIR	25				
	NIR	30				
	NIR	31				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.17	0.17	0.18	0.17	7%	-
Sr-90	2	1.17	1.17	1.18	1.17	1%	Sr-90: 45°/0°
Sr-90/45°	2	0.91	0.91	0.95	0.86	6%	0.77
beta all	6	0.91	0.75	1.18	0.17	62%	-
N-30	2	1.22	1.22	1.22	1.21	0%	N-30: 60°/0°
N-30/60°	2	1.15	1.15	1.18	1.12	4%	0.95
W-80-L	4	1.18	1.19	1.23	1.15	3%	W-80: H/L
W-80-M	2	1.20	1.20	1.20	1.20	0%	1.07
W-80-H	2	1.27	1.27	1.28	1.27	1%	
W-110	2	0.96	0.96	0.96	0.96	0%	-
Cs-137	2	1.13	1.13	1.14	1.12	1%	-
photon all	16	1.19	1.16	1.28	0.96	8%	-
All	22	1.17	1.05	1.28	0.17	29%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 5 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	23 27	8.00 8.00	0.54 0.61	0.07 0.08	outlier outlier
	Sr-90	24 29	5.02 5.02	3.03 3.61	0.60 0.72	OK OK
	Sr-90/45°	1 11	3.00 3.00	1.43 1.48	0.48 0.49	OK OK
	N-30	19 20	5.50 5.50	4.57 4.18	0.83 0.76	OK OK
photon	N-30/60°	31 32	5.10 5.10	3.15 3.92	0.62 0.77	OK OK
	W-80-L	5 6 9 10	4.50 4.50 4.50 4.50	3.19 3.40 3.44 3.44	0.71 0.76 0.76 0.76	OK OK OK OK
		3 4	44.0 44.0	35.05 36.86	0.80 0.84	OK OK
		18 22	357 357	212.63 192.39	0.60 0.54	outlier outlier
	W-110	7 8	4.80 4.80	2.61 3.52	0.54 0.73	OK OK
	Cs-137	16 17	5.00 5.00	3.19 3.62	0.64 0.72	OK OK
	NIR	2				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	12				
	NIR	13				
	NIR	14				
	NIR	15				
	NIR	21				
	NIR	25				
	NIR	26				
	NIR	28				
	NIR	30				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.07	0.07	0.08	0.07	9%	-
Sr-90	2	0.66	0.66	0.72	0.60	12%	Sr-90: 45°/0°
Sr-90/45°	2	0.49	0.49	0.49	0.48	2%	0.73
beta all	6	0.49	0.41	0.72	0.07	67%	-
N-30	2	0.80	0.80	0.83	0.76	6%	N-30: 60°/0°
N-30/60°	2	0.69	0.69	0.77	0.62	15%	0.87
W-80-L	4	0.76	0.75	0.76	0.71	4%	W-80: H/L
W-80-M	2	0.82	0.82	0.84	0.80	4%	0.76
W-80-H	2	0.57	0.57	0.60	0.54	7%	
W-110	2	0.64	0.64	0.73	0.54	21%	-
Cs-137	2	0.68	0.68	0.72	0.64	9%	-
photon all	16	0.74	0.71	0.84	0.54	13%	-
All	22	0.71	0.63	0.84	0.07	33%	-

outliers: 4 of 22

fraction of outliers: 18%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 6 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)		
beta	Kr-85	2	8.00	1.31	0.16	(outlier)	
		16	8.00	1.56	0.20	(outlier)	
	Sr-90	14	5.02	6.64	1.32	(OK)	
		22	5.02	6.91	1.38	(OK)	
	Sr-90/45°	25	3.00	1.91	0.64	(OK)	
		31	3.00	1.92	0.64	(OK)	
photon	N-30	27	5.50	6.95	1.26	OK	
		28	5.50	6.90	1.25	OK	
	N-30/60°	11	4.40	4.97	1.13	OK	
		12	4.40	4.80	1.09	OK	
	W-80-L	10	5.00	6.72	1.34	OK	
		13	5.00	6.68	1.34	OK	
		29	5.00	6.94	1.39	OK	
		30	5.00	6.82	1.36	OK	
	W-80-M	18	49.0	66.77	1.36	OK	
		19	49.0	66.42	1.36	OK	
	W-80-H	17	332	426.20	1.28	OK	
		21	332	428.22	1.29	OK	
	W-110	6	4.80	5.19	1.08	OK	
		8	4.80	5.38	1.12	OK	
	Cs-137	4	5.30	6.65	1.25	OK	
		5	5.30	6.76	1.27	OK	
	NIR	1				Legend for Quality	
	NIR	3				L low dose	
	NIR	7				M medium dose	
	NIR	9				H high dose	
	NIR	15				NIR not irradiated	
	NIR	20				WIR wrongly irradiated	
	NIR	23					
	NIR	24					
	NIR	26					
	NIR	32					

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)	
Kr-85	2	0.18	0.18	0.20	0.16	13%	-	
Sr-90	2	1.35	1.35	1.38	1.32	3%	Sr-90: 45°/0°	
Sr-90/45°	2	0.64	0.64	0.64	0.64	0%	0.47	
beta all	6	0.64	0.72	1.38	0.16	73%	-	
N-30	2	1.26	1.26	1.26	1.25	1%	N-30: 60°/0°	
N-30/60°	2	1.11	1.11	1.13	1.09	3%	0.88	
W-80-L	4	1.35	1.36	1.39	1.34	2%	W-80: H/L	
W-80-M	2	1.36	1.36	1.36	1.36	0%	0.95	
W-80-H	2	1.29	1.29	1.29	1.28	0%		
W-110	2	1.10	1.10	1.12	1.08	3%	-	
Cs-137	2	1.26	1.26	1.27	1.25	1%	-	
photon all	16	1.28	1.26	1.39	1.08	8%	-	
All	22	1.27	1.11	1.39	0.16	33%	-	

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 7 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	8	8.00	1.29	0.16	(outlier)
		21	8.00	1.17	0.15	(outlier)
	Sr-90	3	5.02	3.55	0.71	(OK)
		27	5.02	2.34	0.47	(outlier)
	Sr-90/45°	19	3.00	1.46	0.49	(OK)
		32	3.00	0.97	0.32	(outlier)
photon	N-30	22	4.30	4.70	1.09	OK
		24	4.30	3.98	0.92	OK
	N-30/60°	20	4.80	5.10	1.06	OK
		30	4.80	3.40	0.71	OK
	W-80-L	9	5.70	5.19	0.91	OK
		10	5.70	4.76	0.84	OK
		29	5.70	3.53	0.62	OK
		31	5.70	3.79	0.67	OK
	W-80-M	11	55.0	45.10	0.82	OK
		12	55.0	46.00	0.84	OK
	W-80-H	26	349	228.39	0.65	outlier
		28	349	243.84	0.70	outlier
	W-110	1	5.70	4.26	0.75	OK
		2	5.70	4.00	0.70	OK
	Cs-137	13	4.60	2.73	0.59	OK
		14	4.60	2.70	0.59	OK
	NIR	4				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	5				
	NIR	6				
	NIR	7				
	NIR	17				
	NIR	18				
	NIR	23				
	NIR	25				
	WIR	15				
	WIR	16				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.15	0.15	0.16	0.15	7%	-
Sr-90	2	0.59	0.59	0.71	0.47	29%	Sr-90: 45°/0°
Sr-90/45°	2	0.41	0.41	0.49	0.32	29%	0.69
beta all	6	0.40	0.38	0.71	0.15	56%	-
N-30	2	1.01	1.01	1.09	0.92	12%	N-30: 60°/0°
N-30/60°	2	0.89	0.89	1.06	0.71	28%	
W-80-L	4	0.75	0.76	0.91	0.62	18%	W-80: H/L
W-80-M	2	0.83	0.83	0.84	0.82	1%	
W-80-H	2	0.68	0.68	0.70	0.65	5%	
W-110	2	0.73	0.73	0.75	0.70	4%	-
Cs-137	2	0.59	0.59	0.59	0.59	1%	-
photon all	16	0.73	0.78	1.09	0.59	20%	-
All	22	0.70	0.67	1.09	0.15	37%	-

outliers: 2 of 16

fraction of outliers: 13%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 8 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	13 27	8.00 8.00	0.11 0.11	0.01 0.01	(outlier) (outlier)
	Sr-90	5 31	5.02 5.02	2.32 2.32	0.46 0.46	(outlier) (outlier)
	Sr-90/45°	16 30	3.00 3.00	0.58 0.69	0.19 0.23	(outlier) (outlier)
	N-30	19 20	5.50 5.50	6.06 5.97	1.10 1.09	OK OK
photon	N-30/60°	15 17	5.20 5.20	4.93 5.01	0.95 0.96	OK OK
	W-80-L	1 2 29 32	4.40 4.40 4.40 4.40	3.94 4.29 4.12 4.04	0.90 0.98 0.94 0.92	OK OK OK OK
		4 6	53.0 53.0	52.08 54.52	0.98 1.03	OK OK
		25 26	338 338	335.17 322.82	0.99 0.96	OK OK
	W-110	10 12	4.90 4.90	4.27 4.27	0.87 0.87	OK OK
	Cs-137	7 9	5.20 5.20	3.80 3.80	0.73 0.73	OK OK
	NIR	3				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	8				
	NIR	11				
	NIR	14				
	NIR	18				
	NIR	21				
	NIR	22				
	NIR	23				
	NIR	24				
	NIR	28				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.01	0.01	0.01	0.01	0%	-
Sr-90	2	0.46	0.46	0.46	0.46	0%	Sr-90: 45°/0°
Sr-90/45°	2	0.21	0.21	0.23	0.19	12%	0.46
beta all	6	0.21	0.23	0.46	0.01	88%	-
N-30	2	1.09	1.09	1.10	1.09	1%	N-30: 60°/0°
N-30/60°	2	0.96	0.96	0.96	0.95	1%	
W-80-L	4	0.93	0.93	0.98	0.90	4%	W-80: H/L
W-80-M	2	1.01	1.01	1.03	0.98	3%	
W-80-H	2	0.97	0.97	0.99	0.96	3%	
W-110	2	0.87	0.87	0.87	0.87	0%	-
Cs-137	2	0.73	0.73	0.73	0.73	0%	-
photon all	16	0.95	0.94	1.10	0.73	11%	-
All	22	0.91	0.74	1.10	0.01	47%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 9 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	30 31	8.00 8.00	3.24 3.11	0.41 0.39	outlier outlier
	Sr-90	16 20	5.02 5.02	5.88 5.82	1.17 1.16	OK OK
	Sr-90/45°	15 21	3.00 3.00	2.43 2.58	0.81 0.86	OK OK
	N-30	11 12	5.20 5.20	8.54 8.24	1.64 1.58	OK OK
photon	N-30/60°	22 23	4.40 4.40	6.40 6.75	1.45 1.53	OK OK
	W-80-L	3 5 26	4.40 4.40 4.40	5.87 5.95 5.96	1.33 1.35 1.35	OK OK OK
		27	4.40	5.94	1.35	OK
	W-80-M	4 8	52.0 52.0	73.95 72.29	1.42 1.39	OK OK
	W-80-H	17 18	329 329	473.13 476.84	1.44 1.45	OK OK
	W-110	1 2	5.00 5.00	6.00 5.84	1.20 1.17	OK OK
	Cs-137	9 10	4.50 4.50	4.66 4.64	1.04 1.03	OK OK
	NIR	6				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	7				
	NIR	13				
	NIR	14				
	NIR	19				
	NIR	24				
	NIR	25				
	NIR	28				
	NIR	29				
	NIR	32				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.40	0.40	0.41	0.39	3%	-
Sr-90	2	1.17	1.17	1.17	1.16	1%	Sr-90: 45°/0° 0.72
Sr-90/45°	2	0.84	0.84	0.86	0.81	4%	
beta all	6	0.84	0.80	1.17	0.39	43%	-
N-30	2	1.61	1.61	1.64	1.58	3%	N-30: 60°/0° 0.93
N-30/60°	2	1.49	1.49	1.53	1.45	4%	
W-80-L	4	1.35	1.35	1.35	1.33	1%	W-80: H/L 1.07
W-80-M	2	1.41	1.41	1.42	1.39	2%	
W-80-H	2	1.44	1.44	1.45	1.44	1%	
W-110	2	1.18	1.18	1.20	1.17	2%	-
Cs-137	2	1.03	1.03	1.04	1.03	0%	-
photon all	16	1.37	1.36	1.64	1.03	13%	-
All	22	1.34	1.21	1.64	0.39	28%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 10 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	5	8.00	5.16	0.65	OK
		15	8.00	4.73	0.59	OK
	Sr-90	24	5.02	3.81	0.76	OK
		30	5.02	3.77	0.75	OK
	Sr-90/45°	10	3.00	1.83	0.61	OK
photon		17	3.00	1.80	0.60	OK
	N-30	25	5.20	6.06	1.17	OK
		26	5.20	6.08	1.17	OK
	N-30/60°	31	5.00	5.39	1.08	OK
		32	5.00	5.11	1.02	OK
	W-80-L	6	4.80	5.41	1.13	OK
		7	4.80	5.37	1.12	OK
		11	4.80	5.32	1.11	OK
		12	4.80	5.37	1.12	OK
	W-80-M	8	57.0	63.63	1.12	OK
		9	57.0	63.89	1.12	OK
	W-80-H	21	380	408.94	1.08	OK
		22	380	398.57	1.05	OK
	W-110	1	4.80	4.67	0.97	OK
		2	4.80	4.75	0.99	OK
	Cs-137	13	4.40	3.66	0.83	OK
		14	4.40	3.68	0.84	OK
	NIR	3				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	4				
	NIR	16				
	NIR	18				
	NIR	19				
	NIR	20				
	NIR	23				
	NIR	27				
	NIR	28				
	NIR	29				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.62	0.62	0.65	0.59	6%	-
Sr-90	2	0.75	0.75	0.76	0.75	1%	Sr-90: 45°/0°
Sr-90/45°	2	0.61	0.61	0.61	0.60	1%	0.80
beta all	6	0.63	0.66	0.76	0.59	12%	-
N-30	2	1.17	1.17	1.17	1.17	0%	N-30: 60°/0°
N-30/60°	2	1.05	1.05	1.08	1.02	4%	0.90
W-80-L	4	1.12	1.12	1.13	1.11	1%	W-80: H/L
W-80-M	2	1.12	1.12	1.12	1.12	0%	0.95
W-80-H	2	1.06	1.06	1.08	1.05	2%	
W-110	2	0.98	0.98	0.99	0.97	1%	-
Cs-137	2	0.83	0.83	0.84	0.83	0%	-
photon all	16	1.09	1.06	1.17	0.83	10%	-
All	22	1.04	0.95	1.17	0.59	22%	-

outliers: 0 of 22

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 11 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	12 19	8.00 8.00	2.97 2.88	0.37 0.36	outlier outlier
	Sr-90	8 9	5.02 5.02	3.69 4.50	0.74 0.90	OK OK
	Sr-90/45°	15 23	3.00 3.00	2.21 1.72	0.74 0.57	OK OK
	N-30	16 17	5.20 5.20	4.82 5.06	0.93 0.97	OK OK
photon	N-30/60°	20 22	5.00 5.00	4.32 4.21	0.86 0.84	OK OK
	W-80-L	2 7 29	5.70 5.70 5.70	5.47 5.34 5.43	0.96 0.94 0.95	OK OK OK
		30	5.70	5.41	0.95	OK
	W-80-M	31 32	51.0 51.0	50.97 50.96	1.00 1.00	OK OK
	W-80-H	13 14	371 371	372.53 384.46	1.00 1.04	OK OK
	W-110	4 5	4.60 4.60	4.19 4.54	0.91 0.99	OK OK
	Cs-137	25 26	4.40 4.40	4.36 4.30	0.99 0.98	OK OK
	NIR	1				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	3				
	NIR	6				
	NIR	10				
	NIR	11				
	NIR	18				
	NIR	21				
	NIR	24				
	NIR	27				
	NIR	28				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.37	0.37	0.37	0.36	2%	-
Sr-90	2	0.82	0.82	0.90	0.74	14%	Sr-90: 45°/0°
Sr-90/45°	2	0.66	0.66	0.74	0.57	17%	0.80
beta all	6	0.65	0.61	0.90	0.36	35%	-
N-30	2	0.95	0.95	0.97	0.93	3%	N-30: 60°/0°
N-30/60°	2	0.85	0.85	0.86	0.84	2%	0.90
W-80-L	4	0.95	0.95	0.96	0.94	1%	W-80: H/L
W-80-M	2	1.00	1.00	1.00	1.00	0%	1.07
W-80-H	2	1.02	1.02	1.04	1.00	2%	
W-110	2	0.95	0.95	0.99	0.91	6%	-
Cs-137	2	0.98	0.98	0.99	0.98	1%	-
photon all	16	0.97	0.96	1.04	0.84	5%	-
All	22	0.94	0.86	1.04	0.36	23%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 12 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	26 28	8.00 8.00	1.94 1.97	0.24 0.25	outlier outlier
	Sr-90	13 32	5.02 5.02	4.55 4.13	0.91 0.82	OK OK
	Sr-90/45°	23 31	3.00 3.00	2.30 2.42	0.77 0.81	OK OK
	N-30	20 21	4.70 4.70	4.83 4.73	1.03 1.01	OK OK
	N-30/60°	6 8	4.80 4.80	5.66 3.78	1.18 0.79	OK OK
	W-80-L	5 7 27 30	4.50 4.50 4.50 4.50	3.84 4.04 4.32 3.83	0.85 0.90 0.96 0.85	OK OK OK OK
photon	W-80-M	16 18	55.0 55.0	47.67 54.98	0.87 1.00	OK OK
	W-80-H	15 17	357 357	332.02 304.00	0.93 0.85	OK OK
	W-110	2 3	4.50 4.50	4.57 4.32	1.02 0.96	OK OK
	Cs-137	9 10	5.50 5.50	5.41 5.30	0.98 0.96	OK OK
	NIR	1				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	4				
	NIR	11				
	NIR	12				
	NIR	14				
	NIR	19				
	NIR	22				
	NIR	24				
	NIR	25				
	NIR	29				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.24	0.24	0.25	0.24	1%	-
Sr-90	2	0.86	0.86	0.91	0.82	7%	Sr-90: 45°/0°
Sr-90/45°	2	0.79	0.79	0.81	0.77	4%	0.91
beta all	6	0.79	0.63	0.91	0.24	48%	-
N-30	2	1.02	1.02	1.03	1.01	1%	N-30: 60°/0°
N-30/60°	2	0.98	0.98	1.18	0.79	28%	
W-80-L	4	0.88	0.89	0.96	0.85	6%	W-80: H/L
W-80-M	2	0.93	0.93	1.00	0.87	10%	
W-80-H	2	0.89	0.89	0.93	0.85	6%	
W-110	2	0.99	0.99	1.02	0.96	4%	-
Cs-137	2	0.97	0.97	0.98	0.96	1%	-
photon all	16	0.96	0.95	1.18	0.79	10%	-
All	22	0.90	0.86	1.18	0.24	26%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 13 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	20 29	8.00 8.00	2.20 2.54	0.28 0.32	outlier outlier
	Sr-90	12 14	5.02 5.02	4.58 4.54	0.91 0.90	OK OK
	Sr-90/45°	16 27	3.00 3.00	1.82 1.73	0.61 0.58	OK OK
	N-30	10 11	4.50 4.50	4.52 4.40	1.00 0.98	OK OK
	N-30/60°	17 18	5.30 5.30	4.28 3.64	0.81 0.69	OK OK
photon	W-80-L	1 2 19 21	4.90 4.90 4.90 4.90	4.11 4.34 4.24 4.26	0.84 0.89 0.87 0.87	OK OK OK OK
		28 30	46.0 46.0	39.32 43.45	0.85 0.94	OK OK
		8 9	316 316	296.01 302.33	0.94 0.96	OK OK
		23 24	5.10 5.10	3.96 3.73	0.78 0.73	OK OK
	W-110	6 25	5.70 5.70	5.49 5.37	0.96 0.94	OK OK
	Cs-137	3 7 13 15 22 26 31 32 4 5				
						Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.30	0.30	0.32	0.28	10%	-
Sr-90	2	0.91	0.91	0.91	0.90	1%	Sr-90: 45°/0°
Sr-90/45°	2	0.59	0.59	0.61	0.58	4%	0.65
beta all	6	0.59	0.60	0.91	0.28	46%	-
N-30	2	0.99	0.99	1.00	0.98	2%	N-30: 60°/0°
N-30/60°	2	0.75	0.75	0.81	0.69	11%	0.75
W-80-L	4	0.87	0.86	0.89	0.84	2%	W-80: H/L
W-80-M	2	0.90	0.90	0.94	0.85	7%	1.09
W-80-H	2	0.95	0.95	0.96	0.94	1%	
W-110	2	0.75	0.75	0.78	0.73	4%	-
Cs-137	2	0.95	0.95	0.96	0.94	2%	-
photon all	16	0.88	0.88	1.00	0.69	10%	-
All	22	0.87	0.80	1.00	0.28	25%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 14 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	5 23	8.00 8.00	0.80 0.62	0.10 0.08	outlier outlier
	Sr-90	2 19	5.02 5.02	6.59 6.98	1.31 1.39	OK OK
	Sr-90/45°	12 32	3.00 3.00	2.30 2.23	0.77 0.74	OK OK
	N-30	18 20	5.20 5.20	4.50 4.73	0.87 0.91	OK OK
photon	N-30/60°	26 27	4.60 4.60	3.68 3.96	0.80 0.86	OK OK
		3 6 16 17	4.30 4.30 4.30 4.30	4.33 4.42 4.26 4.35	1.01 1.03 0.99 1.01	OK OK OK OK
	W-80-L	21 22	55.0 55.0	55.72 58.22	1.01 1.06	OK OK
	W-80-M	28 29	356 356	376.02 360.02	1.06 1.01	OK OK
	W-80-H	9 10	5.20 5.20	5.19 5.05	1.00 0.97	OK OK
	W-110	7 8	5.50 5.50	7.10 7.33	1.29 1.33	OK OK
	Cs-137	1 4 11 13 14 15 24 25 30 31				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.09	0.09	0.10	0.08	18%	-
Sr-90	2	1.35	1.35	1.39	1.31	4%	Sr-90: 45°/0°
Sr-90/45°	2	0.76	0.76	0.77	0.74	2%	0.56
beta all	6	0.76	0.73	1.39	0.08	77%	-
N-30	2	0.89	0.89	0.91	0.87	4%	N-30: 60°/0°
N-30/60°	2	0.83	0.83	0.86	0.80	5%	0.94
W-80-L	4	1.01	1.01	1.03	0.99	2%	W-80: H/L
W-80-M	2	1.04	1.04	1.06	1.01	3%	1.02
W-80-H	2	1.03	1.03	1.06	1.01	3%	
W-110	2	0.98	0.98	1.00	0.97	2%	-
Cs-137	2	1.31	1.31	1.33	1.29	2%	-
photon all	16	1.01	1.01	1.33	0.80	14%	-
All	22	1.00	0.94	1.39	0.08	35%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 15 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	20 26	8.00 8.00	1.41 1.56	0.18 0.20	outlier outlier
	Sr-90	6 12	5.02 5.02	5.64 5.66	1.12 1.13	OK OK
	Sr-90/45°	17 18	3.00 3.00	2.49 2.52	0.83 0.84	OK OK
	N-30	23 24	5.00 5.00	3.74 3.76	0.75 0.75	OK OK
photon	N-30/60°	8 9	5.50 5.50	3.75 3.86	0.68 0.70	OK OK
	W-80-L	3 4 10 13	5.40 5.40 5.40 5.40	4.27 4.18 4.35 4.44	0.79 0.77 0.81 0.82	OK OK OK OK
		15 16	49.0 49.0	38.82 38.22	0.79 0.78	OK OK
		21 22	378 378	304.72 307.72	0.81 0.81	OK OK
	W-110	5 7	4.70 4.70	3.76 3.85	0.80 0.82	OK OK
	Cs-137	27 29	4.80 4.80	4.55 4.50	0.95 0.94	OK OK
	NIR	1				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	2				
	NIR	11				
	NIR	14				
	NIR	19				
	NIR	25				
	NIR	28				
	NIR	30				
	NIR	31				
	NIR	32				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.19	0.19	0.20	0.18	7%	-
Sr-90	2	1.13	1.13	1.13	1.12	0%	Sr-90: 45°/0° 0.74
Sr-90/45°	2	0.84	0.84	0.84	0.83	1%	
beta all	6	0.84	0.72	1.13	0.18	60%	-
N-30	2	0.75	0.75	0.75	0.75	0%	N-30: 60°/0° 0.92
N-30/60°	2	0.69	0.69	0.70	0.68	2%	
W-80-L	4	0.80	0.80	0.82	0.77	3%	W-80: H/L 1.01
W-80-M	2	0.79	0.79	0.79	0.78	1%	
W-80-H	2	0.81	0.81	0.81	0.81	1%	
W-110	2	0.81	0.81	0.82	0.80	2%	-
Cs-137	2	0.94	0.94	0.95	0.94	1%	-
photon all	16	0.80	0.80	0.95	0.68	9%	-
All	22	0.80	0.78	1.13	0.18	29%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 16 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	17 27	8.00 8.00	7.90 8.17	0.99 1.02	OK OK
	Sr-90	7 9	5.02 5.02	4.98 4.99	0.99 0.99	OK OK
	Sr-90/45°	2 14	3.00 3.00	2.65 2.73	0.88 0.91	OK OK
	N-30	21 22	5.30 5.30	4.80 4.93	0.91 0.93	OK OK
photon	N-30/60°	31 32	4.40 4.40	4.47 4.54	1.02 1.03	OK OK
	W-80-L	8 10 19 20	4.50 4.50 4.50 4.50	4.36 4.43 4.38 4.37	0.97 0.98 0.97 0.97	OK OK OK OK
		23 25	53.0 53.0	53.55 52.79	1.01 1.00	OK OK
		5 6	333 333	326.49 335.06	0.98 1.01	OK OK
	W-110	3 4	5.70 5.70	4.89 4.81	0.86 0.84	OK OK
	Cs-137	28 29	4.60 4.60	4.58 4.59	1.00 1.00	OK OK
	NIR	1				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	11				
	NIR	12				
	NIR	13				
	NIR	15				
	NIR	16				
	NIR	18				
	NIR	24				
	NIR	26				
	NIR	30				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	1.00	1.00	1.02	0.99	2%	-
Sr-90	2	0.99	0.99	0.99	0.99	0%	Sr-90: 45°/0°
Sr-90/45°	2	0.90	0.90	0.91	0.88	2%	0.90
beta all	6	0.99	0.96	1.02	0.88	6%	-
N-30	2	0.92	0.92	0.93	0.91	2%	N-30: 60°/0°
N-30/60°	2	1.02	1.02	1.03	1.02	1%	1.12
W-80-L	4	0.97	0.97	0.98	0.97	1%	W-80: H/L
W-80-M	2	1.00	1.00	1.01	1.00	1%	1.02
W-80-H	2	0.99	0.99	1.01	0.98	2%	
W-110	2	0.85	0.85	0.86	0.84	1%	-
Cs-137	2	1.00	1.00	1.00	1.00	0%	-
photon all	16	0.98	0.97	1.03	0.84	6%	-
All	22	0.99	0.97	1.03	0.84	6%	-

outliers: 0 of 22

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 17 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)		
beta	Kr-85	1	4.51	0.41	0.09	(outlier)	
		17	4.51	0.62	0.14	(outlier)	
	Sr-90	9	3.00	1.86	0.62	(OK)	
		25	3.00	2.49	0.83	(OK)	
	Sr-90/45°	10	2.80	0.99	0.35	(outlier)	
		31	2.80	1.16	0.41	(OK)	
photon	N-30	13	4.40	2.40	0.55	OK	
		14	4.40	2.73	0.62	OK	
	N-30/60°	15	5.20	2.60	0.50	outlier	
		18	5.20	2.89	0.56	OK	
	W-80-L	16	4.40	1.65	0.38	outlier	
		19	4.40	2.46	0.56	OK	
		27	4.40	2.08	0.47	outlier	
		29	4.40	2.13	0.48	outlier	
	W-80-M	24	43.0	22.68	0.53	outlier	
		28	43.0	24.83	0.58	outlier	
	W-80-H	30	340	245.32	0.72	OK	
		32	340	201.50	0.59	outlier	
	W-110	6	4.70	2.74	0.58	OK	
		7	4.70	2.81	0.60	OK	
	Cs-137	3	5.50	3.92	0.71	OK	
		4	5.50	4.69	0.85	OK	
	NIR	2				Legend for Quality	
	NIR	5				L low dose	
	NIR	8				M medium dose	
	NIR	11				H high dose	
	NIR	12				NIR not irradiated	
	NIR	20				WIR wrongly irradiated	
	NIR	21					
	NIR	22					
	NIR	23					
	NIR	26					

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.11	0.11	0.14	0.09	29%	-
Sr-90	2	0.73	0.73	0.83	0.62	20%	Sr-90: 45°/0°
Sr-90/45°	2	0.38	0.38	0.41	0.35	11%	0.53
beta all	6	0.38	0.41	0.83	0.09	70%	-
N-30	2	0.58	0.58	0.62	0.55	9%	N-30: 60°/0°
N-30/60°	2	0.53	0.53	0.56	0.50	7%	0.91
W-80-L	4	0.48	0.47	0.56	0.38	16%	W-80: H/L
W-80-M	2	0.55	0.55	0.58	0.53	6%	1.39
W-80-H	2	0.66	0.66	0.72	0.59	14%	
W-110	2	0.59	0.59	0.60	0.58	2%	-
Cs-137	2	0.78	0.78	0.85	0.71	13%	-
photon all	16	0.57	0.58	0.85	0.38	19%	-
All	22	0.56	0.53	0.85	0.09	35%	-

outliers: 7 of 16

fraction of outliers: 44%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 18 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	15 29	4.51 4.51	1.33 1.25	0.29 0.28	outlier outlier
	Sr-90	19 26	3.00 3.00	4.32 4.01	1.44 1.34	OK OK
	Sr-90/45°	2 7	2.80 2.80	3.14 3.28	1.12 1.17	OK OK
	N-30	1 4	5.50 5.50	5.06 4.99	0.92 0.91	OK OK
photon	N-30/60°	13 14	5.00 5.00	4.94 4.89	0.99 0.98	OK OK
	W-80-L	16 20 21 22	5.20 5.20 5.20 5.20	5.14 4.61 4.66 5.18	0.99 0.89 0.90 1.00	OK OK OK OK
		17 18	52.0 52.0	45.91 48.75	0.88 0.94	OK OK
		24 25	373 373	379.21 345.18	1.02 0.93	OK OK
	W-110	8 9	5.30 5.30	4.47 4.73	0.84 0.89	OK OK
	Cs-137	10 11	4.40 4.40	4.82 5.38	1.09 1.22	OK OK
	NIR NIR NIR NIR NIR NIR NIR NIR NIR NIR	3 5 6 12 23 27 28 30 31 32				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.29	0.29	0.29	0.28	4%	-
Sr-90	2	1.39	1.39	1.44	1.34	5%	Sr-90: 45°/0°
Sr-90/45°	2	1.15	1.15	1.17	1.12	3%	0.83
beta all	6	1.15	0.94	1.44	0.28	55%	-
N-30	2	0.91	0.91	0.92	0.91	1%	N-30: 60°/0°
N-30/60°	2	0.98	0.98	0.99	0.98	1%	1.08
W-80-L	4	0.94	0.94	1.00	0.89	6%	W-80: H/L
W-80-M	2	0.91	0.91	0.94	0.88	4%	1.03
W-80-H	2	0.97	0.97	1.02	0.93	7%	
W-110	2	0.87	0.87	0.89	0.84	4%	-
Cs-137	2	1.16	1.16	1.22	1.09	8%	-
photon all	16	0.93	0.96	1.22	0.84	10%	-
All	22	0.96	0.96	1.44	0.28	28%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 19 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	9	4.51	0.30	0.07	(outlier)
		23	4.51	0.35	0.08	(outlier)
	Sr-90	16	3.00	3.51	1.17	(OK)
		17	3.00	3.38	1.13	(OK)
	Sr-90/45°	5	2.80	2.23	0.79	(OK)
		8	2.80	2.40	0.86	(OK)
	N-30	18	4.70	4.44	0.94	OK
		19	4.70	4.65	0.99	OK
photon	N-30/60°	22	5.70	5.65	0.99	OK
		26	5.70	6.01	1.05	OK
	W-80-L	2	5.50	4.47	0.81	OK
		4	5.50	5.44	0.99	OK
		27	5.50	6.03	1.10	OK
		30	5.50	5.68	1.03	OK
	W-80-M	3	49.0	51.59	1.05	OK
		6	49.0	63.91	1.30	OK
	W-80-H	12	371	438.41	1.18	OK
		15	371	390.22	1.05	OK
	W-110	28	5.30	5.47	1.03	OK
		29	5.30	6.54	1.23	OK
	Cs-137	7	4.70	5.89	1.25	OK
		10	4.70	7.20	1.53	OK
	NIR	1				
	NIR	11				
	NIR	13				
	NIR	14				
	NIR	20				
	NIR	21				
	NIR	24				
	NIR	25				
	NIR	31				
	NIR	32				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.07	0.07	0.08	0.07	11%	-
Sr-90	2	1.15	1.15	1.17	1.13	3%	Sr-90: 45°/0°
Sr-90/45°	2	0.83	0.83	0.86	0.79	5%	0.72
beta all	6	0.83	0.68	1.17	0.07	73%	-
N-30	2	0.97	0.97	0.99	0.94	3%	N-30: 60°/0°
N-30/60°	2	1.02	1.02	1.05	0.99	4%	1.06
W-80-L	4	1.01	0.98	1.10	0.81	12%	W-80: H/L
W-80-M	2	1.18	1.18	1.30	1.05	15%	1.14
W-80-H	2	1.12	1.12	1.18	1.05	8%	
W-110	2	1.13	1.13	1.23	1.03	13%	-
Cs-137	2	1.39	1.39	1.53	1.25	14%	-
photon all	16	1.05	1.10	1.53	0.81	16%	-
All	22	1.04	0.98	1.53	0.07	34%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 20 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	16	4.51	0.06	0.01	(outlier)
		24	4.51	0.43	0.10	(outlier)
	Sr-90	10	3.00	3.92	1.31	(OK)
		22	3.00	3.74	1.25	(OK)
	Sr-90/45°	4	2.80	1.21	0.43	(OK)
photon	N-30	17	5.70	5.36	0.94	OK
		19	5.70	4.38	0.77	OK
	N-30/60°	26	4.70	5.80	1.23	OK
		27	4.70	4.23	0.90	OK
	W-80-L	11	4.40	4.53	1.03	OK
		13	4.40	3.70	0.84	OK
		31	4.40	3.79	0.86	OK
		32	4.40	5.23	1.19	OK
	W-80-M	14	46.0	48.38	1.05	OK
		15	46.0	54.94	1.19	OK
	W-80-H	1	336	417.81	1.24	OK
		2	336	328.70	0.98	OK
	W-110	21	4.60	3.63	0.79	OK
		28	4.60	4.21	0.91	OK
	Cs-137	3	4.50	4.80	1.07	OK
		5	4.50	6.10	1.35	OK
		6				
	NIR	7				
	NIR	8				
	NIR	9				
	NIR	12				
	NIR	18				
	NIR	20				
	NIR	23				
	NIR	29				
	NIR	30				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.06	0.06	0.10	0.01	105%	-
Sr-90	2	1.28	1.28	1.31	1.25	3%	Sr-90: 45°/0°
Sr-90/45°	2	0.50	0.50	0.57	0.43	20%	0.39
beta all	6	0.50	0.61	1.31	0.01	91%	-
N-30	2	0.85	0.85	0.94	0.77	14%	N-30: 60°/0°
N-30/60°	2	1.07	1.07	1.23	0.90	22%	1.25
W-80-L	4	0.95	0.98	1.19	0.84	17%	W-80: H/L
W-80-M	2	1.12	1.12	1.19	1.05	9%	1.13
W-80-H	2	1.11	1.11	1.24	0.98	17%	
W-110	2	0.85	0.85	0.91	0.79	11%	-
Cs-137	2	1.21	1.21	1.35	1.07	17%	-
photon all	16	1.00	1.02	1.35	0.77	17%	-
All	22	0.96	0.91	1.35	0.01	40%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 21: Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)		
beta	Kr-85	12	4.51	0.32	0.07	(outlier)	
		32	4.51	0.16	0.04	(outlier)	
	Sr-90	25	3.00	2.17	0.72	(OK)	
		31	3.00	4.13	1.38	(OK)	
	Sr-90/45°	11	2.80	2.63	0.94	(OK)	
photon		15	2.80	2.26	0.81	(OK)	
	N-30	28	5.50	4.72	0.86	OK	
		29	5.50	7.67	1.39	OK	
	N-30/60°	1	4.50	6.09	1.35	OK	
		5	4.50	5.37	1.19	OK	
		7	4.60	4.94	1.07	OK	
	W-80-L	9	4.60	4.44	0.97	OK	
		14	4.60	3.27	0.71	OK	
		17	4.60	4.59	1.00	OK	
	W-80-M	26	51.0	56.78	1.11	OK	
		27	51.0	61.33	1.20	OK	
	W-80-H	22	333	203.03	0.61	outlier	
		23	333	314.14	0.94	OK	
	W-110	16	4.40	4.24	0.96	OK	
		18	4.40	4.68	1.06	OK	
	Cs-137	3	5.40	5.58	1.03	OK	
		4	5.40	4.62	0.85	OK	
	NIR	2				Legend for Quality	
	NIR	6				L low dose	
	NIR	8				M medium dose	
	NIR	10				H high dose	
	NIR	13				NIR not irradiated	
	NIR	19				WIR wrongly irradiated	
	NIR	20					
	NIR	21					
	NIR	24					
	NIR	30					

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)	
Kr-85	2	0.05	0.05	0.07	0.04	46%	-	
Sr-90	2	1.05	1.05	1.38	0.72	44%	Sr-90: 45°/0°	
Sr-90/45°	2	0.87	0.87	0.94	0.81	11%	0.83	
beta all	6	0.76	0.66	1.38	0.04	79%	-	
N-30	2	1.13	1.13	1.39	0.86	34%	N-30: 60°/0°	
N-30/60°	2	1.27	1.27	1.35	1.19	9%	1.13	
W-80-L	4	0.98	0.94	1.07	0.71	17%	W-80: H/L	
W-80-M	2	1.16	1.16	1.20	1.11	5%	0.83	
W-80-H	2	0.78	0.78	0.94	0.61	30%		
W-110	2	1.01	1.01	1.06	0.96	7%	-	
Cs-137	2	0.94	0.94	1.03	0.85	13%	-	
photon all	16	1.02	1.02	1.39	0.61	21%	-	
All	22	0.96	0.92	1.39	0.04	38%	-	

outliers: 1 of 16

fraction of outliers: 6%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 22 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	5 23	4.51 4.51	2.88 3.24	0.64 0.72	OK OK
	Sr-90	2 20	3.00 3.00	4.12 4.27	1.37 1.42	OK OK
	Sr-90/45°	4 21	2.80 2.80	2.38 2.23	0.85 0.80	OK OK
	N-30	11 12	5.70 5.70	6.28 6.68	1.10 1.17	OK OK
photon	N-30/60°	13 14	4.70 4.70	5.33 5.17	1.13 1.10	OK OK
	W-80-L	18 24 28	5.30 5.30 5.30	6.72 6.66 6.25	1.27 1.26 1.18	OK OK OK
		31	5.30	6.75	1.27	OK
	W-80-M	3 7	47.0 47.0	56.31 54.55	1.20 1.16	OK OK
	W-80-H	29	336	392.95	1.17	OK
		30	336	408.27	1.22	OK
	W-110	8 9	5.40 5.40	5.15 5.14	0.95 0.95	OK OK
	Cs-137	15 16	5.40 5.40	6.02 5.90	1.11 1.09	OK OK
	NIR	1				
	NIR	6				
	NIR	10				
	NIR	17				
	NIR	19				
	NIR	22				
	NIR	25				
	NIR	26				
	NIR	27				
	NIR	32				


Legend for Quality

- L low dose
- M medium dose
- H high dose
- NIR not irradiated
- WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.68	0.68	0.72	0.64	8%	-
Sr-90	2	1.40	1.40	1.42	1.37	3%	Sr-90: 45°/0°
Sr-90/45°	2	0.82	0.82	0.85	0.80	5%	0.59
beta all	6	0.82	0.97	1.42	0.64	35%	-
N-30	2	1.14	1.14	1.17	1.10	4%	N-30: 60°/0°
N-30/60°	2	1.12	1.12	1.13	1.10	2%	0.98
W-80-L	4	1.26	1.24	1.27	1.18	4%	W-80: H/L
W-80-M	2	1.18	1.18	1.20	1.16	2%	0.96
W-80-H	2	1.19	1.19	1.22	1.17	3%	
W-110	2	0.95	0.95	0.95	0.95	0%	-
Cs-137	2	1.10	1.10	1.11	1.09	1%	-
photon all	16	1.17	1.15	1.27	0.95	8%	-
All	22	1.15	1.10	1.42	0.64	18%	-

outliers: 0 of 22

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 23 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	6	4.51	4.91	1.09	OK
		8	4.51	4.53	1.00	OK
	Sr-90	13	3.00	4.16	1.39	OK
		30	3.00	4.22	1.41	OK
photon	Sr-90/45°	14	2.80	3.20	1.14	OK
		28	2.80	3.38	1.21	OK
	N-30	31	5.60	7.06	1.26	OK
		32	5.60	7.11	1.27	OK
	N-30/60°	16	5.60	7.58	1.35	OK
		17	5.60	7.45	1.33	OK
	W-80-L	4	5.10	6.11	1.20	OK
		5	5.10	6.37	1.25	OK
		18	5.10	6.41	1.26	OK
		22	5.10	6.62	1.30	OK
	W-80-M	12	48.0	56.68	1.18	OK
		15	48.0	62.04	1.29	OK
	W-80-H	21	356	475.97	1.34	OK
		23	356	455.66	1.28	OK
	W-110	25	5.70	6.21	1.09	OK
		26	5.70	6.10	1.07	OK
	Cs-137	9	5.10	5.53	1.08	OK
		11	5.10	5.74	1.13	OK
	NIR	1				
	NIR	2				
	NIR	3				
	NIR	7				
	NIR	10				
	NIR	19				
	NIR	20				
	NIR	24				
	NIR	27				
	NIR	29				


Legend for Quality

- L low dose
- M medium dose
- H high dose
- NIR not irradiated
- WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	1.05	1.05	1.09	1.00	6%	-
Sr-90	2	1.40	1.40	1.41	1.39	1%	Sr-90: 45°/0°
Sr-90/45°	2	1.18	1.18	1.21	1.14	4%	0.84
beta all	6	1.18	1.21	1.41	1.00	13%	-
N-30	2	1.27	1.27	1.27	1.26	0%	N-30: 60°/0°
N-30/60°	2	1.34	1.34	1.35	1.33	1%	1.06
W-80-L	4	1.25	1.25	1.30	1.20	3%	W-80: H/L
W-80-M	2	1.24	1.24	1.29	1.18	6%	1.05
W-80-H	2	1.31	1.31	1.34	1.28	3%	
W-110	2	1.08	1.08	1.09	1.07	1%	-
Cs-137	2	1.10	1.10	1.13	1.08	3%	-
photon all	16	1.26	1.23	1.35	1.07	8%	-
All	22	1.25	1.22	1.41	1.00	9%	-

outliers: 0 of 22

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 24 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	16	4.51	2.70	0.60	(OK)
		27	4.51	3.10	0.69	(OK)
	Sr-90	4	3.00	4.29	1.43	(OK)
		19	3.00	4.06	1.35	(OK)
	Sr-90/45°	18	2.80	1.55	0.55	(OK)
photon		28	2.80	1.51	0.54	(OK)
	N-30	7	4.90	5.76	1.18	OK
		11	4.90	5.81	1.18	OK
	N-30/60°	12	4.80	5.72	1.19	OK
		13	4.80	5.81	1.21	OK
	W-80-L	9	4.40	5.97	1.36	OK
		14	4.40	5.77	1.31	OK
		23	4.40	5.89	1.34	OK
		29	4.40	5.94	1.35	OK
	W-80-M	24	53.0	71.33	1.35	OK
		25	53.0	72.25	1.36	OK
	W-80-H	8	378	426.31	1.13	OK
		10	378	432.76	1.14	OK
	W-110	31	4.30	4.65	1.08	OK
		32	4.30	4.65	1.08	OK
	Cs-137	2	5.00	5.82	1.16	OK
		3	5.00	5.82	1.16	OK
	NIR	1				
	NIR	5				
	NIR	6				
	NIR	15				
	NIR	17				
	NIR	20				
	NIR	21				
	NIR	22				
	NIR	26				
	NIR	30				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.64	0.64	0.69	0.60	10%	-
Sr-90	2	1.39	1.39	1.43	1.35	4%	Sr-90: 45°/0°
Sr-90/45°	2	0.55	0.55	0.55	0.54	2%	0.39
beta all	6	0.64	0.86	1.43	0.54	48%	-
N-30	2	1.18	1.18	1.18	1.18	1%	N-30: 60°/0°
N-30/60°	2	1.20	1.20	1.21	1.19	1%	1.02
W-80-L	4	1.34	1.34	1.36	1.31	2%	W-80: H/L
W-80-M	2	1.35	1.35	1.36	1.35	1%	0.85
W-80-H	2	1.14	1.14	1.14	1.13	1%	
W-110	2	1.08	1.08	1.08	1.08	0%	-
Cs-137	2	1.16	1.16	1.16	1.16	0%	-
photon all	16	1.19	1.22	1.36	1.08	8%	-
All	22	1.18	1.12	1.43	0.54	24%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 25 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	14 21	4.51 4.51	1.12 1.17	0.25 0.26	outlier outlier
	Sr-90	4 28	3.00 3.00	1.04 2.28	0.35 0.76	outlier OK
	Sr-90/45°	20 26	2.80 2.80	1.98 2.09	0.71 0.75	OK OK
	N-30	29 31	5.50 5.50	4.50 4.13	0.82 0.75	OK OK
	N-30/60°	30 32	4.30 4.30	3.63 3.28	0.84 0.76	OK OK
photon	W-80-L	6 8 10 16	4.90 4.90 4.90 4.90	3.11 3.86 3.74 3.70	0.63 0.79 0.76 0.76	OK OK OK OK
		17 19	46.0 46.0	35.16 35.25	0.76 0.77	OK OK
		22 23	362 362	278.60 253.50	0.77 0.70	OK outlier
		1 3	4.70 4.70	2.54 2.48	0.54 0.53	OK OK
	W-110	15 18	4.90 4.90	3.16 3.49	0.64 0.71	OK OK
	Cs-137	2 5 7 9 11 12 13 24 25 27	2 5 7 9 11 12 13 24 25 27			Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.25	0.25	0.26	0.25	3%	-
Sr-90	2	0.55	0.55	0.76	0.35	53%	Sr-90: 45°/0°
Sr-90/45°	2	0.73	0.73	0.75	0.71	4%	1.31
beta all	6	0.53	0.51	0.76	0.25	49%	-
N-30	2	0.78	0.78	0.82	0.75	6%	N-30: 60°/0°
N-30/60°	2	0.80	0.80	0.84	0.76	7%	1.02
W-80-L	4	0.76	0.74	0.79	0.63	9%	W-80: H/L
W-80-M	2	0.77	0.77	0.77	0.76	0%	1.00
W-80-H	2	0.73	0.73	0.77	0.70	7%	
W-110	2	0.53	0.53	0.54	0.53	2%	-
Cs-137	2	0.68	0.68	0.71	0.64	7%	-
photon all	16	0.76	0.72	0.84	0.53	13%	-
All	22	0.75	0.66	0.84	0.25	26%	-

outliers: 4 of 22

fraction of outliers: 18%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 26 : Extremity - photon dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	11 15	4.51 4.51	1.07 1.17	0.24 0.26	(outlier) (outlier)
	Sr-90	5 21	3.00 3.00	3.14 3.14	1.05 1.05	(OK) (OK)
	Sr-90/45°	18 22	2.80 2.80	1.82 1.76	0.65 0.63	(OK) (OK)
	N-30	6 16	5.20 5.20	5.13 4.95	0.99 0.95	OK OK
	N-30/60°	14 20	5.40 5.40	5.31 5.13	0.98 0.95	OK OK
photon	W-80-L	17 19 28 29	5.20 5.20 5.20 5.20	5.67 5.70 5.92 5.90	1.09 1.10 1.14 1.13	OK OK OK OK
		24 27	54.0 54.0	62.13 62.37	1.15 1.16	OK OK
		1 2	317 317	354.20 383.04	1.12 1.21	OK OK
		12 13	5.50 5.50	4.79 5.06	0.87 0.92	OK OK
	W-110	8 9	5.50 5.50	5.03 4.92	0.91 0.89	OK OK
	Cs-137	3 4 7 10 23 25 26 30 31 32				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
						beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.25	0.25	0.26	0.24	6%	-
Sr-90	2	1.05	1.05	1.05	1.05	0%	Sr-90: 45°/0°
Sr-90/45°	2	0.64	0.64	0.65	0.63	2%	0.61
beta all	6	0.64	0.64	1.05	0.24	55%	-
N-30	2	0.97	0.97	0.99	0.95	3%	N-30: 60°/0°
N-30/60°	2	0.97	0.97	0.98	0.95	2%	1.00
W-80-L	4	1.12	1.11	1.14	1.09	2%	W-80: H/L
W-80-M	2	1.15	1.15	1.16	1.15	0%	1.04
W-80-H	2	1.16	1.16	1.21	1.12	6%	
W-110	2	0.90	0.90	0.92	0.87	4%	-
Cs-137	2	0.90	0.90	0.91	0.89	2%	-
photon all	16	1.04	1.04	1.21	0.87	11%	-
All	22	0.98	0.93	1.21	0.24	29%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 27 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	2	4.51	1.40	0.31	outlier
		20	4.51	1.35	0.30	outlier
	Sr-90	16	3.00	2.65	0.88	OK
		17	3.00	3.60	1.20	OK
	Sr-90/45°	3	2.80	2.70	0.96	OK
photon		13	2.80	2.90	1.04	OK
	N-30	8	4.40	5.15	1.17	OK
		9	4.40	5.30	1.20	OK
	N-30/60°	29	5.30	1.95	0.37	outlier
		30	5.30	6.30	1.19	OK
		7	5.10	7.05	1.38	OK
	W-80-L	11	5.10	6.65	1.30	OK
		14	5.10	6.95	1.36	OK
		15	5.10	6.05	1.19	OK
	W-80-M	19	53.0	73.00	1.38	OK
		24	53.0	73.00	1.38	OK
	W-80-H	22	358	600.00	1.68	outlier
		23	358	600.00	1.68	outlier
	W-110	21	4.50	4.80	1.07	OK
		26	4.50	4.50	1.00	OK
	Cs-137	1	5.50	6.40	1.16	OK
		4	5.50	4.70	0.85	OK
	NIR	5				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	6				
	NIR	10				
	NIR	12				
	NIR	18				
	NIR	25				
	NIR	27				
	NIR	28				
	NIR	31				
	NIR	32				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.30	0.30	0.31	0.30	3%	-
Sr-90	2	1.04	1.04	1.20	0.88	21%	Sr-90: 45°/0°
Sr-90/45°	2	1.00	1.00	1.04	0.96	5%	0.96
beta all	6	0.92	0.78	1.20	0.30	49%	-
N-30	2	1.19	1.19	1.20	1.17	2%	N-30: 60°/0°
N-30/60°	2	0.78	0.78	1.19	0.37	75%	
W-80-L	4	1.33	1.31	1.38	1.19	7%	W-80: H/L
W-80-M	2	1.38	1.38	1.38	1.38	0%	
W-80-H	2	1.68	1.68	1.68	1.68	0%	1.28
W-110	2	1.03	1.03	1.07	1.00	5%	-
Cs-137	2	1.01	1.01	1.16	0.85	22%	-
photon all	16	1.20	1.21	1.68	0.37	26%	-
All	22	1.18	1.09	1.68	0.30	35%	-

outliers: 5 of 22

fraction of outliers: 23%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 28 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	8 9	4.51 4.51	3.45 3.10	0.76 0.69	OK OK
	Sr-90	10 24	3.00 3.00	2.65 2.65	0.88 0.88	OK OK
	Sr-90/45°	3 23	2.80 2.80	2.45 2.85	0.88 1.02	OK OK
	N-30	30 31	4.60 4.60	5.25 4.10	1.14 0.89	OK OK
	N-30/60°	19 20	4.10 4.10	3.55 3.65	0.87 0.89	OK OK
photon	W-80-L	11 12 26 27	5.30 5.30 5.30 5.30	6.15 6.30 6.05 5.90	1.16 1.19 1.14 1.11	OK OK OK OK
		28 29	48.0 48.0	46.20 50.00	0.96 1.04	OK OK
		2 7	375 375	391.75 400.00	1.04 1.07	OK OK
		13 16	4.40 4.40	4.00 3.95	0.91 0.90	OK OK
	W-110	4 5	4.70 4.70	3.45 3.95	0.73 0.84	OK OK
	Cs-137	1 6 14 15 17 18 21 22 25 32				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.73	0.73	0.76	0.69	8%	-
Sr-90	2	0.88	0.88	0.88	0.88	0%	Sr-90: 45°/0°
Sr-90/45°	2	0.95	0.95	1.02	0.88	11%	1.07
beta all	6	0.88	0.85	1.02	0.69	13%	-
N-30	2	1.02	1.02	1.14	0.89	17%	N-30: 60°/0°
N-30/60°	2	0.88	0.88	0.89	0.87	2%	0.86
W-80-L	4	1.15	1.15	1.19	1.11	3%	W-80: H/L
W-80-M	2	1.00	1.00	1.04	0.96	6%	0.92
W-80-H	2	1.06	1.06	1.07	1.04	1%	
W-110	2	0.90	0.90	0.91	0.90	1%	-
Cs-137	2	0.79	0.79	0.84	0.73	10%	-
photon all	16	1.00	0.99	1.19	0.73	14%	-
All	22	0.90	0.95	1.19	0.69	15%	-

outliers: 0 of 22

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 29 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	17 20	4.51 4.51	5.51 5.08	1.22 1.13	OK OK
	Sr-90	7 23	3.00 3.00	4.04 4.47	1.35 1.49	OK OK
	Sr-90/45°	27 32	2.80 2.80	3.77 3.72	1.35 1.33	OK OK
	N-30	24 28	5.50 5.50	7.16 8.90	1.30 1.62	OK OK
photon	N-30/60°	4 29	5.20 5.20	8.23 7.11	1.58 1.37	OK OK
	W-80-L	3 5 16 18	4.50 4.50 4.50 4.50	6.41 6.29 6.34 6.14	1.43 1.40 1.41 1.36	OK OK OK OK
		11 12	55.0 55.0	72.23 76.24	1.31 1.39	OK OK
		2 6	369 369	503.13 571.25	1.36 1.55	OK OK
	W-110	21 22	5.10 5.10	5.08 5.62	1.00 1.10	OK OK
	Cs-137	8 10	4.70 4.70	6.04 6.20	1.29 1.32	OK OK
	NIR	1				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	9				
	NIR	13				
	NIR	14				
	NIR	15				
	NIR	19				
	NIR	25				
	NIR	26				
	NIR	30				
	NIR	31				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	1.17	1.17	1.22	1.13	6%	-
Sr-90	2	1.42	1.42	1.49	1.35	7%	Sr-90: 45°/0°
Sr-90/45°	2	1.34	1.34	1.35	1.33	1%	0.94
beta all	6	1.34	1.31	1.49	1.13	9%	-
N-30	2	1.46	1.46	1.62	1.30	15%	N-30: 60°/0°
N-30/60°	2	1.47	1.47	1.58	1.37	10%	1.01
W-80-L	4	1.40	1.40	1.43	1.36	2%	W-80: H/L
W-80-M	2	1.35	1.35	1.39	1.31	4%	1.04
W-80-H	2	1.46	1.46	1.55	1.36	9%	
W-110	2	1.05	1.05	1.10	1.00	7%	-
Cs-137	2	1.30	1.30	1.32	1.29	2%	-
photon all	16	1.37	1.36	1.62	1.00	12%	-
All	22	1.36	1.35	1.62	1.00	11%	-

outliers: 0 of 22

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 30 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results			
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)			
beta	Kr-85	4 27	4.51 4.51	4.03 4.00	0.89 0.89	OK OK		
	Sr-90	3 19	3.00 3.00	3.90 3.75	1.30 1.25	OK OK		
	Sr-90/45°	16 32	2.80 2.80	3.04 2.83	1.09 1.01	OK OK		
	N-30	28 29	5.20 5.20	5.59 5.74	1.08 1.10	OK OK		
	N-30/60°	10 12	5.30 5.30	6.11 6.03	1.15 1.14	OK OK		
	W-80-L		5 6 14 15	5.60 5.60 5.60 5.60	5.99 6.11 6.17 6.17	1.07 1.09 1.10 1.10		
photon			1 2	55.0 55.0	61.00 62.00	1.11 1.13		
			25 26	356 356	400.70 397.80	1.13 1.12		
			23 24	4.90 4.90	4.29 4.16	0.88 0.85		
			20 21	4.60 4.60	4.34 4.48	0.94 0.97		
NIR	7				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated			
NIR	8							
NIR	9							
NIR	11							
NIR	13							
NIR	17							
NIR	18							
NIR	22							
NIR	30							
NIR	31							

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.89	0.89	0.89	0.89	1%	-
Sr-90	2	1.28	1.28	1.30	1.25	3%	Sr-90: 45°/0°
Sr-90/45°	2	1.05	1.05	1.09	1.01	5%	0.82
beta all	6	1.05	1.07	1.30	0.89	16%	-
N-30	2	1.09	1.09	1.10	1.08	2%	N-30: 60°/0°
N-30/60°	2	1.15	1.15	1.15	1.14	1%	
W-80-L	4	1.10	1.09	1.10	1.07	1%	W-80: H/L
W-80-M	2	1.12	1.12	1.13	1.11	1%	
W-80-H	2	1.12	1.12	1.13	1.12	1%	-
W-110	2	0.86	0.86	0.88	0.85	2%	-
Cs-137	2	0.96	0.96	0.97	0.94	2%	-
photon all	16	1.10	1.06	1.15	0.85	9%	-
All	22	1.10	1.06	1.30	0.85	11%	-

outliers: 0 of 22

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 31: Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	17 25	4.51 4.51	0.69 0.62	0.15 0.14	(outlier) (outlier)
	Sr-90	10 29	3.00 3.00	3.27 3.28	1.09 1.09	(OK) (OK)
	Sr-90/45°	15 19	2.80 2.80	2.21 2.29	0.79 0.82	(OK) (OK)
	N-30	22 23	4.40 4.40	5.06 5.18	1.15 1.18	OK OK
	N-30/60°	4 6	4.90 4.90	5.77 5.75	1.18 1.17	OK OK
photon	W-80-L	5 7 20	5.50 5.50 5.50	6.45 6.37 6.21	1.17 1.16 1.13	OK OK OK
		21	5.50	6.30	1.14	OK
		W-80-M	8 9	50.0 50.0	57.46 60.13	1.15 1.20
	W-80-H	14 18	340 340	420.40 407.50	1.24 1.20	OK OK
		W-110	12 13	5.50 5.50	5.07 5.28	0.92 0.96
	Cs-137	1 2	5.00 5.00	5.47 5.59	1.09 1.12	OK OK
	NIR	3				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	11				
	NIR	16				
	NIR	24				
	NIR	26				
	NIR	27				
	NIR	28				
	NIR	30				
	NIR	31				
	NIR	32				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.15	0.15	0.15	0.14	7%	-
Sr-90	2	1.09	1.09	1.09	1.09	0%	Sr-90: 45°/0°
Sr-90/45°	2	0.80	0.80	0.82	0.79	2%	0.73
beta all	6	0.80	0.68	1.09	0.14	64%	-
N-30	2	1.16	1.16	1.18	1.15	2%	N-30: 60°/0°
N-30/60°	2	1.18	1.18	1.18	1.17	0%	
W-80-L	4	1.15	1.15	1.17	1.13	2%	W-80: H/L
W-80-M	2	1.18	1.18	1.20	1.15	3%	
W-80-H	2	1.22	1.22	1.24	1.20	2%	
W-110	2	0.94	0.94	0.96	0.92	3%	-
Cs-137	2	1.11	1.11	1.12	1.09	2%	-
photon all	16	1.15	1.14	1.24	0.92	7%	-
All	22	1.14	1.01	1.24	0.14	30%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 32 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	25 27	4.51 4.51	4.70 4.90	1.04 1.09	OK OK
	Sr-90	22 31	3.00 3.00	4.20 4.00	1.40 1.33	OK OK
	Sr-90/45°	21 28	2.80 2.80	3.10 2.90	1.11 1.04	OK OK
	N-30	24 29	5.10 5.10	6.40 6.20	1.25 1.22	OK OK
photon	N-30/60°	4 5	5.30 5.30	6.70 6.70	1.26 1.26	OK OK
	W-80-L	9 10 11 15	4.60 4.60 4.60 4.60	5.60 5.70 5.40 6.00	1.22 1.24 1.17 1.30	OK OK OK OK
		2 3	49.0 49.0	57.10 59.00	1.17 1.20	OK OK
		13 14	321 321	382.00 400.30	1.19 1.25	OK OK
	W-110	30 32	4.50 4.50	4.40 4.40	0.98 0.98	OK OK
	Cs-137	16 17	4.60 4.60	4.90 4.90	1.07 1.07	OK OK
	NIR	1				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	6				
	NIR	7				
	NIR	8				
	NIR	12				
	NIR	18				
	NIR	19				
	NIR	20				
	NIR	23				
	NIR	26				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	1.06	1.06	1.09	1.04	3%	-
Sr-90	2	1.37	1.37	1.40	1.33	3%	Sr-90: 45°/0°
Sr-90/45°	2	1.07	1.07	1.11	1.04	5%	0.78
beta all	6	1.10	1.17	1.40	1.04	14%	-
N-30	2	1.24	1.24	1.25	1.22	2%	N-30: 60°/0°
N-30/60°	2	1.26	1.26	1.26	1.26	0%	
W-80-L	4	1.23	1.23	1.30	1.17	4%	W-80: H/L
W-80-M	2	1.18	1.18	1.20	1.17	2%	
W-80-H	2	1.22	1.22	1.25	1.19	3%	
W-110	2	0.98	0.98	0.98	0.98	0%	-
Cs-137	2	1.07	1.07	1.07	1.07	0%	-
photon all	16	1.21	1.18	1.30	0.98	9%	-
All	22	1.20	1.17	1.40	0.98	10%	-

outliers: 0 of 22

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 33 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	22	4.51	4.60	1.02	OK
		28	4.51	4.34	0.96	OK
	Sr-90	12	3.00	4.13	1.38	OK
		20	3.00	4.29	1.43	OK
	Sr-90/45°	26	2.80	3.25	1.16	OK
		31	2.80	3.17	1.13	OK
photon	N-30	23	4.50	6.08	1.35	OK
		24	4.50	6.24	1.39	OK
	N-30/60°	1	5.30	7.08	1.34	OK
		3	5.30	7.85	1.48	OK
	W-80-L	2	5.30	6.59	1.24	OK
		4	5.30	6.63	1.25	OK
		8	5.30	6.57	1.24	OK
		10	5.30	6.43	1.21	OK
	W-80-M	25	44.0	51.70	1.18	OK
		27	44.0	56.10	1.28	OK
	W-80-H	14	320	406.00	1.27	OK
		16	320	416.00	1.30	OK
	W-110	13	5.00	5.30	1.06	OK
		15	5.00	5.03	1.01	OK
	Cs-137	5	5.40	5.98	1.11	OK
		6	5.40	6.22	1.15	OK
	NIR	7				
	NIR	9				
	NIR	11				
	NIR	17				
	NIR	18				
	NIR	19				
	NIR	21				
	NIR	29				
	NIR	30				
	NIR	32				


Legend for Quality

- L low dose
- M medium dose
- H high dose
- NIR not irradiated
- WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.99	0.99	1.02	0.96	4%	-
Sr-90	2	1.40	1.40	1.43	1.38	3%	Sr-90: 45°/0°
Sr-90/45°	2	1.15	1.15	1.16	1.13	2%	0.82
beta all	6	1.15	1.18	1.43	0.96	16%	-
N-30	2	1.37	1.37	1.39	1.35	2%	N-30: 60°/0°
N-30/60°	2	1.41	1.41	1.48	1.34	7%	1.03
W-80-L	4	1.24	1.24	1.25	1.21	1%	W-80: H/L
W-80-M	2	1.23	1.23	1.28	1.18	6%	1.04
W-80-H	2	1.28	1.28	1.30	1.27	2%	
W-110	2	1.03	1.03	1.06	1.01	4%	-
Cs-137	2	1.13	1.13	1.15	1.11	3%	-
photon all	16	1.25	1.24	1.48	1.01	10%	-
All	22	1.24	1.22	1.48	0.96	12%	-

outliers: 0 of 22

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 34 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	13	4.51	0.34	0.07	(outlier)
		16	4.51	1.05	0.23	(outlier)
	Sr-90	10	3.00	4.64	1.55	(OK)
		28	3.00	3.97	1.32	(OK)
	Sr-90/45°	4	2.80	2.58	0.92	(OK)
photon	N-30	2	5.50	8.74	1.59	OK
		3	5.50	8.83	1.61	OK
	N-30/60°	7	4.50	6.79	1.51	OK
		8	4.50	6.42	1.43	OK
	W-80-L	9	4.30	7.13	1.66	OK
		11	4.30	5.99	1.39	OK
		20	4.30	6.68	1.55	OK
		25	4.30	7.38	1.72	OK
	W-80-M	17	57.0	81.22	1.42	OK
		19	57.0	87.23	1.53	OK
	W-80-H	22	369	559.08	1.52	OK
		23	369	513.95	1.39	OK
	W-110	27	5.30	6.24	1.18	OK
		32	5.30	6.45	1.22	OK
	Cs-137	5	5.00	5.98	1.20	OK
		6	5.00	7.61	1.52	OK
	NIR	1				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	12				
	NIR	14				
	NIR	18				
	NIR	21				
	NIR	24				
	NIR	26				
	NIR	29				
	NIR	30				
	NIR	31				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.15	0.15	0.23	0.07	73%	-
Sr-90	2	1.44	1.44	1.55	1.32	11%	Sr-90: 45°/0°
Sr-90/45°	2	1.02	1.02	1.12	0.92	14%	0.71
beta all	6	1.02	0.87	1.55	0.07	68%	-
N-30	2	1.60	1.60	1.61	1.59	1%	N-30: 60°/0° 0.92
N-30/60°	2	1.47	1.47	1.51	1.43	4%	
W-80-L	4	1.61	1.58	1.72	1.39	9%	W-80: H/L 0.92
W-80-M	2	1.48	1.48	1.53	1.42	5%	
W-80-H	2	1.45	1.45	1.52	1.39	6%	
W-110	2	1.20	1.20	1.22	1.18	2%	-
Cs-137	2	1.36	1.36	1.52	1.20	17%	-
photon all	16	1.51	1.46	1.72	1.18	11%	-
All	22	1.43	1.30	1.72	0.07	32%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 35 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	7 31	4.51 4.51	3.94 4.02	0.87 0.89	OK OK
	Sr-90	10 13	3.00 3.00	3.60 3.35	1.20 1.12	OK OK
	Sr-90/45°	2 8	2.80 2.80	3.25 3.25	1.16 1.16	OK OK
	N-30	27 28	5.30 5.30	5.61 5.42	1.06 1.02	OK OK
photon	N-30/60°	12 14	4.10 4.10	4.31 4.44	1.05 1.08	OK OK
	W-80-L	17 19 26 32	5.10 5.10 5.10 5.10	6.31 6.18 6.21 6.05	1.24 1.21 1.22 1.19	OK OK OK OK
		20 22	55.0 55.0	69.13 66.90	1.26 1.22	OK OK
		24 25	324 324	392.44 377.20	1.21 1.16	OK OK
	W-110	29 30	4.40 4.40	4.34 4.69	0.99 1.07	OK OK
	Cs-137	4 6	5.00 5.00	4.82 4.94	0.96 0.99	OK OK
	NIR	1				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	3				
	NIR	5				
	NIR	9				
	NIR	15				
	NIR	16				
	NIR	18				
	NIR	23				
	WIR	11				
	WIR	21				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.88	0.88	0.89	0.87	1%	-
Sr-90	2	1.16	1.16	1.20	1.12	5%	Sr-90: 45°/0°
Sr-90/45°	2	1.16	1.16	1.16	1.16	0%	1.00
beta all	6	1.14	1.07	1.20	0.87	14%	-
N-30	2	1.04	1.04	1.06	1.02	2%	N-30: 60°/0°
N-30/60°	2	1.07	1.07	1.08	1.05	2%	1.03
W-80-L	4	1.21	1.21	1.24	1.19	2%	W-80: H/L
W-80-M	2	1.24	1.24	1.26	1.22	2%	0.98
W-80-H	2	1.19	1.19	1.21	1.16	3%	
W-110	2	1.03	1.03	1.07	0.99	5%	-
Cs-137	2	0.98	0.98	0.99	0.96	2%	-
photon all	16	1.12	1.12	1.26	0.96	9%	-
All	22	1.14	1.11	1.26	0.87	10%	-

outliers: 0 of 22

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 36 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	3	4.51	5.85	1.30	OK
		28	4.51	4.79	1.06	OK
	Sr-90	13	3.00	4.81	1.60	OK
		27	3.00	4.44	1.48	OK
photon	Sr-90/45°	17	2.80	3.93	1.40	OK
		20	2.80	4.07	1.45	OK
	N-30	16	4.90	7.17	1.46	OK
		18	4.90	6.52	1.33	OK
	N-30/60°	8	5.00	7.54	1.51	OK
		11	5.00	7.67	1.53	OK
	W-80-L	9	4.80	6.98	1.45	OK
		10	4.80	7.15	1.49	OK
		21	4.80	7.09	1.48	OK
		22	4.80	7.31	1.52	OK
	W-80-M	12	45.0	69.93	1.55	OK
		15	45.0	66.93	1.49	OK
	W-80-H	31	373	566.93	1.52	OK
		32	373	576.93	1.55	OK
	W-110	5	4.40	4.83	1.10	OK
		7	4.40	4.87	1.11	OK
	Cs-137	2	5.50	6.28	1.14	OK
		4	5.50	7.77	1.41	OK
	NIR	1				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	6				
	NIR	14				
	NIR	19				
	NIR	23				
	NIR	24				
	NIR	25				
	NIR	26				
	NIR	29				
	NIR	30				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	1.18	1.18	1.30	1.06	14%	-
Sr-90	2	1.54	1.54	1.60	1.48	6%	Sr-90: 45°/0° 0.93
Sr-90/45°	2	1.43	1.43	1.45	1.40	2%	
beta all	6	1.43	1.38	1.60	1.06	13%	-
N-30	2	1.40	1.40	1.46	1.33	7%	N-30: 60°/0° 1.09
N-30/60°	2	1.52	1.52	1.53	1.51	1%	
W-80-L	4	1.48	1.49	1.52	1.45	2%	W-80: H/L 1.03
W-80-M	2	1.52	1.52	1.55	1.49	3%	
W-80-H	2	1.53	1.53	1.55	1.52	1%	
W-110	2	1.10	1.10	1.11	1.10	1%	-
Cs-137	2	1.28	1.28	1.41	1.14	15%	-
photon all	16	1.48	1.42	1.55	1.10	11%	-
All	22	1.47	1.41	1.60	1.06	12%	-

outliers: 0 of 22

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 37 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)		
beta	Kr-85	2	4.51	0.38	0.08	(outlier)	
		5	4.51	0.37	0.08	(outlier)	
	Sr-90	6	3.00	1.94	0.65	(OK)	
		14	3.00	2.08	0.69	(OK)	
	Sr-90/45°	3	2.80	1.37	0.49	(OK)	
photon		18	2.80	1.00	0.36	(outlier)	
	N-30	22	4.30	2.97	0.69	OK	
		23	4.30	3.11	0.72	OK	
	N-30/60°	31	4.70	3.00	0.64	OK	
		32	4.70	2.91	0.62	OK	
		1	4.40	3.15	0.72	OK	
	W-80-L	4	4.40	3.20	0.73	OK	
		16	4.40	3.48	0.79	OK	
		19	4.40	3.40	0.77	OK	
	W-80-M	29	49.0	35.40	0.72	OK	
		30	49.0	35.10	0.72	OK	
	W-80-H	10	332	156.00	0.47	outlier	
		11	332	154.00	0.46	outlier	
	W-110	8	4.40	2.85	0.65	OK	
		9	4.40	2.63	0.60	OK	
	Cs-137	26	5.20	3.51	0.68	OK	
		27	5.20	3.74	0.72	OK	
	NIR	7				Legend for Quality	
	NIR	12				L low dose	
	NIR	13				M medium dose	
	NIR	15				H high dose	
	NIR	17				NIR not irradiated	
	NIR	20				WIR wrongly irradiated	
	NIR	21					
	NIR	24					
	NIR	25					
	NIR	28					

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.08	0.08	0.08	0.08	2%	-
Sr-90	2	0.67	0.67	0.69	0.65	5%	Sr-90: 45°/0°
Sr-90/45°	2	0.42	0.42	0.49	0.36	22%	0.63
beta all	6	0.42	0.39	0.69	0.08	68%	-
N-30	2	0.71	0.71	0.72	0.69	3%	N-30: 60°/0°
N-30/60°	2	0.63	0.63	0.64	0.62	2%	0.89
W-80-L	4	0.75	0.75	0.79	0.72	5%	W-80: H/L
W-80-M	2	0.72	0.72	0.72	0.72	1%	0.62
W-80-H	2	0.47	0.47	0.47	0.46	1%	
W-110	2	0.62	0.62	0.65	0.60	6%	-
Cs-137	2	0.70	0.70	0.72	0.68	4%	-
photon all	16	0.70	0.67	0.79	0.46	14%	-
All	22	0.66	0.59	0.79	0.08	33%	-

outliers: 2 of 16

fraction of outliers: 13%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 38 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	7	4.51	4.60	1.02	(OK)
		17	4.51	2.95	0.65	(OK)
	Sr-90	4	3.00	3.74	1.25	(OK)
		30	3.00	4.13	1.38	(OK)
	Sr-90/45°	22	2.80	1.27	0.45	(OK)
photon		24	2.80	1.22	0.44	(OK)
	N-30	8	4.80	6.05	1.26	OK
		9	4.80	5.93	1.23	OK
	N-30/60°	18	4.60	5.38	1.17	OK
		20	4.60	5.44	1.18	OK
	W-80-L	1	5.00	6.61	1.32	OK
		6	5.00	6.80	1.36	OK
		11	5.00	6.77	1.35	OK
		16	5.00	6.80	1.36	OK
	W-80-M	23	46.0	62.16	1.35	OK
		25	46.0	61.41	1.34	OK
	W-80-H	31	382	423.88	1.11	OK
		32	382	436.05	1.14	OK
	W-110	26	5.30	5.69	1.07	OK
		28	5.30	5.38	1.01	OK
	Cs-137	14	4.60	5.21	1.13	OK
		15	4.60	5.34	1.16	OK
	NIR	2				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	3				
	NIR	5				
	NIR	10				
	NIR	12				
	NIR	13				
	NIR	19				
	NIR	21				
	NIR	27				
	NIR	29				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.84	0.84	1.02	0.65	31%	-
Sr-90	2	1.31	1.31	1.38	1.25	7%	Sr-90: 45°/0°
Sr-90/45°	2	0.44	0.44	0.45	0.44	3%	0.34
beta all	6	0.84	0.86	1.38	0.44	47%	-
N-30	2	1.25	1.25	1.26	1.23	1%	N-30: 60°/0°
N-30/60°	2	1.18	1.18	1.18	1.17	1%	
W-80-L	4	1.36	1.35	1.36	1.32	1%	W-80: H/L
W-80-M	2	1.34	1.34	1.35	1.34	1%	
W-80-H	2	1.13	1.13	1.14	1.11	2%	-
W-110	2	1.04	1.04	1.07	1.01	4%	-
Cs-137	2	1.15	1.15	1.16	1.13	2%	-
photon all	16	1.21	1.22	1.36	1.01	9%	-
All	22	1.18	1.12	1.38	0.44	24%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 39 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	14 28	4.51 4.51	4.38 4.37	0.97 0.97	OK OK
	Sr-90	16 25	3.00 3.00	4.30 3.67	1.43 1.22	OK OK
	Sr-90/45°	3 15	2.80 2.80	1.86 1.69	0.66 0.60	OK OK
	N-30	8 10	4.60 4.60	4.69 4.69	1.02 1.02	OK OK
	N-30/60°	1 4	4.90 4.90	5.97 5.50	1.22 1.12	OK OK
photon	W-80-L	2 5 7	4.40 4.40 4.40	5.64 5.20 5.27	1.28 1.18 1.20	OK OK OK
		9	4.40	5.23	1.19	OK
		22 24	56.0 56.0	67.54 60.36	1.21 1.08	OK OK
	W-80-H	29	345	359.24	1.04	OK
		30	345	364.49	1.06	OK
	W-110	19	4.60	4.10	0.89	OK
		23	4.60	3.64	0.79	OK
	Cs-137	12 17	5.60 5.60	5.78 5.53	1.03 0.99	OK OK
	NIR NIR NIR NIR NIR NIR NIR NIR NIR NIR	6 11 13 18 20 21 26 27 31 32				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.97	0.97	0.97	0.97	0%	-
Sr-90	2	1.33	1.33	1.43	1.22	11%	Sr-90: 45°/0°
Sr-90/45°	2	0.63	0.63	0.66	0.60	7%	0.48
beta all	6	0.97	0.98	1.43	0.60	33%	-
N-30	2	1.02	1.02	1.02	1.02	0%	N-30: 60°/0°
N-30/60°	2	1.17	1.17	1.22	1.12	6%	1.15
W-80-L	4	1.19	1.21	1.28	1.18	4%	W-80: H/L
W-80-M	2	1.14	1.14	1.21	1.08	8%	0.87
W-80-H	2	1.05	1.05	1.06	1.04	1%	
W-110	2	0.84	0.84	0.89	0.79	8%	-
Cs-137	2	1.01	1.01	1.03	0.99	3%	-
photon all	16	1.07	1.08	1.28	0.79	12%	-
All	22	1.05	1.05	1.43	0.60	19%	-

outliers: 0 of 22

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 40 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	25 28	4.51 4.51	5.23 4.99	1.16 1.11	OK OK
	Sr-90	6 29	3.00 3.00	5.05 4.89	1.68 1.63	OK OK
	Sr-90/45°	12 21	2.80 2.80	4.86 4.86	1.74 1.74	OK OK
	N-30	10 11	5.00 5.00	4.93 4.93	0.99 0.99	OK OK
	N-30/60°	18 19	4.80 4.80	4.62 4.59	0.96 0.96	OK OK
photon	W-80-L	20 22 23 26	5.50 5.50 5.50 5.50	5.58 5.58 5.50 5.53	1.01 1.01 1.00 1.01	OK OK OK OK
		30 31	54.0 54.0	54.33 53.32	1.01 0.99	OK OK
		13 14	378 378	395.90 385.90	1.05 1.02	OK OK
		3 9	4.60 4.60	4.46 4.43	0.97 0.96	OK OK
	W-110	5 7	5.40 5.40	5.40 5.41	1.00 1.00	OK OK
	Cs-137	1 2 4 8 15 16 17 24 27 32				
						Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	1.13	1.13	1.16	1.11	3%	-
Sr-90	2	1.66	1.66	1.68	1.63	2%	Sr-90: 45°/0°
Sr-90/45°	2	1.74	1.74	1.74	1.74	0%	1.05
beta all	6	1.66	1.51	1.74	1.11	19%	-
N-30	2	0.99	0.99	0.99	0.99	0%	N-30: 60°/0°
N-30/60°	2	0.96	0.96	0.96	0.96	0%	0.97
W-80-L	4	1.01	1.01	1.01	1.00	1%	W-80: H/L
W-80-M	2	1.00	1.00	1.01	0.99	1%	1.03
W-80-H	2	1.03	1.03	1.05	1.02	2%	
W-110	2	0.97	0.97	0.97	0.96	0%	-
Cs-137	2	1.00	1.00	1.00	1.00	0%	-
photon all	16	1.00	1.00	1.05	0.96	2%	-
All	22	1.01	1.14	1.74	0.96	24%	-

outliers: 0 of 22

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 41: Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)		
beta	Kr-85	17	4.51	0.51	0.11	(outlier)	
		29	4.51	0.51	0.11	(outlier)	
	Sr-90	4	3.00	1.28	0.43	(OK)	
		10	3.00	1.81	0.60	(OK)	
	Sr-90/45°	2	2.80	0.51	0.18	(outlier)	
photon	N-30	1	4.80	4.74	0.99	OK	
		7	4.80	4.65	0.97	OK	
	N-30/60°	11	5.00	5.21	1.04	OK	
		12	5.00	5.13	1.03	OK	
	W-80-L	6	5.30	4.74	0.89	OK	
		8	5.30	4.80	0.91	OK	
		19	5.30	4.83	0.91	OK	
		20	5.30	3.95	0.74	OK	
	W-80-M	23	48.0	45.18	0.94	OK	
		24	48.0	36.70	0.76	OK	
	W-80-H	22	369	341.21	0.92	OK	
		26	369	334.69	0.91	OK	
	W-110	13	4.30	3.57	0.83	OK	
		14	4.30	3.28	0.76	OK	
	Cs-137	27	4.70	3.17	0.68	OK	
		28	4.70	2.93	0.62	OK	
	NIR	3				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated	
	NIR	5					
	NIR	9					
	NIR	16					
	NIR	18					
	NIR	21					
	NIR	25					
	NIR	30					
	NIR	31					
	NIR	32					

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)	
Kr-85	2	0.11	0.11	0.11	0.11	0%	-	
Sr-90	2	0.52	0.52	0.60	0.43	24%	Sr-90: 45°/0°	
Sr-90/45°	2	0.25	0.25	0.32	0.18	39%	0.49	
beta all	6	0.25	0.29	0.60	0.11	67%	-	
N-30	2	0.98	0.98	0.99	0.97	1%	N-30: 60°/0°	
N-30/60°	2	1.03	1.03	1.04	1.03	1%		
W-80-L	4	0.90	0.86	0.91	0.74	9%	W-80: H/L	
W-80-M	2	0.85	0.85	0.94	0.76	15%		
W-80-H	2	0.92	0.92	0.92	0.91	1%		
W-110	2	0.80	0.80	0.83	0.76	6%	-	
Cs-137	2	0.65	0.65	0.68	0.62	6%	-	
photon all	16	0.91	0.87	1.04	0.62	14%	-	
All	22	0.80	0.71	1.04	0.11	42%	-	

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 42 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)		
beta	Kr-85	18	4.51	0.40	0.09	(outlier)	
		24	4.51	0.40	0.09	(outlier)	
	Sr-90	12	3.00	3.20	1.07	(OK)	
		25	3.00	3.40	1.13	(OK)	
	Sr-90/45°	15	2.80	2.10	0.75	(OK)	
		32	2.80	2.20	0.79	(OK)	
photon	N-30	21	5.60	10.50	1.88	outlier	
		27	5.60	10.10	1.80	OK	
	N-30/60°	16	4.90	9.10	1.86	OK	
		17	4.90	6.40	1.31	OK	
	W-80-L	7	5.70	9.40	1.65	OK	
		9	5.70	10.10	1.77	OK	
		28	5.70	9.30	1.63	OK	
	W-80-M	30	45.0	73.50	1.63	OK	
		31	45.0	68.00	1.51	OK	
	W-80-H	11	344	605.50	1.76	outlier	
		19	344	670.00	1.95	outlier	
	W-110	2	5.80	6.10	1.05	OK	
		8	5.80	8.80	1.52	OK	
	Cs-137	3	4.60	5.20	1.13	OK	
		4	4.60	4.70	1.02	OK	
	NIR	1				Legend for Quality	
	NIR	5				L low dose	
	NIR	6				M medium dose	
	NIR	10				H high dose	
	NIR	13				NIR not irradiated	
	NIR	14				WIR wrongly irradiated	
	NIR	20					
	NIR	22					
	NIR	23					
	NIR	26					

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.09	0.09	0.09	0.09	0%	-
Sr-90	2	1.10	1.10	1.13	1.07	4%	Sr-90: 45°/0°
Sr-90/45°	2	0.77	0.77	0.79	0.75	3%	0.70
beta all	6	0.77	0.65	1.13	0.09	71%	-
N-30	2	1.84	1.84	1.88	1.80	3%	N-30: 60°/0°
N-30/60°	2	1.58	1.58	1.86	1.31	25%	0.86
W-80-L	4	1.66	1.68	1.77	1.63	4%	W-80: H/L
W-80-M	2	1.57	1.57	1.63	1.51	5%	1.10
W-80-H	2	1.85	1.85	1.95	1.76	7%	
W-110	2	1.28	1.28	1.52	1.05	26%	-
Cs-137	2	1.08	1.08	1.13	1.02	7%	-
photon all	16	1.64	1.57	1.95	1.02	19%	-
All	22	1.51	1.32	1.95	0.09	41%	-

outliers: 3 of 16

fraction of outliers: 19%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 43 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)		
beta	Kr-85	11	4.51	0.51	0.11	(outlier)	
		17	4.51	0.59	0.13	(outlier)	
	Sr-90	20	3.00	2.22	0.74	(OK)	
		25	3.00	2.38	0.79	(OK)	
	Sr-90/45°	19	2.80	1.45	0.52	(OK)	
photon	N-30	6	4.90	5.07	1.03	OK	
		7	4.90	5.71	1.16	OK	
	N-30/60°	1	5.10	5.46	1.07	OK	
		4	5.10	6.28	1.23	OK	
	W-80-L	21	4.30	4.79	1.11	OK	
		22	4.30	4.89	1.14	OK	
		23	4.30	4.58	1.06	OK	
		32	4.30	5.13	1.19	OK	
	W-80-M	16	54.0	64.39	1.19	OK	
		18	54.0	54.90	1.02	OK	
	W-80-H	12	366	436.96	1.19	OK	
		14	366	405.98	1.11	OK	
	W-110	30	5.10	4.30	0.84	OK	
		31	5.10	4.54	0.89	OK	
	Cs-137	9	4.70	3.95	0.84	OK	
		10	4.70	3.46	0.74	OK	
	NIR	2				Legend for Quality	
	NIR	3				L low dose	
	NIR	5				M medium dose	
	NIR	8				H high dose	
	NIR	13				NIR not irradiated	
	NIR	15				WIR wrongly irradiated	
	NIR	24					
	NIR	26					
	NIR	28					
	NIR	29					

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)	
Kr-85	2	0.12	0.12	0.13	0.11	10%	-	
Sr-90	2	0.77	0.77	0.79	0.74	5%	Sr-90: 45°/0°	
Sr-90/45°	2	0.53	0.53	0.55	0.52	4%	0.69	
beta all	6	0.53	0.47	0.79	0.11	62%	-	
N-30	2	1.10	1.10	1.16	1.03	8%	N-30: 60°/0°	
N-30/60°	2	1.15	1.15	1.23	1.07	10%	1.05	
W-80-L	4	1.13	1.13	1.19	1.06	5%	W-80: H/L	
W-80-M	2	1.10	1.10	1.19	1.02	11%	1.02	
W-80-H	2	1.15	1.15	1.19	1.11	5%		
W-110	2	0.87	0.87	0.89	0.84	4%	-	
Cs-137	2	0.79	0.79	0.84	0.74	9%	-	
photon all	16	1.09	1.05	1.23	0.74	14%	-	
All	22	1.03	0.89	1.23	0.11	36%	-	

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 44 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	7	4.51	0.46	0.10	(outlier)
		19	4.51	0.54	0.12	(outlier)
	Sr-90	3	3.00	1.22	0.41	(outlier)
		18	3.00	1.28	0.43	(OK)
	Sr-90/45°	12	2.80	0.44	0.16	(outlier)
photon	N-30	25	4.80	3.50	0.73	OK
		28	4.80	3.42	0.71	OK
	N-30/60°	16	5.00	3.08	0.62	OK
		17	2.80	3.83	1.37	OK
	W-80-L	21	5.30	3.45	0.65	OK
		22	5.30	3.47	0.65	OK
		30	5.30	4.15	0.78	OK
		31	5.30	3.82	0.72	OK
	W-80-M	8	46.0	29.73	0.65	outlier
		9	46.0	30.46	0.66	outlier
	W-80-H	1	345	254.33	0.74	OK
		2	345	249.99	0.72	OK
	W-110	14	5.60	3.31	0.59	OK
		15	5.60	3.72	0.66	OK
	Cs-137	26	5.10	1.92	0.38	outlier
		27	5.10	2.22	0.44	outlier
	NIR	4				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	5				
	NIR	6				
	NIR	10				
	NIR	11				
	NIR	13				
	NIR	20				
	NIR	23				
	NIR	24				
	NIR	32				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.11	0.11	0.12	0.10	11%	-
Sr-90	2	0.42	0.42	0.43	0.41	3%	Sr-90: 45°/0°
Sr-90/45°	2	0.16	0.16	0.16	0.16	3%	0.39
beta all	6	0.16	0.23	0.43	0.10	64%	-
N-30	2	0.72	0.72	0.73	0.71	2%	N-30: 60°/0°
N-30/60°	2	0.99	0.99	1.37	0.62	54%	
W-80-L	4	0.69	0.70	0.78	0.65	9%	W-80: H/L
W-80-M	2	0.65	0.65	0.66	0.65	2%	
W-80-H	2	0.73	0.73	0.74	0.72	1%	
W-110	2	0.63	0.63	0.66	0.59	8%	-
Cs-137	2	0.41	0.41	0.44	0.38	10%	-
photon all	16	0.66	0.69	1.37	0.38	30%	-
All	22	0.65	0.57	1.37	0.10	50%	-

outliers: 4 of 16

fraction of outliers: 25%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 45 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	13	4.51	1.00	0.22	(outlier)
		14	4.51	1.10	0.24	(outlier)
	Sr-90	20	3.00	3.20	1.07	(OK)
		31	3.00	2.90	0.97	(OK)
	Sr-90/45°	5	2.80	2.00	0.71	(OK)
photon	N-30	1	4.90	6.90	1.41	OK
		16	4.90	5.90	1.20	OK
	N-30/60°	21	5.10	5.50	1.08	OK
		22	5.10	6.30	1.24	OK
	W-80-L	3	4.70	7.10	1.51	OK
		8	4.70	7.00	1.49	OK
		26	4.70	7.20	1.53	OK
		29	4.70	7.00	1.49	OK
	W-80-M	15	51.0	70.70	1.39	OK
		27	51.0	78.60	1.54	OK
	W-80-H	2	369	592.10	1.60	OK
		4	369	578.40	1.57	OK
	W-110	11	4.50	6.20	1.38	OK
		12	4.50	6.10	1.36	OK
	Cs-137	24	5.60	6.00	1.07	OK
		25	5.60	4.90	0.88	OK
	NIR	6				
	NIR	7				
	NIR	10				
	NIR	17				
	NIR	18				
	NIR	19				
	NIR	23				
	NIR	28				
	NIR	30				
	NIR	32				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.23	0.23	0.24	0.22	7%	-
Sr-90	2	1.02	1.02	1.07	0.97	7%	Sr-90: 45°/0°
Sr-90/45°	2	0.71	0.71	0.71	0.71	0%	0.70
beta all	6	0.71	0.65	1.07	0.22	54%	-
N-30	2	1.31	1.31	1.41	1.20	11%	N-30: 60°/0°
N-30/60°	2	1.16	1.16	1.24	1.08	10%	0.89
W-80-L	4	1.50	1.51	1.53	1.49	1%	W-80: H/L
W-80-M	2	1.46	1.46	1.54	1.39	7%	1.05
W-80-H	2	1.59	1.59	1.60	1.57	2%	
W-110	2	1.37	1.37	1.38	1.36	1%	-
Cs-137	2	0.97	0.97	1.07	0.88	14%	-
photon all	16	1.40	1.36	1.60	0.88	15%	-
All	22	1.30	1.17	1.60	0.22	35%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 46 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	16 23	4.51 4.51	0.31 0.00	0.07 0.00	outlier outlier
	Sr-90	13 18	3.00 3.00	3.14 2.20	1.05 0.73	OK OK
	Sr-90/45°	10 26	2.80 2.80	1.34 1.03	0.48 0.37	OK outlier
	N-30	31 32	4.50 4.50	7.28 9.09	1.62 2.02	OK outlier
	N-30/60°	17 19	4.40 4.40	9.28 5.55	2.11 1.26	outlier OK
photon	W-80-L	2 5 24	4.50 4.50 4.50	6.79 5.36 7.72	1.51 1.19 1.72	OK OK OK
		25	4.50	7.14	1.59	OK
		W-80-M	6 8	45.0 45.0	70.45 61.99	1.57 1.38
	W-80-H	27 28	333 333	486.69 539.20	1.46 1.62	OK OK
		W-110	12 14	4.60 4.60	3.81 6.22	0.83 1.35
	Cs-137	21 22	4.80 4.80	4.77 5.22	0.99 1.09	OK OK
	NIR	1				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	3				
	NIR	4				
	NIR	7				
	NIR	9				
	NIR	11				
	NIR	15				
	NIR	20				
	NIR	29				
	NIR	30				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.03	0.03	0.07	0.00	141%	-
Sr-90	2	0.89	0.89	1.05	0.73	25%	Sr-90: 45°/0°
Sr-90/45°	2	0.42	0.42	0.48	0.37	18%	0.48
beta all	6	0.42	0.45	1.05	0.00	89%	-
N-30	2	1.82	1.82	2.02	1.62	16%	N-30: 60°/0°
N-30/60°	2	1.69	1.69	2.11	1.26	36%	0.93
W-80-L	4	1.55	1.50	1.72	1.19	15%	W-80: H/L
W-80-M	2	1.47	1.47	1.57	1.38	9%	1.03
W-80-H	2	1.54	1.54	1.62	1.46	7%	
W-110	2	1.09	1.09	1.35	0.83	34%	-
Cs-137	2	1.04	1.04	1.09	0.99	6%	-
photon all	16	1.49	1.46	2.11	0.83	24%	-
All	22	1.31	1.18	2.11	0.00	49%	-

outliers: 5 of 22

fraction of outliers: 23%



Results: IC2019_{ext eye}

2 values out of diagramme range (>2!)

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 47 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	3	4.51	0.70	0.16	(outlier)
		30	4.51	0.88	0.20	(outlier)
	Sr-90	8	3.00	2.80	0.93	(OK)
		19	3.00	2.90	0.97	(OK)
	Sr-90/45°	5	2.80	1.76	0.63	(OK)
		12	2.80	1.75	0.63	(OK)
photon	N-30	26	4.40	5.80	1.32	OK
		28	4.40	5.89	1.34	OK
	N-30/60°	15	5.40	6.74	1.25	OK
		24	5.40	6.81	1.26	OK
	W-80-L	1	5.20	6.32	1.22	OK
		4	5.20	6.00	1.15	OK
		17	5.20	6.39	1.23	OK
		21	5.20	6.30	1.21	OK
	W-80-M	6	51.0	62.33	1.22	OK
		7	51.0	61.78	1.21	OK
	W-80-H	9	317	395.68	1.25	OK
		10	317	390.56	1.23	OK
	W-110	16	5.00	4.99	1.00	OK
		18	5.00	5.32	1.06	OK
	Cs-137	11	4.90	4.18	0.85	OK
		13	4.90	4.07	0.83	OK
	NIR	2				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	14				
	NIR	20				
	NIR	22				
	NIR	23				
	NIR	25				
	NIR	27				
	NIR	29				
	NIR	31				
	NIR	32				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.18	0.18	0.20	0.16	16%	-
Sr-90	2	0.95	0.95	0.97	0.93	2%	Sr-90: 45°/0°
Sr-90/45°	2	0.63	0.63	0.63	0.63	0%	0.66
beta all	6	0.63	0.58	0.97	0.16	60%	-
N-30	2	1.33	1.33	1.34	1.32	1%	N-30: 60°/0° 0.94
N-30/60°	2	1.25	1.25	1.26	1.25	1%	
W-80-L	4	1.21	1.20	1.23	1.15	3%	W-80: H/L 1.03
W-80-M	2	1.22	1.22	1.22	1.21	1%	
W-80-H	2	1.24	1.24	1.25	1.23	1%	
W-110	2	1.03	1.03	1.06	1.00	5%	-
Cs-137	2	0.84	0.84	0.85	0.83	2%	-
photon all	16	1.22	1.16	1.34	0.83	13%	-
All	22	1.18	1.01	1.34	0.16	34%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 48 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	17	4.51	0.69	0.15	outlier
		30	4.51	0.47	0.10	outlier
	Sr-90	4	3.00	2.37	0.79	OK
		15	3.00	2.33	0.78	OK
photon	Sr-90/45°	6	2.80	1.83	0.65	OK
		16	2.80	1.59	0.57	OK
	N-30	9	4.80	6.11	1.27	OK
		10	4.80	6.63	1.38	OK
	N-30/60°	19	4.70	6.38	1.36	OK
		20	4.70	8.03	1.71	OK
	W-80-L	12	4.90	6.10	1.24	OK
		18	4.90	6.57	1.34	OK
		24	4.90	6.17	1.26	OK
		26	4.90	6.57	1.34	OK
	W-80-M	13	44.0	56.00	1.27	OK
		14	44.0	54.18	1.23	OK
	W-80-H	25	315	391.53	1.24	OK
		27	315	403.90	1.28	OK
	W-110	1	4.90	5.11	1.04	OK
		2	4.90	5.92	1.21	OK
	Cs-137	21	5.40	4.80	0.89	OK
		23	5.40	4.44	0.82	OK
		31				
		32				
						Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.13	0.13	0.15	0.10	27%	-
Sr-90	2	0.78	0.78	0.79	0.78	1%	Sr-90: 45°/0°
Sr-90/45°	2	0.61	0.61	0.65	0.57	10%	0.78
beta all	6	0.61	0.51	0.79	0.10	60%	-
N-30	2	1.33	1.33	1.38	1.27	6%	N-30: 60°/0°
N-30/60°	2	1.53	1.53	1.71	1.36	16%	1.16
W-80-L	4	1.30	1.30	1.34	1.24	4%	W-80: H/L
W-80-M	2	1.25	1.25	1.27	1.23	2%	0.97
W-80-H	2	1.26	1.26	1.28	1.24	2%	
W-110	2	1.13	1.13	1.21	1.04	10%	-
Cs-137	2	0.86	0.86	0.89	0.82	6%	-
photon all	16	1.27	1.24	1.71	0.82	16%	-
All	22	1.24	1.04	1.71	0.10	39%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 49 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	10 16	4.51 4.51	0.47 0.62	0.10 0.14	outlier outlier
	Sr-90	2 31	3.00 3.00	2.39 2.22	0.80 0.74	OK OK
	Sr-90/45°	14 27	2.80 2.80	1.19 1.29	0.43 0.46	OK OK
	N-30	20 25	4.30 4.30	5.01 4.89	1.16 1.14	OK OK
	N-30/60°	23 24	4.40 4.40	4.87 4.91	1.11 1.11	OK OK
photon	W-80-L	11 12 18 19	5.70 5.70 5.70 5.70	5.85 5.80 5.97 6.21	1.03 1.02 1.05 1.09	OK OK OK OK
		21 22	58.0 58.0	63.94 58.77	1.10 1.01	OK OK
		7 8	358 358	445.13 445.94	1.24 1.25	OK OK
		28 29	4.70 4.70	4.12 3.98	0.88 0.85	OK OK
	W-110	9 17	4.60 4.60	3.33 3.15	0.72 0.68	OK OK
	Cs-137	1 3 4 5 6 13 15 26 30 32				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.12	0.12	0.14	0.10	19%	-
Sr-90	2	0.77	0.77	0.80	0.74	5%	Sr-90: 45°/0°
Sr-90/45°	2	0.44	0.44	0.46	0.43	6%	0.58
beta all	6	0.44	0.44	0.80	0.10	65%	-
N-30	2	1.15	1.15	1.16	1.14	2%	N-30: 60°/0°
N-30/60°	2	1.11	1.11	1.11	1.11	0%	0.97
W-80-L	4	1.04	1.04	1.09	1.02	3%	W-80: H/L
W-80-M	2	1.06	1.06	1.10	1.01	6%	1.19
W-80-H	2	1.24	1.24	1.25	1.24	0%	
W-110	2	0.86	0.86	0.88	0.85	3%	-
Cs-137	2	0.70	0.70	0.72	0.68	4%	-
photon all	16	1.07	1.03	1.25	0.68	16%	-
All	22	1.02	0.87	1.25	0.10	38%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 50 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	11 29	4.51 4.51	0.57 0.60	0.13 0.13	outlier outlier
	Sr-90	12 21	3.00 3.00	2.63 2.71	0.88 0.90	OK OK
	Sr-90/45°	6 30	2.80 2.80	1.72 1.50	0.61 0.54	OK OK
	N-30	1 2	5.70 5.70	7.60 7.29	1.33 1.28	OK OK
photon	N-30/60°	8 9	5.10 5.10	6.48 6.46	1.27 1.27	OK OK
	W-80-L	5 10 24	5.70 5.70 5.70	6.67 6.77 6.52	1.17 1.19 1.14	OK OK OK
		27	5.70	6.82	1.20	OK
	W-80-M	31 32	50.0 50.0	59.60 57.70	1.19 1.15	OK OK
	W-80-H	13 14	366 366	463.00 424.00	1.27 1.16	OK OK
	W-110	20 22	5.50 5.50	5.32 5.24	0.97 0.95	OK OK
	Cs-137	25 26	4.40 4.40	3.50 3.44	0.80 0.78	OK OK
	NIR	3				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	4				
	NIR	7				
	NIR	15				
	NIR	16				
	NIR	17				
	NIR	18				
	NIR	19				
	NIR	23				
	NIR	28				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.13	0.13	0.13	0.13	4%	-
Sr-90	2	0.89	0.89	0.90	0.88	2%	Sr-90: 45°/0°
Sr-90/45°	2	0.58	0.58	0.61	0.54	10%	
beta all	6	0.58	0.53	0.90	0.13	64%	-
N-30	2	1.31	1.31	1.33	1.28	3%	N-30: 60°/0°
N-30/60°	2	1.27	1.27	1.27	1.27	0%	
W-80-L	4	1.18	1.17	1.20	1.14	2%	W-80: H/L
W-80-M	2	1.17	1.17	1.19	1.15	2%	
W-80-H	2	1.21	1.21	1.27	1.16	6%	-
W-110	2	0.96	0.96	0.97	0.95	1%	-
Cs-137	2	0.79	0.79	0.80	0.78	1%	-
photon all	16	1.18	1.13	1.33	0.78	15%	-
All	22	1.15	0.97	1.33	0.13	36%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 51 : Extremity - photon/beta dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	22	4.51	0.73	0.16	outlier
		26	4.51	0.69	0.15	outlier
	Sr-90	4	3.00	2.50	0.83	OK
		13	3.00	2.44	0.81	OK
photon	Sr-90/45°	1	2.80	1.90	0.68	OK
		17	2.80	1.91	0.68	OK
	N-30	30	4.80	5.99	1.25	OK
		31	4.80	5.99	1.25	OK
	N-30/60°	24	4.70	6.53	1.39	OK
		25	4.70	6.25	1.33	OK
	W-80-L	18	4.50	4.94	1.10	OK
		23	4.50	5.28	1.17	OK
		27	4.50	5.36	1.19	OK
		29	4.50	5.37	1.19	OK
	W-80-M	8	48.0	58.39	1.22	OK
		10	48.0	56.11	1.17	OK
	W-80-H	9	350	429.40	1.23	OK
		11	350	443.27	1.27	OK
	W-110	20	5.40	5.31	0.98	OK
		28	5.40	5.58	1.03	OK
	Cs-137	5	4.30	3.51	0.82	OK
		6	4.30	3.45	0.80	OK
		2				
	NIR	3				
	NIR	7				
	NIR	12				
	NIR	14				
	NIR	15				
	NIR	16				
	NIR	19				
	NIR	21				
	NIR	32				

Legend for Quality
 L low dose
 M medium dose
 H high dose
 NIR not irradiated
 WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.16	0.16	0.16	0.15	4%	-
Sr-90	2	0.82	0.82	0.83	0.81	2%	Sr-90: 45°/0°
Sr-90/45°	2	0.68	0.68	0.68	0.68	0%	0.83
beta all	6	0.68	0.55	0.83	0.15	57%	-
N-30	2	1.25	1.25	1.25	1.25	0%	N-30: 60°/0°
N-30/60°	2	1.36	1.36	1.39	1.33	3%	1.09
W-80-L	4	1.18	1.16	1.19	1.10	4%	W-80: H/L
W-80-M	2	1.19	1.19	1.22	1.17	3%	1.07
W-80-H	2	1.25	1.25	1.27	1.23	2%	
W-110	2	1.01	1.01	1.03	0.98	4%	-
Cs-137	2	0.81	0.81	0.82	0.80	1%	-
photon all	16	1.19	1.15	1.39	0.80	14%	-
All	22	1.13	0.99	1.39	0.15	35%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 52 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	5	4.51	0.46	0.10	(outlier)
		7	4.51	0.52	0.12	(outlier)
	Sr-90	6	3.00	2.62	0.87	(OK)
		25	3.00	2.29	0.76	(OK)
	Sr-90/45°	13	2.80	1.35	0.48	(OK)
photon	N-30	17	4.40	5.13	1.17	OK
		22	4.40	5.06	1.15	OK
	N-30/60°	9	5.20	5.63	1.08	OK
		10	5.20	5.78	1.11	OK
	W-80-L	12	4.50	4.59	1.02	OK
		14	4.50	5.59	1.24	OK
		23	4.50	4.73	1.05	OK
		24	4.50	5.06	1.12	OK
	W-80-M	26	56.0	59.01	1.05	OK
		27	56.0	61.06	1.09	OK
	W-80-H	30	324	354.57	1.09	OK
		31	324	352.60	1.09	OK
	W-110	19	4.80	4.03	0.84	OK
		20	4.80	4.34	0.90	OK
	Cs-137	4	4.70	3.78	0.80	OK
		8	5.30	3.97	0.75	OK
	NIR	1				
	NIR	2				
	NIR	3				
	NIR	11				
	NIR	16				
	NIR	18				
	NIR	21				
	NIR	28				
	NIR	29				
	NIR	32				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.11	0.11	0.12	0.10	9%	-
Sr-90	2	0.82	0.82	0.87	0.76	10%	Sr-90: 45°/0°
Sr-90/45°	2	0.48	0.48	0.48	0.48	1%	0.58
beta all	6	0.48	0.47	0.87	0.10	68%	-
N-30	2	1.16	1.16	1.17	1.15	1%	
N-30/60°	2	1.10	1.10	1.11	1.08	2%	0.95
W-80-L	4	1.09	1.11	1.24	1.02	9%	
W-80-M	2	1.07	1.07	1.09	1.05	2%	
W-80-H	2	1.09	1.09	1.09	1.09	0%	
W-110	2	0.87	0.87	0.90	0.84	5%	-
Cs-137	2	0.78	0.78	0.80	0.75	5%	-
photon all	16	1.09	1.04	1.24	0.75	13%	-
All	22	1.04	0.88	1.24	0.10	37%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 53 : Extremity - photon dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	15 24	4.51 4.51	0.47 0.48	0.10 0.11	(outlier) (outlier)
	Sr-90	6 18	3.00 3.00	2.64 2.59	0.88 0.86	(OK) (OK)
	Sr-90/45°	19 27	2.80 2.80	1.25 1.38	0.45 0.49	(OK) (OK)
	N-30	21 22	5.30 5.30	6.96 7.35	1.31 1.39	OK OK
	N-30/60°	13 14	5.40 5.40	6.29 6.58	1.16 1.22	OK OK
photon	W-80-L	2 4 29 30	4.90 4.90 4.90 4.90	5.84 5.76 5.43 5.63	1.19 1.18 1.11 1.15	OK OK OK OK
		28 31	49.0 49.0	55.40 56.10	1.13 1.14	OK OK
		3 7	330 330	417.00 417.00	1.26 1.26	OK OK
		1 5	4.60 4.60	4.36 4.58	0.95 1.00	OK OK
	W-110	16 20	4.60 4.60	4.02 3.70	0.87 0.80	OK OK
	Cs-137	8 9 10 11 12 17 23 25 26 32				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
						beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.11	0.11	0.11	0.10	1%	-
Sr-90	2	0.87	0.87	0.88	0.86	1%	Sr-90: 45°/0°
Sr-90/45°	2	0.47	0.47	0.49	0.45	7%	0.54
beta all	6	0.47	0.48	0.88	0.10	71%	-
N-30	2	1.35	1.35	1.39	1.31	4%	N-30: 60°/0°
N-30/60°	2	1.19	1.19	1.22	1.16	3%	0.88
W-80-L	4	1.16	1.16	1.19	1.11	3%	W-80: H/L
W-80-M	2	1.14	1.14	1.14	1.13	1%	1.09
W-80-H	2	1.26	1.26	1.26	1.26	0%	
W-110	2	0.97	0.97	1.00	0.95	3%	-
Cs-137	2	0.84	0.84	0.87	0.80	6%	-
photon all	16	1.16	1.13	1.39	0.80	14%	-
All	22	1.12	0.96	1.39	0.10	38%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 54 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	20 24	4.51 4.51	0.55 0.59	0.12 0.13	outlier outlier
	Sr-90	6 17	3.00 3.00	2.08 2.29	0.69 0.76	OK OK
	Sr-90/45°	3 7	2.80 2.80	1.15 1.57	0.41 0.56	OK OK
	N-30	21 23	5.00 5.00	5.39 6.16	1.08 1.23	OK OK
photon	N-30/60°	31 32	4.90 4.90	6.02 5.60	1.23 1.14	OK OK
	W-80-L	2 5 11 18	5.60 5.60 5.60 5.60	5.48 5.67 6.28 5.52	0.98 1.01 1.12 0.99	OK OK OK OK
		16 19	43.0 43.0	50.05 49.08	1.16 1.14	OK OK
		8 9	375 375	424.32 409.82	1.13 1.09	OK OK
	W-110	29 30	4.40 4.40	3.85 3.61	0.88 0.82	OK OK
	Cs-137	13 14	5.60 5.60	3.96 3.79	0.71 0.68	OK OK
	NIR	1				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	4				
	NIR	10				
	NIR	12				
	NIR	15				
	NIR	22				
	NIR	25				
	NIR	26				
	NIR	27				
	NIR	28				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.13	0.13	0.13	0.12	5%	-
Sr-90	2	0.73	0.73	0.76	0.69	7%	Sr-90: 45°/0° 0.67
Sr-90/45°	2	0.49	0.49	0.56	0.41	22%	
beta all	6	0.49	0.45	0.76	0.12	62%	-
N-30	2	1.16	1.16	1.23	1.08	9%	N-30: 60°/0° 1.03
N-30/60°	2	1.19	1.19	1.23	1.14	5%	
W-80-L	4	1.00	1.02	1.12	0.98	6%	W-80: H/L 1.09
W-80-M	2	1.15	1.15	1.16	1.14	1%	
W-80-H	2	1.11	1.11	1.13	1.09	2%	
W-110	2	0.85	0.85	0.88	0.82	5%	-
Cs-137	2	0.69	0.69	0.71	0.68	3%	-
photon all	16	1.09	1.02	1.23	0.68	17%	-
All	22	0.98	0.87	1.23	0.12	38%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 55 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	23 25	4.51 4.51	4.10 4.68	0.91 1.04	OK OK
	Sr-90	7 16	3.00 3.00	3.16 3.01	1.05 1.00	OK OK
	Sr-90/45°	8 18	2.80 2.80	1.65 1.62	0.59 0.58	OK OK
	N-30	1 2	5.50 5.50	5.36 5.35	0.97 0.97	OK OK
photon	N-30/60°	28 29	4.60 4.60	4.24 4.18	0.92 0.91	OK OK
	W-80-L	17 20 30 31	5.70 5.70 5.70 5.70	6.46 6.31 6.51 6.44	1.13 1.11 1.14 1.13	OK OK OK OK
		9 10	56.0 56.0	65.06 63.29	1.16 1.13	OK OK
		12 13	330 330	397.19 398.20	1.20 1.21	OK OK
	W-110	6 11	4.80 4.80	4.66 4.66	0.97 0.97	OK OK
	Cs-137	4 5	4.80 4.80	3.90 3.82	0.81 0.80	OK OK
	NIR	3				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	14				
	NIR	15				
	NIR	19				
	NIR	21				
	NIR	22				
	NIR	24				
	NIR	26				
	NIR	27				
	NIR	32				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.97	0.97	1.04	0.91	9%	-
Sr-90	2	1.03	1.03	1.05	1.00	3%	Sr-90: 45°/0°
Sr-90/45°	2	0.58	0.58	0.59	0.58	1%	
beta all	6	0.96	0.86	1.05	0.58	26%	-
N-30	2	0.97	0.97	0.97	0.97	0%	N-30: 60°/0°
N-30/60°	2	0.92	0.92	0.92	0.91	1%	
W-80-L	4	1.13	1.13	1.14	1.11	1%	W-80: H/L
W-80-M	2	1.15	1.15	1.16	1.13	2%	
W-80-H	2	1.21	1.21	1.21	1.20	0%	
W-110	2	0.97	0.97	0.97	0.97	0%	-
Cs-137	2	0.80	0.80	0.81	0.80	1%	-
photon all	16	1.04	1.03	1.21	0.80	13%	-
All	22	0.99	0.99	1.21	0.58	18%	-

outliers: 0 of 22

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 56 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	30 31	4.51 4.51	0.43 0.44	0.10 0.10	outlier outlier
	Sr-90	13 29	3.00 3.00	2.77 2.83	0.92 0.94	OK OK
	Sr-90/45°	18 22	2.80 2.80	1.27 1.27	0.45 0.45	OK OK
	N-30	14 16	5.00 5.00	8.04 6.89	1.61 1.38	OK OK
	N-30/60°	7 8	4.60 4.60	6.50 6.73	1.41 1.46	OK OK
photon	W-80-L	11 12 26 32	4.50 4.50 4.50 4.50	6.94 5.60 6.39 6.37	1.54 1.24 1.42 1.42	OK OK OK OK
		1 2	51.0 51.0	71.74 68.91	1.41 1.35	OK OK
		23 25	371 371	561.27 533.07	1.51 1.44	OK OK
		24 27	5.20 5.20	5.58 6.22	1.07 1.20	OK OK
	W-110	5 6	4.50 4.50	4.21 4.21	0.94 0.94	OK OK
	NIR	3 4 9 10 15 17 19 20 21 28				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.10	0.10	0.10	0.10	2%	-
Sr-90	2	0.93	0.93	0.94	0.92	2%	Sr-90: 45°/0°
Sr-90/45°	2	0.45	0.45	0.45	0.45	0%	0.49
beta all	6	0.45	0.49	0.94	0.10	76%	-
N-30	2	1.49	1.49	1.61	1.38	11%	N-30: 60°/0°
N-30/60°	2	1.44	1.44	1.46	1.41	2%	0.96
W-80-L	4	1.42	1.41	1.54	1.24	9%	W-80: H/L
W-80-M	2	1.38	1.38	1.41	1.35	3%	1.05
W-80-H	2	1.47	1.47	1.51	1.44	4%	
W-110	2	1.13	1.13	1.20	1.07	8%	-
Cs-137	2	0.94	0.94	0.94	0.94	0%	-
photon all	16	1.41	1.33	1.61	0.94	15%	-
All	22	1.30	1.10	1.61	0.10	41%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 57 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)		
beta	Kr-85	4	4.51	0.33	0.07	(outlier)	
		30	4.51	0.51	0.11	(outlier)	
	Sr-90	8	3.00	1.91	0.64	(OK)	
		23	3.00	2.04	0.68	(OK)	
	Sr-90/45°	21	2.80	1.19	0.43	(OK)	
photon		22	2.80	1.06	0.38	(outlier)	
	N-30	15	4.90	4.79	0.98	OK	
		16	4.90	4.83	0.99	OK	
	N-30/60°	5	5.10	4.93	0.97	OK	
		7	5.10	4.99	0.98	OK	
		10	5.70	5.59	0.98	OK	
	W-80-L	11	5.70	5.34	0.94	OK	
		26	5.70	5.68	1.00	OK	
		27	5.70	5.30	0.93	OK	
	W-80-M	17	54.0	52.38	0.97	OK	
		31	54.0	51.01	0.94	OK	
	W-80-H	1	344	322.94	0.94	OK	
		24	344	340.14	0.99	OK	
	W-110	28	5.20	4.37	0.84	OK	
		29	5.20	4.75	0.91	OK	
	Cs-137	2	4.80	3.47	0.72	OK	
		3	4.80	3.73	0.78	OK	
	NIR	6				Legend for Quality	
	NIR	9				L low dose	
	NIR	12				M medium dose	
	NIR	13				H high dose	
	NIR	14				NIR not irradiated	
	NIR	18				WIR wrongly irradiated	
	NIR	19					
	NIR	20					
	NIR	25					
	NIR	32					

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)	
Kr-85	2	0.09	0.09	0.11	0.07	30%	-	
Sr-90	2	0.66	0.66	0.68	0.64	5%	Sr-90: 45°/0°	
Sr-90/45°	2	0.40	0.40	0.43	0.38	8%	0.61	
beta all	6	0.40	0.38	0.68	0.07	66%	-	
N-30	2	0.98	0.98	0.99	0.98	1%	N-30: 60°/0°	
N-30/60°	2	0.97	0.97	0.98	0.97	1%	0.99	
W-80-L	4	0.96	0.96	1.00	0.93	3%	W-80: H/L	
W-80-M	2	0.96	0.96	0.97	0.94	2%	1.00	
W-80-H	2	0.96	0.96	0.99	0.94	4%		
W-110	2	0.88	0.88	0.91	0.84	6%	-	
Cs-137	2	0.75	0.75	0.78	0.72	5%	-	
photon all	16	0.96	0.93	1.00	0.72	9%	-	
All	22	0.93	0.78	1.00	0.07	37%	-	

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 58 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	4	4.51	0.72	0.16	(outlier)
		24	4.51	0.80	0.18	(outlier)
	Sr-90	19	3.00	4.48	1.49	(OK)
		31	3.00	4.23	1.41	(OK)
	Sr-90/45°	26	2.80	2.80	1.00	(OK)
		32	2.80	3.08	1.10	(OK)
photon	N-30	8	4.80	4.64	0.97	OK
		9	4.80	4.97	1.04	OK
	N-30/60°	11	5.00	4.63	0.93	OK
		14	5.00	4.84	0.97	OK
	W-80-L	20	5.00	4.87	0.97	OK
		22	5.00	5.23	1.05	OK
		29	5.00	4.88	0.98	OK
		30	5.00	4.96	0.99	OK
	W-80-M	16	57.0	52.46	0.92	OK
		17	57.0	55.30	0.97	OK
	W-80-H	21	320	287.94	0.90	OK
		23	320	278.62	0.87	OK
	W-110	27	4.70	4.38	0.93	OK
		28	4.70	4.98	1.06	OK
	Cs-137	3	5.60	7.53	1.34	OK
		5	5.60	7.37	1.32	OK
	NIR	1				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	2				
	NIR	6				
	NIR	7				
	NIR	10				
	NIR	12				
	NIR	13				
	NIR	15				
	NIR	18				
	NIR	25				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.17	0.17	0.18	0.16	7%	-
Sr-90	2	1.45	1.45	1.49	1.41	4%	Sr-90: 45°/0°
Sr-90/45°	2	1.05	1.05	1.10	1.00	7%	0.72
beta all	6	1.05	0.89	1.49	0.16	66%	-
N-30	2	1.00	1.00	1.04	0.97	5%	N-30: 60°/0°
N-30/60°	2	0.95	0.95	0.97	0.93	3%	
W-80-L	4	0.98	1.00	1.05	0.97	3%	W-80: H/L
W-80-M	2	0.95	0.95	0.97	0.92	4%	
W-80-H	2	0.89	0.89	0.90	0.87	2%	
W-110	2	1.00	1.00	1.06	0.93	9%	-
Cs-137	2	1.33	1.33	1.34	1.32	2%	-
photon all	16	0.97	1.01	1.34	0.87	13%	-
All	22	0.98	0.98	1.49	0.16	32%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 59 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	8	4.51	0.43	0.10	(outlier)
		10	4.51	0.39	0.09	(outlier)
	Sr-90	20	3.00	4.80	1.60	(OK)
		27	3.00	4.77	1.59	(OK)
	Sr-90/45°	15	2.80	2.49	0.89	(OK)
photon		19	2.80	3.17	1.13	(OK)
	N-30	21	4.40	5.54	1.26	OK
		22	4.40	5.43	1.23	OK
	N-30/60°	16	5.40	6.20	1.15	OK
		17	5.40	6.39	1.18	OK
		24	5.00	6.57	1.31	OK
	W-80-L	25	5.00	6.41	1.28	OK
		30	5.00	6.32	1.26	OK
		31	5.00	6.47	1.29	OK
	W-80-M	12	43.0	55.98	1.30	OK
		23	43.0	56.58	1.32	OK
	W-80-H	26	329	428.58	1.30	OK
		28	329	430.08	1.31	OK
	W-110	11	5.10	6.32	1.24	OK
		13	5.10	6.53	1.28	OK
	Cs-137	2	5.10	8.19	1.61	OK
		4	5.10	8.10	1.59	OK
	NIR	1				
	NIR	3				
	NIR	5				
	NIR	6				
	NIR	7				
	NIR	9				
	NIR	14				
	NIR	18				
	NIR	29				
	NIR	32				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.09	0.09	0.10	0.09	7%	-
Sr-90	2	1.60	1.60	1.60	1.59	0%	Sr-90: 45°/0°
Sr-90/45°	2	1.01	1.01	1.13	0.89	17%	0.63
beta all	6	1.01	0.90	1.60	0.09	76%	-
N-30	2	1.25	1.25	1.26	1.23	1%	N-30: 60°/0°
N-30/60°	2	1.17	1.17	1.18	1.15	2%	0.94
W-80-L	4	1.29	1.29	1.31	1.26	2%	W-80: H/L
W-80-M	2	1.31	1.31	1.32	1.30	1%	1.01
W-80-H	2	1.30	1.30	1.31	1.30	0%	
W-110	2	1.26	1.26	1.28	1.24	2%	-
Cs-137	2	1.60	1.60	1.61	1.59	1%	-
photon all	16	1.29	1.31	1.61	1.15	9%	-
All	22	1.28	1.20	1.61	0.09	33%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 60 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	5 6	4.51 4.51	0.30 0.30	0.07 0.07	outlier outlier
	Sr-90	2 3	3.00 3.00	3.06 3.11	1.02 1.04	OK OK
	Sr-90/45°	4 12	2.80 2.80	1.77 1.78	0.63 0.64	OK OK
	N-30	20 22	5.60 5.60	4.47 4.20	0.80 0.75	OK OK
photon	N-30/60°	24 27	3.00 5.50	4.30 4.38	1.43 0.80	OK OK
	W-80-L	13 14 21 23	5.30 5.30 5.30 5.50	4.48 4.47 4.37 4.75	0.85 0.84 0.83 0.86	OK OK OK OK
		1 7	45.0 45.0	37.81 38.91	0.84 0.86	OK OK
		15 16	321 321	272.71 266.71	0.85 0.83	OK OK
	W-110	30 32	4.30 4.30	3.49 3.46	0.81 0.81	OK OK
	Cs-137	10 11	4.80 4.80	4.99 4.97	1.04 1.04	OK OK
	NIR NIR NIR NIR NIR NIR NIR NIR NIR NIR	8 9 17 18 19 25 26 28 29 31			Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated	

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.07	0.07	0.07	0.07	0%	-
Sr-90	2	1.03	1.03	1.04	1.02	1%	Sr-90: 45°/0°
Sr-90/45°	2	0.64	0.64	0.64	0.63	0%	0.62
beta all	6	0.64	0.58	1.04	0.07	75%	-
N-30	2	0.77	0.77	0.80	0.75	4%	N-30: 60°/0°
N-30/60°	2	1.12	1.12	1.43	0.80	40%	1.44
W-80-L	4	0.84	0.84	0.86	0.83	2%	W-80: H/L
W-80-M	2	0.85	0.85	0.86	0.84	2%	0.99
W-80-H	2	0.84	0.84	0.85	0.83	2%	
W-110	2	0.81	0.81	0.81	0.81	1%	-
Cs-137	2	1.04	1.04	1.04	1.04	0%	-
photon all	16	0.84	0.89	1.43	0.75	19%	-
All	22	0.84	0.80	1.43	0.07	36%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 61: Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)		
beta	Kr-85	16 25	4.51 4.51	0.10 0.12	0.02 0.03	(outlier) (outlier)	
	Sr-90	14 28	3.00 3.00	0.67 0.66	0.22 0.22	(outlier) (outlier)	
	Sr-90/45°	18 21	2.80 2.80	0.29 0.35	0.10 0.13	(outlier) (outlier)	
photon	N-30	27 29	5.70 5.70	4.48 4.52	0.79 0.79	OK OK	
	N-30/60°	12 13	5.70 5.70	3.63 3.94	0.64 0.69	OK OK	
	W-80-L	1 3 20 22	5.60 5.60 5.60 5.60	6.11 5.72 6.09 5.99	1.09 1.02 1.09 1.07	OK OK OK OK	
		9 10	43.0 43.0	47.48 49.52	1.10 1.15	OK OK	
		4 5	315 315	357.23 320.61	1.13 1.02	OK OK	
		17 19	4.70 4.70	5.19 5.03	1.10 1.07	OK OK	
	Cs-137	6 7	5.70 5.70	4.93 4.81	0.86 0.84	OK OK	
	NIR	2				Legend for Quality	
	NIR	8				L low dose	
	NIR	11				M medium dose	
	NIR	15				H high dose	
	NIR	23				NIR not irradiated	
	NIR	24				WIR wrongly irradiated	
	NIR	26					
	NIR	30					
	NIR	31					
	NIR	32					

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)	
Kr-85	2	0.02	0.02	0.03	0.02	13%	-	
Sr-90	2	0.22	0.22	0.22	0.22	1%	Sr-90: 45°/0°	
Sr-90/45°	2	0.11	0.11	0.13	0.10	13%	0.52	
beta all	6	0.11	0.12	0.22	0.02	74%	-	
N-30	2	0.79	0.79	0.79	0.79	1%	N-30: 60°/0°	
N-30/60°	2	0.66	0.66	0.69	0.64	6%	0.84	
W-80-L	4	1.08	1.07	1.09	1.02	3%	W-80: H/L	
W-80-M	2	1.13	1.13	1.15	1.10	3%	1.01	
W-80-H	2	1.08	1.08	1.13	1.02	8%		
W-110	2	1.09	1.09	1.10	1.07	2%	-	
Cs-137	2	0.85	0.85	0.86	0.84	2%	-	
photon all	16	1.05	0.97	1.15	0.64	18%	-	
All	22	0.85	0.74	1.15	0.02	56%	-	

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 62 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	1 27	4.51 4.51	0.80 0.71	0.18 0.16	outlier outlier
	Sr-90	13 16	3.00 3.00	4.54 4.93	1.51 1.64	OK OK
	Sr-90/45°	3 8	2.80 2.80	2.70 3.35	0.96 1.20	OK OK
	N-30	14 15	5.00 5.00	4.80 4.65	0.96 0.93	OK OK
photon	N-30/60°	30 31	5.50 5.50	6.69 7.01	1.22 1.27	OK OK
	W-80-L	4 9 20 26	4.90 4.90 4.90 4.90	6.04 4.92 5.77 5.01	1.23 1.00 1.18 1.02	OK OK OK OK
		19 29	53.0 53.0	53.13 69.37	1.00 1.31	OK OK
		21 22	350 350	381.04 445.87	1.09 1.27	OK OK
	W-110	2 18	5.10 5.10	5.53 5.16	1.08 1.01	OK OK
	Cs-137	6 7	4.70 4.70	7.17 6.92	1.53 1.47	OK OK
	NIR	5				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	10				
	NIR	11				
	NIR	12				
	NIR	17				
	NIR	23				
	NIR	24				
	NIR	25				
	NIR	28				
	NIR	32				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.17	0.17	0.18	0.16	8%	-
Sr-90	2	1.58	1.58	1.64	1.51	6%	Sr-90: 45°/0°
Sr-90/45°	2	1.08	1.08	1.20	0.96	15%	0.68
beta all	6	1.08	0.94	1.64	0.16	69%	-
N-30	2	0.95	0.95	0.96	0.93	2%	N-30: 60°/0°
N-30/60°	2	1.25	1.25	1.27	1.22	3%	
W-80-L	4	1.10	1.11	1.23	1.00	10%	W-80: H/L
W-80-M	2	1.16	1.16	1.31	1.00	19%	
W-80-H	2	1.18	1.18	1.27	1.09	11%	
W-110	2	1.05	1.05	1.08	1.01	5%	-
Cs-137	2	1.50	1.50	1.53	1.47	3%	-
photon all	16	1.13	1.16	1.53	0.93	15%	-
All	22	1.13	1.10	1.64	0.16	33%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 63 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	14 31	4.51 4.51	0.58 0.52	0.13 0.12	outlier outlier
	Sr-90	22 30	3.00 3.00	3.66 3.59	1.22 1.20	OK OK
	Sr-90/45°	13 32	2.80 2.80	1.81 2.01	0.65 0.72	OK OK
	N-30	27 28	4.90 4.90	3.75 4.10	0.77 0.84	OK OK
photon	N-30/60°	1 3	4.30 4.30	3.71 4.33	0.86 1.01	OK OK
	W-80-L	2 7 20	5.70 5.70 5.70	5.03 5.08 5.46	0.88 0.89 0.96	OK OK OK
		21	5.70	4.77	0.84	OK
	W-80-M	15 17	47.0 47.0	43.76 41.96	0.93 0.89	OK OK
	W-80-H	18 19	362 362	340.72 343.20	0.94 0.95	OK OK
	W-110	8 10	5.60 5.60	4.63 4.99	0.83 0.89	OK OK
	Cs-137	4 5	5.30 5.30	6.05 5.62	1.14 1.06	OK OK
	NIR	6				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	9				
	NIR	11				
	NIR	12				
	NIR	16				
	NIR	23				
	NIR	24				
	NIR	25				
	NIR	26				
	NIR	29				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.12	0.12	0.13	0.12	8%	-
Sr-90	2	1.21	1.21	1.22	1.20	1%	Sr-90: 45°/0°
Sr-90/45°	2	0.68	0.68	0.72	0.65	7%	0.56
beta all	6	0.68	0.67	1.22	0.12	73%	-
N-30	2	0.80	0.80	0.84	0.77	6%	N-30: 60°/0°
N-30/60°	2	0.93	0.93	1.01	0.86	11%	1.17
W-80-L	4	0.89	0.89	0.96	0.84	6%	W-80: H/L
W-80-M	2	0.91	0.91	0.93	0.89	3%	1.06
W-80-H	2	0.94	0.94	0.95	0.94	1%	
W-110	2	0.86	0.86	0.89	0.83	5%	-
Cs-137	2	1.10	1.10	1.14	1.06	5%	-
photon all	16	0.89	0.92	1.14	0.77	10%	-
All	22	0.89	0.85	1.22	0.12	32%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 64 : Extremity - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)		
beta	Kr-85	20	4.51	0.49	0.11	(outlier)	
		26	4.51	0.16	0.04	(outlier)	
	Sr-90	10	3.00	0.17	0.06	(outlier)	
		22	3.00	0.56	0.19	(outlier)	
	Sr-90/45°	24	2.80	0.17	0.06	(outlier)	
photon	N-30	13	4.60	4.33	0.94	OK	
		14	4.60	3.47	0.75	OK	
	N-30/60°	27	5.20	3.62	0.70	OK	
		32	5.20	2.73	0.53	outlier	
	W-80-L	12	5.70	7.64	1.34	OK	
		15	5.70	5.65	0.99	OK	
		29	5.70	4.68	0.82	OK	
		30	5.70	5.77	1.01	OK	
	W-80-M	7	46.0	61.43	1.34	OK	
		8	46.0	63.73	1.39	OK	
	W-80-H	5	382	565.40	1.48	OK	
		6	382	481.94	1.26	OK	
	W-110	16	4.70	5.54	1.18	OK	
		17	4.70	5.89	1.25	OK	
	Cs-137	3	5.30	6.08	1.15	OK	
		4	5.30	4.91	0.93	OK	
	NIR	1				Legend for Quality	
	NIR	2				L low dose	
	NIR	9				M medium dose	
	NIR	11				H high dose	
	NIR	18				NIR not irradiated	
	NIR	19				WIR wrongly irradiated	
	NIR	21					
	NIR	23					
	NIR	28					
	NIR	31					

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)	
Kr-85	2	0.07	0.07	0.11	0.04	72%	-	
Sr-90	2	0.12	0.12	0.19	0.06	76%	Sr-90: 45°/0°	
Sr-90/45°	2	0.06	0.06	0.06	0.05	9%	0.47	
beta all	6	0.06	0.08	0.19	0.04	67%	-	
N-30	2	0.85	0.85	0.94	0.75	16%	N-30: 60°/0°	
N-30/60°	2	0.61	0.61	0.70	0.53	20%	0.72	
W-80-L	4	1.00	1.04	1.34	0.82	21%	W-80: H/L	
W-80-M	2	1.36	1.36	1.39	1.34	3%	1.32	
W-80-H	2	1.37	1.37	1.48	1.26	11%		
W-110	2	1.22	1.22	1.25	1.18	4%	-	
Cs-137	2	1.04	1.04	1.15	0.93	15%	-	
photon all	16	1.08	1.07	1.48	0.53	26%	-	
All	22	0.93	0.80	1.48	0.04	63%	-	

outliers: 1 of 16

fraction of outliers: 6%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 65 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	20 21	4.51 4.51	4.34 4.04	0.96 0.90	OK OK
	Sr-90	9 19	3.00 3.00	4.49 4.04	1.50 1.35	OK OK
	Sr-90/45°	15 25	2.80 2.80	3.97 3.98	1.42 1.42	OK OK
	N-30	18 23	4.80 4.80	6.33 5.92	1.32 1.23	OK OK
	N-30/60°	1 26	5.60 5.60	7.10 6.86	1.27 1.23	OK OK
photon	W-80-L	4 5 14 16	4.70 4.70 4.70 4.70	7.30 7.31 7.16 7.60	1.55 1.56 1.52 1.62	OK OK OK OK
		31 32	53.0 53.0	82.17 81.42	1.55 1.54	OK OK
		10 11	370 370	580.60 617.00	1.57 1.67	OK OK
		3 7	4.30 4.30	5.46 5.74	1.27 1.33	OK OK
	W-110	8 12	5.70 5.70	7.27 7.28	1.28 1.28	OK OK
	Cs-137	2 6 13 17 28 29 30 22 24 27				
						Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.93	0.93	0.96	0.90	5%	-
Sr-90	2	1.42	1.42	1.50	1.35	7%	Sr-90: 45°/0°
Sr-90/45°	2	1.42	1.42	1.42	1.42	0%	1.00
beta all	6	1.38	1.26	1.50	0.90	21%	-
N-30	2	1.28	1.28	1.32	1.23	5%	N-30: 60°/0°
N-30/60°	2	1.25	1.25	1.27	1.23	2%	0.98
W-80-L	4	1.55	1.56	1.62	1.52	3%	W-80: H/L
W-80-M	2	1.54	1.54	1.55	1.54	1%	1.04
W-80-H	2	1.62	1.62	1.67	1.57	4%	
W-110	2	1.30	1.30	1.33	1.27	4%	-
Cs-137	2	1.28	1.28	1.28	1.28	0%	-
photon all	16	1.43	1.42	1.67	1.23	11%	-
All	22	1.38	1.38	1.67	0.90	14%	-

outliers: 0 of 22

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 66 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	13 29	4.51 4.51	0.49 0.34	0.11 0.07	outlier outlier
	Sr-90	12 20	3.00 3.00	3.82 3.77	1.27 1.26	OK OK
	Sr-90/45°	17 18	2.80 2.80	2.95 2.36	1.05 0.84	OK OK
	N-30	27 28	4.90 4.90	6.25 6.10	1.27 1.25	OK OK
photon	N-30/60°	19 21	5.00 5.00	6.42 5.87	1.28 1.17	OK OK
	W-80-L	6 10 15 16	5.60 5.60 5.60 5.60	7.34 7.45 7.49 7.61	1.31 1.33 1.34 1.36	OK OK OK OK
		31 32	52.0 52.0	70.14 69.97	1.35 1.35	OK OK
		3 4	329 329	423.99 408.54	1.29 1.24	OK OK
	W-110	22 23	5.40 5.40	5.81 5.65	1.08 1.05	OK OK
	Cs-137	8 9	4.40 4.40	5.33 5.42	1.21 1.23	OK OK
	NIR	1				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	2				
	NIR	5				
	NIR	7				
	NIR	11				
	NIR	14				
	NIR	24				
	NIR	25				
	NIR	26				
	NIR	30				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.09	0.09	0.11	0.07	26%	-
Sr-90	2	1.27	1.27	1.27	1.26	1%	Sr-90: 45°/0° 0.75
Sr-90/45°	2	0.95	0.95	1.05	0.84	16%	
beta all	6	0.95	0.77	1.27	0.07	71%	-
N-30	2	1.26	1.26	1.27	1.25	2%	N-30: 60°/0° 0.97
N-30/60°	2	1.23	1.23	1.28	1.17	6%	
W-80-L	4	1.33	1.33	1.36	1.31	2%	W-80: H/L 0.95
W-80-M	2	1.35	1.35	1.35	1.35	0%	
W-80-H	2	1.27	1.27	1.29	1.24	3%	
W-110	2	1.06	1.06	1.08	1.05	2%	-
Cs-137	2	1.22	1.22	1.23	1.21	1%	-
photon all	16	1.28	1.26	1.36	1.05	7%	-
All	22	1.25	1.12	1.36	0.07	32%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv


Reporting number 67 : Extremity - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	2	4.51	1.79	0.40	outlier
		17	4.51	1.58	0.35	outlier
	Sr-90	1	3.00	2.98	0.99	OK
		32	3.00	2.89	0.96	OK
photon	Sr-90/45°	28	2.80	2.45	0.88	OK
		30	2.80	2.44	0.87	OK
	N-30	8	4.80	4.48	0.93	OK
		9	4.80	4.86	1.01	OK
	N-30/60°	22	4.60	4.07	0.88	OK
		24	4.60	4.00	0.87	OK
	W-80-L	5	5.20	4.22	0.81	OK
		6	5.20	4.28	0.82	OK
		27	5.20	4.89	0.94	OK
		29	5.20	4.52	0.87	OK
	W-80-M	10	47.0	44.27	0.94	OK
		11	47.0	44.62	0.95	OK
	W-80-H	15	341	328.30	0.96	OK
		16	341	286.80	0.84	OK
	W-110	20	4.60	2.71	0.59	OK
		21	4.60	3.13	0.68	OK
	Cs-137	3	5.80	4.91	0.85	OK
		4	5.80	5.05	0.87	OK
	NIR	7				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	12				
	NIR	13				
	NIR	14				
	NIR	18				
	NIR	19				
	NIR	23				
	NIR	25				
	NIR	26				
	NIR	31				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.37	0.37	0.40	0.35	9%	-
Sr-90	2	0.98	0.98	0.99	0.96	2%	Sr-90: 45°/0°
Sr-90/45°	2	0.87	0.87	0.88	0.87	0%	0.89
beta all	6	0.87	0.74	0.99	0.35	39%	-
N-30	2	0.97	0.97	1.01	0.93	6%	N-30: 60°/0°
N-30/60°	2	0.88	0.88	0.88	0.87	1%	0.90
W-80-L	4	0.85	0.86	0.94	0.81	7%	W-80: H/L
W-80-M	2	0.95	0.95	0.95	0.94	1%	1.05
W-80-H	2	0.90	0.90	0.96	0.84	10%	
W-110	2	0.63	0.63	0.68	0.59	10%	-
Cs-137	2	0.86	0.86	0.87	0.85	2%	-
photon all	16	0.87	0.86	1.01	0.59	12%	-
All	22	0.87	0.83	1.01	0.35	21%	-

outliers: 2 of 22

fraction of outliers: 9%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 68 : Extremity - photon/beta dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(0.07)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Kr-85	8 20	4.51 4.51	3.40 3.80	0.75 0.84	OK OK
	Sr-90	7 17	3.00 3.00	3.60 3.50	1.20 1.17	OK OK
	Sr-90/45°	23 27	2.80 2.80	3.20 2.80	1.14 1.00	OK OK
	N-30	30 31	4.60 4.60	5.10 5.20	1.11 1.13	OK OK
photon	N-30/60°	22 24	5.60 5.60	5.90 6.10	1.05 1.09	OK OK
		3 9 14 15	5.00 5.00 5.00 5.00	5.70 6.00 5.90 5.80	1.14 1.20 1.18 1.16	OK OK OK OK
	W-80-L	12 13	47.0 47.0	56.10 54.70	1.19 1.16	OK OK
	W-80-M	10 11	327 327	384.50 393.10	1.18 1.20	OK OK
	W-80-H	18 19	5.70 5.70	5.50 5.60	0.96 0.98	OK OK
	W-110	18 19	5.70 5.70	5.50 5.60	0.96 0.98	OK OK
	Cs-137	5 6	5.50 5.50	5.70 5.80	1.04 1.05	OK OK
	NIR	1				
	NIR	2				
	NIR	4				
	NIR	16				
	NIR	21				
	NIR	25				
	NIR	26				
	NIR	28				
	NIR	29				
	NIR	32				

Legend for Quality
 L low dose
 M medium dose
 H high dose
 NIR not irradiated
 WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Kr-85	2	0.80	0.80	0.84	0.75	8%	-
Sr-90	2	1.18	1.18	1.20	1.17	2%	Sr-90: 45°/0°
Sr-90/45°	2	1.07	1.07	1.14	1.00	9%	0.91
beta all	6	1.07	1.02	1.20	0.75	18%	-
N-30	2	1.12	1.12	1.13	1.11	1%	N-30: 60°/0°
N-30/60°	2	1.07	1.07	1.09	1.05	2%	0.96
W-80-L	4	1.17	1.17	1.20	1.14	2%	W-80: H/L
W-80-M	2	1.18	1.18	1.19	1.16	2%	1.02
W-80-H	2	1.19	1.19	1.20	1.18	2%	
W-110	2	0.97	0.97	0.98	0.96	1%	-
Cs-137	2	1.05	1.05	1.05	1.04	1%	-
photon all	16	1.14	1.11	1.20	0.96	7%	-
All	22	1.14	1.09	1.20	0.75	11%	-

outliers: 0 of 22

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 1.0 mSv

Reporting number 69 : Eye - photon dosimeter



	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	11 31	2.20 2.20	2.19 1.66	1.00 0.75	(OK) (OK)
photon	N-30	7	4.50	2.91	0.65	OK
		8	4.50	3.80	0.84	OK
	N-30/60°	29	4.40	3.49	0.79	OK
		30	4.40	3.16	0.72	OK
	W-80-L	15	4.40	3.66	0.83	OK
		16	4.40	3.15	0.72	OK
		26	4.40	3.33	0.76	OK
		28	4.40	2.74	0.62	outlier
	W-80-M	12	43.0	32.84	0.76	OK
		13	43.0	31.61	0.74	OK
	W-80-H	1	324	283.46	0.87	OK
		2	324	216.09	0.67	outlier
	W-110	5	4.60	3.44	0.75	OK
		6	4.60	2.89	0.63	outlier
	Cs-137	24 25	5.20 5.20	4.27 4.86	0.82 0.93	OK OK
						Legend for Quality
						L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.88	0.88	1.00	0.75	20%	-
N-30	2	0.74	0.74	0.84	0.65	19%	N-30: 60°/0° 1.01
N-30/60°	2	0.76	0.76	0.79	0.72	7%	
W-80-L	4	0.74	0.73	0.83	0.62	12%	W-80: H/L 1.02
W-80-M	2	0.75	0.75	0.76	0.74	3%	
W-80-H	2	0.77	0.77	0.87	0.67	19%	
W-110	2	0.69	0.69	0.75	0.63	12%	-
Cs-137	2	0.88	0.88	0.93	0.82	9%	-
photon all	16	0.75	0.76	0.93	0.62	12%	-
All	18	0.76	0.77	1.00	0.62	13%	-

outliers: 3 of 16

fraction of outliers: 19%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv


Reporting number 70 : Eye - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)		
beta	Sr-90	9 21	2.20 2.20	2.03 2.16	0.92 0.98	OK OK	
photon	N-30 WIR					WIR WIR	
	N-30/60°	26 28	5.00 5.00	5.32 5.00	1.06 1.00	OK OK	
	W-80-L	6 8 30 31	5.60 5.60 5.60 5.60	4.88 5.13 5.19 5.19	0.87 0.92 0.93 0.93	OK OK OK OK	
		20 22	53.0 53.0	44.09 44.06	0.83 0.83	OK OK	
		14 15	320 320	281.38 303.67	0.88 0.95	OK OK	
		1 2	5.00 5.00	4.14 4.38	0.83 0.88	OK OK	
	Cs-137	16 17	4.80 4.80	5.09 5.24	1.06 1.09	OK OK	
	NIR	3				Legend for Quality	
	NIR	4				L low dose	
	NIR	5				M medium dose	
	NIR	7				H high dose	
	NIR	12				NIR not irradiated	
	NIR	13				WIR wrongly irradiated	
	NIR	18					
	NIR	19					
	NIR	23					
	NIR	24					
	NIR	25					
	NIR	27					
	NIR	29					
	NIR	32					

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.95	0.95	0.98	0.92	4%	-
N-30 WIR	0						
N-30/60°	2	1.03	1.03	1.06	1.00	4%	
W-80-L	4	0.92	0.91	0.93	0.87	3%	W-80: H/L 0.91
W-80-M	2	0.83	0.83	0.83	0.83	0%	
W-80-H	2	0.91	0.91	0.95	0.88	5%	
W-110	2	0.85	0.85	0.88	0.83	4%	-
Cs-137	2	1.08	1.08	1.09	1.06	2%	-
photon all	14	0.92	0.93	1.09	0.83	10%	-
All	16	0.92	0.93	1.09	0.83	9%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 71: Eye - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)		
beta	Sr-90	7	2.20	1.94	0.88	OK	
		17	2.20	2.18	0.99	OK	
	N-30	20	5.60	6.13	1.09	OK	
		21	5.60	5.97	1.07	OK	
	N-30/60°	4	5.40	5.98	1.11	OK	
		5	5.40	6.10	1.13	OK	
	W-80-L	2	8.50	13.09	1.54	OK	
		3	8.50	12.79	1.50	OK	
		11	8.50	11.64	1.37	OK	
		15	8.50	10.23	1.20	OK	
	W-80-M	26	48.0	65.57	1.37	OK	
		28	48.0	63.65	1.33	OK	
	W-80-H	22	364	463.81	1.27	OK	
		24	364	501.23	1.38	OK	
	W-110	13	5.50	6.06	1.10	OK	
		14	5.50	6.29	1.14	OK	
photon	Cs-137	30	4.80	5.62	1.17	OK	
		31	4.80	5.45	1.14	OK	

Reporting number 72 : Eye - photon dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	11	2.20	1.07	0.49	(outlier)
photon	N-30	12	2.20	1.05	0.48	(outlier)
		25	5.50	3.41	0.62	outlier
	N-30/60°	26	5.50	3.37	0.61	outlier
		20	4.80	3.10	0.65	outlier
	W-80-L	21	4.80	3.13	0.65	OK
		15	4.90	22.52	4.60	outlier
		17	4.90	2.89	0.59	outlier
		23	4.90	3.77	0.77	OK
	W-80-M	24	4.90	2.99	0.61	outlier
		1	57.0	3.60	0.06	outlier
	W-80-H	4	57.0	26.19	0.46	outlier
		18	375	140.63	0.38	outlier
	W-110	19	375	143.35	0.38	outlier
		29	5.40	2.65	0.49	outlier
	Cs-137	30	5.40	2.88	0.53	outlier
		5	5.50	3.31	0.60	outlier
		7	5.50	2.61	0.47	outlier
	NIR	2				
	NIR	3				
	NIR	6				
	NIR	8				
	NIR	9				
	NIR	10				
	NIR	13				
	NIR	14				
	NIR	16				
	NIR	22				
	NIR	27				
	NIR	28				
	NIR	31				
	NIR	32				



beta results (xx) are greyed and were not considered

Legend for Quality
 L low dose
 M medium dose
 H high dose
 NIR not irradiated
 WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.48	0.48	0.49	0.48	1%	-
N-30	2	0.62	0.62	0.62	0.61	1%	N-30: 60°/0°
N-30/60°	2	0.65	0.65	0.65	0.65	1%	1.05
W-80-L	4	0.69	1.64	4.60	0.59	120%	W-80: H/L
W-80-M	2	0.26	0.26	0.46	0.06	107%	0.16
W-80-H	2	0.38	0.38	0.38	0.38	1%	
W-110	2	0.51	0.51	0.53	0.49	6%	-
Cs-137	2	0.54	0.54	0.60	0.47	17%	-
photon all	16	0.60	0.78	4.60	0.06	132%	-
All	18	0.56	0.75	4.60	0.06	130%	-

outliers: 14 of 16

fraction of outliers: 88%



Results: IC2019_{ext eye}

1 values out of diagramme range (>2)!

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 73 : Eye - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	5	2.20	0.55	0.25	(outlier)
		20	2.20	0.52	0.24	(outlier)
photon	N-30	19	5.60	5.99	1.07	OK
		21	5.60	5.73	1.02	OK
	N-30/60°	22	5.50	6.08	1.11	OK
		23	5.50	6.03	1.10	OK
	W-80-L	10	4.70	3.83	0.81	OK
		12	4.70	4.14	0.88	OK
		24	4.70	2.59	0.55	outlier
		25	4.70	4.13	0.88	OK
	W-80-M	31	56.0	51.54	0.92	OK
		32	56.0	50.15	0.90	OK
	W-80-H	3	375	338.02	0.90	OK
		6	375	299.94	0.80	OK
	W-110	14	5.10	3.80	0.75	OK
		18	5.10	3.76	0.74	OK
	Cs-137	15	5.60	3.49	0.62	outlier
		16	5.60	3.89	0.69	OK
						Legend for Quality
						L low dose
						M medium dose
						H high dose
						NIR not irradiated
						WIR wrongly irradiated

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.24	0.24	0.25	0.24	3%	-
N-30	2	1.05	1.05	1.07	1.02	3%	N-30: 60°/0°
N-30/60°	2	1.10	1.10	1.11	1.10	1%	
W-80-L	4	0.85	0.78	0.88	0.55	20%	W-80: H/L
W-80-M	2	0.91	0.91	0.92	0.90	2%	
W-80-H	2	0.85	0.85	0.90	0.80	8%	-
W-110	2	0.74	0.74	0.75	0.74	1%	-
Cs-137	2	0.66	0.66	0.69	0.62	8%	-
photon all	16	0.88	0.86	1.11	0.55	19%	-
All	18	0.85	0.79	1.11	0.24	32%	-

outliers: 2 of 16

fraction of outliers: 13%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 74 : Eye - photon dosimeter



	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	4	2.20	1.46	0.66	(OK)
photon	N-30	32	2.20	1.62	0.74	(OK)
		21	4.50	6.25	1.39	OK
	N-30/60°	26	4.50	6.19	1.38	OK
		25	5.50	7.69	1.40	OK
	W-80-L	27	5.50	8.48	1.54	OK
		8	4.50	6.22	1.38	OK
		12	4.50	5.53	1.23	OK
		18	4.50	5.42	1.20	OK
	W-80-M	20	4.50	6.56	1.46	OK
		10	57.0	86.81	1.52	OK
	W-80-H	11	57.0	86.01	1.51	OK
		3	320	424.83	1.33	OK
	W-110	5	320	435.63	1.36	OK
		1	5.10	5.71	1.12	OK
	Cs-137	2	5.10	5.29	1.04	OK
		15	4.40	5.52	1.25	OK
		16	4.40	5.74	1.30	OK
						Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.70	0.70	0.74	0.66	7%	-
N-30	2	1.38	1.38	1.39	1.38	1%	N-30: 60°/0° 1.06
N-30/60°	2	1.47	1.47	1.54	1.40	7%	
W-80-L	4	1.31	1.32	1.46	1.20	9%	W-80: H/L 1.15
W-80-M	2	1.52	1.52	1.52	1.51	1%	
W-80-H	2	1.34	1.34	1.36	1.33	2%	
W-110	2	1.08	1.08	1.12	1.04	5%	-
Cs-137	2	1.28	1.28	1.30	1.25	3%	-
photon all	16	1.37	1.34	1.54	1.04	11%	-
All	18	1.34	1.27	1.54	0.66	19%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 75 : Eye - photon dosimeter



	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)		
beta	Sr-90	12	2.20	3.42	1.55	(OK)	
photon	N-30	21	2.20	3.25	1.48	(OK)	
		24	4.50	3.78	0.84	OK	
	N-30/60°	25	4.50	3.65	0.81	OK	
		3	4.40	3.70	0.84	OK	
	W-80-L	5	4.40	3.97	0.90	OK	
		1	5.30	4.57	0.86	OK	
		2	5.30	4.39	0.83	OK	
		20	5.30	4.57	0.86	OK	
	W-80-M	22	5.30	4.54	0.86	OK	
		17	43.0	36.59	0.85	OK	
	W-80-H	19	43.0	34.90	0.81	OK	
		28	321	182.25	0.57	outlier	
	W-110	29	321	187.82	0.59	outlier	
		10	4.80	3.21	0.67	OK	
	Cs-137	11	4.80	3.52	0.73	OK	
		31	4.80	3.59	0.75	OK	
		32	4.80	4.07	0.85	OK	
	NIR	4				Legend for Quality	
	NIR	6				L low dose	
	NIR	7				M medium dose	
	NIR	8				H high dose	
	NIR	9				NIR not irradiated	
	NIR	13				WIR wrongly irradiated	
	NIR	14					
	NIR	15					
	NIR	16					
	NIR	18					
	NIR	23					
	NIR	26					
	NIR	27					
	NIR	30					

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)	
Sr-90	2	1.52	1.52	1.55	1.48	4%	-	
N-30	2	0.83	0.83	0.84	0.81	2%	N-30: 60°/0°	
N-30/60°	2	0.87	0.87	0.90	0.84	5%	1.06	
W-80-L	4	0.86	0.85	0.86	0.83	2%	W-80: H/L	
W-80-M	2	0.83	0.83	0.85	0.81	3%	0.98	
W-80-H	2	0.58	0.58	0.59	0.57	2%		
W-110	2	0.70	0.70	0.73	0.67	7%	-	
Cs-137	2	0.80	0.80	0.85	0.75	9%	-	
photon all	16	0.83	0.79	0.90	0.57	13%	-	
All	18	0.84	0.87	1.55	0.57	29%	-	

outliers: 2 of 16

fraction of outliers: 13%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 76 : Eye - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	4 21	2.20 2.20	6.28 5.85	2.85 2.66	(outlier) (outlier)
photon	N-30	19 20	4.90 4.90	6.40 6.34	1.31 1.29	OK OK
	N-30/60°	1 2	4.30 4.30	5.60 5.52	1.30 1.28	OK OK
	W-80-L	22 25 26 28	5.00 5.00 5.00 5.00	6.09 6.16 6.05 6.01	1.22 1.23 1.21 1.20	OK OK OK OK
		5 6	47.0 47.0	58.46 60.05	1.24 1.28	OK OK
	W-80-H	16 27	378 378	403.82 400.47	1.07 1.06	OK OK
	W-110	30 32	5.20 5.20	5.46 5.46	1.05 1.05	OK OK
	Cs-137	9 10	4.90 4.90	5.69 5.69	1.16 1.16	OK OK
	NIR	3				
	NIR	7				
	NIR	8				
	NIR	11				
	NIR	12				
	NIR	13				
	NIR	14				
	NIR	15				
	NIR	17				
	NIR	18				
	NIR	23				
	NIR	24				
	NIR	29				
	NIR	31				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	2.76	2.76	2.85	2.66	5%	-
N-30	2	1.30	1.30	1.31	1.29	1%	N-30: 60°/0°
N-30/60°	2	1.29	1.29	1.30	1.28	1%	0.99
W-80-L	4	1.21	1.22	1.23	1.20	1%	W-80: H/L
W-80-M	2	1.26	1.26	1.28	1.24	2%	1.04
W-80-H	2	1.06	1.06	1.07	1.06	1%	
W-110	2	1.05	1.05	1.05	1.05	0%	-
Cs-137	2	1.16	1.16	1.16	1.16	0%	-
photon all	16	1.21	1.19	1.31	1.05	8%	-
All	18	1.22	1.37	2.85	1.05	38%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

2 values out of diagramme range (>2)!

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 77 : Eye - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	2 17	2.20 2.20	1.26 1.57	0.57 0.71	(outlier) (OK)
photon	N-30	11 12	5.50 5.50	4.71 5.19	0.86 0.94	OK OK
	N-30/60°	1 31	5.40 5.40	4.49 4.07	0.83 0.75	OK OK
	W-80-L	7 10 14 15	4.30 4.30 4.30 4.30	4.48 4.60 4.37 4.38	1.04 1.07 1.02 1.02	OK OK OK OK
		25 26	50.0 50.0	51.42 52.25	1.03 1.05	OK OK
		18 19	353 353	369.10 380.20	1.05 1.08	OK OK
		5 6	5.30 5.30	4.62 4.91	0.87 0.93	OK OK
	Cs-137	8 9	4.40 4.40	3.91 3.78	0.89 0.86	OK OK
	NIR	3				
	NIR	4				
	NIR	13				
	NIR	16				
	NIR	20				
	NIR	21				
	NIR	22				
	NIR	23				
	NIR	24				
	NIR	27				
	NIR	28				
	NIR	29				
	NIR	30				
	NIR	32				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.64	0.64	0.71	0.57	15%	-
N-30	2	0.90	0.90	0.94	0.86	7%	N-30: 60°/0°
N-30/60°	2	0.79	0.79	0.83	0.75	7%	0.88
W-80-L	4	1.03	1.04	1.07	1.02	2%	W-80: H/L
W-80-M	2	1.04	1.04	1.05	1.03	1%	1.00
W-80-H	2	1.06	1.06	1.08	1.05	2%	
W-110	2	0.90	0.90	0.93	0.87	4%	-
Cs-137	2	0.87	0.87	0.89	0.86	2%	-
photon all	16	0.98	0.95	1.08	0.75	11%	-
All	18	0.94	0.92	1.08	0.57	15%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv


Reporting number 78 : Eye - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	19 24	2.20 2.20	7.00 7.21	3.18 3.28	outlier outlier
photon	N-30	5 6	4.80 4.80	7.09 6.32	1.48 1.32	OK OK
	N-30/60°	14 15	4.60 4.60	7.67 7.27	1.67 1.58	OK OK
	W-80-L	12 13 17 18	4.60 4.60 4.60 4.60	5.48 5.61 5.64 5.48	1.19 1.22 1.23 1.19	OK OK OK OK
		2 3	51.0 51.0	58.75 63.34	1.15 1.24	OK OK
	W-80-H	9 10	364 364	435.24 420.04	1.20 1.15	OK OK
	W-110	30 31	4.70 4.70	4.61 4.71	0.98 1.00	OK OK
		26 27	5.30 5.30	5.82 6.13	1.10 1.16	OK OK
	Cs-137	1 4 7 8 11 16 20 21 22 23 25 28 29 32	1 4 7 8 11 16 20 21 22 23 25 28 29 32			Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	3.23	3.23	3.28	3.18	2%	-
N-30	2	1.40	1.40	1.48	1.32	8%	N-30: 60°/0°
N-30/60°	2	1.62	1.62	1.67	1.58	4%	1.16
W-80-L	4	1.21	1.21	1.23	1.19	2%	W-80: H/L
W-80-M	2	1.20	1.20	1.24	1.15	5%	0.99
W-80-H	2	1.17	1.17	1.20	1.15	3%	
W-110	2	0.99	0.99	1.00	0.98	2%	-
Cs-137	2	1.13	1.13	1.16	1.10	4%	-
photon all	16	1.19	1.24	1.67	0.98	15%	-
All	18	1.21	1.46	3.28	0.98	46%	-

outliers: 2 of 18

fraction of outliers: 11%



Results: IC2019_{ext eye}

2 values out of diagramme range (>2)!

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 79 : Eye - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	17 31	2.20 2.20	3.28 2.83	1.49 1.29	(OK) (OK)
photon	N-30	24	5.20	5.12	0.99	OK
		25	5.20	5.25	1.01	OK
	N-30/60°	1	4.60	4.50	0.98	OK
		5	4.60	4.32	0.94	OK
	W-80-L	14	5.30	5.24	0.99	OK
		15	5.30	5.32	1.00	OK
		20	5.30	5.67	1.07	OK
		21	5.30	5.56	1.05	OK
	W-80-M	18	50.0	49.74	0.99	OK
		19	50.0	50.32	1.01	OK
	W-80-H	6	342	330.20	0.97	OK
		8	342	319.50	0.93	OK
	W-110	26	4.90	5.10	1.04	OK
		27	4.90	4.97	1.01	OK
	Cs-137	9 10	4.30 4.30	4.21 4.27	0.98 0.99	OK OK
	NIR	2				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	3				
	NIR	4				
	NIR	7				
	NIR	11				
	NIR	12				
	NIR	13				
	NIR	16				
	NIR	22				
	NIR	23				
	NIR	28				
	NIR	29				
	NIR	30				
	NIR	32				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	1.39	1.39	1.49	1.29	10%	-
N-30	2	1.00	1.00	1.01	0.99	2%	N-30: 60°/0° 0.96
N-30/60°	2	0.96	0.96	0.98	0.94	3%	
W-80-L	4	1.03	1.03	1.07	0.99	4%	W-80: H/L 0.97
W-80-M	2	1.00	1.00	1.01	0.99	1%	
W-80-H	2	0.95	0.95	0.97	0.93	2%	
W-110	2	1.03	1.03	1.04	1.01	2%	-
Cs-137	2	0.99	0.99	0.99	0.98	1%	-
photon all	16	0.99	1.00	1.07	0.93	4%	-
All	18	1.00	1.04	1.49	0.93	13%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 80 : Eye - photon dosimeter



	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)		
beta	Sr-90	18	2.20	5.30	2.41	(outlier)	
photon	N-30	27	2.20	5.32	2.42	(outlier)	
		31	5.20	6.41	1.23	OK	
	N-30/60°	32	5.20	6.27	1.20	OK	
		11	4.70	6.58	1.40	OK	
	W-80-L	13	4.70	6.53	1.39	OK	
		22	4.80	5.04	1.05	OK	
		23	4.80	5.16	1.08	OK	
		24	4.80	5.16	1.08	OK	
	W-80-M	25	4.80	5.18	1.08	OK	
		29	53.0	56.66	1.07	OK	
	W-80-H	30	53.0	56.02	1.06	OK	
		20	320	338.80	1.06	OK	
	W-110	21	320	352.20	1.10	OK	
		1	5.80	5.31	0.92	OK	
	Cs-137	2	5.80	5.16	0.89	OK	
		9	4.40	4.47	1.02	OK	
		10	4.40	4.40	1.00	OK	
	NIR	3				Legend for Quality	
	NIR	4				L low dose	
	NIR	5				M medium dose	
	NIR	6				H high dose	
	NIR	7				NIR not irradiated	
	NIR	8				WIR wrongly irradiated	
	NIR	12					
	NIR	14					
	NIR	15					
	NIR	16					
	NIR	17					
	NIR	19					
	NIR	26					
	NIR	28					

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	2.41	2.41	2.42	2.41	0%	-
N-30	2	1.22	1.22	1.23	1.20	2%	N-30: 60°/0°
N-30/60°	2	1.40	1.40	1.40	1.39	1%	
W-80-L	4	1.08	1.07	1.08	1.05	1%	W-80: H/L
W-80-M	2	1.06	1.06	1.07	1.06	1%	
W-80-H	2	1.08	1.08	1.10	1.06	3%	
W-110	2	0.90	0.90	0.92	0.89	2%	-
Cs-137	2	1.01	1.01	1.02	1.00	1%	-
photon all	16	1.07	1.10	1.40	0.89	13%	-
All	18	1.08	1.25	2.42	0.89	36%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

2 values out of diagramme range (>2)!

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 81: Eye - photon dosimeter



	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)		
beta	Sr-90	18	2.20	1.19	0.54	(outlier)	
photon	N-30	21	2.20	1.24	0.56	(outlier)	
		27	5.20	5.74	1.10	OK	
	N-30/60°	28	5.20	5.75	1.11	OK	
		16	5.40	6.36	1.18	OK	
	W-80-L	17	5.40	6.14	1.14	OK	
		6	5.00	5.30	1.06	OK	
		9	5.00	5.59	1.12	OK	
		25	5.00	5.65	1.13	OK	
	W-80-M	26	5.00	5.30	1.06	OK	
		3	47.0	42.53	0.90	OK	
	W-80-H	4	47.0	38.55	0.82	OK	
		31	333	378.80	1.14	OK	
	W-110	32	333	360.22	1.08	OK	
		7	4.90	4.45	0.91	OK	
	Cs-137	8	4.90	4.35	0.89	OK	
		10	5.20	5.09	0.98	OK	
		11	5.20	5.30	1.02	OK	
					Legend for Quality		
					L low dose		
					M medium dose		
					H high dose		
					NIR not irradiated		
					WIR wrongly irradiated		

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.55	0.55	0.56	0.54	3%	-
N-30	2	1.10	1.10	1.11	1.10	0%	N-30: 60°/0°
N-30/60°	2	1.16	1.16	1.18	1.14	2%	
W-80-L	4	1.09	1.09	1.13	1.06	3%	W-80: H/L
W-80-M	2	0.86	0.86	0.90	0.82	7%	
W-80-H	2	1.11	1.11	1.14	1.08	4%	
W-110	2	0.90	0.90	0.91	0.89	2%	-
Cs-137	2	1.00	1.00	1.02	0.98	3%	-
photon all	16	1.07	1.04	1.18	0.82	10%	-
All	18	1.06	0.99	1.18	0.54	19%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv


Reporting number 82 : Eye - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	1 20	2.20 2.20	5.34 5.27	2.43 2.40	outlier outlier
photon	N-30	12 13	4.60 4.60	4.53 5.62	0.98 1.22	OK OK
	N-30/60°	14 15	5.10 5.10	5.28 5.09	1.04 1.00	OK OK
	W-80-L	3 4 9 10	5.30 5.30 5.30 5.30	5.71 5.92 5.31 5.62	1.08 1.12 1.00 1.06	OK OK OK OK
		24 25	46.0 46.0	47.45 42.46	1.03 0.92	OK OK
		22 23	358 358	260.17 270.58	0.73 0.76	OK OK
		7 8	5.50 5.50	4.40 5.18	0.80 0.94	OK OK
	Cs-137	31 32	5.50 5.50	4.77 6.16	0.87 1.12	OK OK
	NIR	2				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	5				
	NIR	6				
	NIR	11				
	NIR	16				
	NIR	17				
	NIR	18				
	NIR	19				
	NIR	21				
	NIR	26				
	NIR	27				
	NIR	28				
	NIR	29				
	NIR	30				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	2.41	2.41	2.43	2.40	1%	-
N-30	2	1.10	1.10	1.22	0.98	15%	N-30: 60°/0°
N-30/60°	2	1.02	1.02	1.04	1.00	3%	
W-80-L	4	1.07	1.06	1.12	1.00	4%	W-80: H/L
W-80-M	2	0.98	0.98	1.03	0.92	8%	
W-80-H	2	0.74	0.74	0.76	0.73	3%	
W-110	2	0.87	0.87	0.94	0.80	12%	-
Cs-137	2	0.99	0.99	1.12	0.87	18%	-
photon all	16	1.00	0.98	1.22	0.73	14%	-
All	18	1.02	1.14	2.43	0.73	42%	-

outliers: 2 of 18

fraction of outliers: 11%



Results: IC2019_{ext eye}

2 values out of diagramme range (>2)!

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv


Reporting number 83 : Eye - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	10	2.20	1.22	0.55	outlier
		14	2.20	1.28	0.58	OK
	N-30	13	5.30	7.02	1.32	OK
		15	5.30	7.22	1.36	OK
	N-30/60°	16	4.30	5.95	1.38	OK
		17	4.30	5.66	1.32	OK
	W-80-L	1	4.40	5.22	1.19	OK
		2	4.40	5.21	1.18	OK
		29	4.40	5.24	1.19	OK
		30	4.40	5.18	1.18	OK
	W-80-M	11	56.0	68.96	1.23	OK
		12	56.0	70.57	1.26	OK
	W-80-H	27	364	545.76	1.50	OK
		28	364	570.39	1.57	OK
	W-110	25	4.80	5.00	1.04	OK
		26	4.80	4.78	0.99	OK
	Cs-137	3	5.60	6.50	1.16	OK
		4	5.60	6.58	1.17	OK
photon	NIR	5				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	6				
	NIR	7				
	NIR	8				
	NIR	9				
	NIR	18				
	NIR	19				
	NIR	20				
	NIR	21				
	NIR	22				
	NIR	23				
	NIR	24				
	NIR	31				
	NIR	32				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.57	0.57	0.58	0.55	4%	-
N-30	2	1.34	1.34	1.36	1.32	2%	N-30: 60°/0°
N-30/60°	2	1.35	1.35	1.38	1.32	4%	
W-80-L	4	1.19	1.18	1.19	1.18	0%	W-80: H/L
W-80-M	2	1.25	1.25	1.26	1.23	2%	
W-80-H	2	1.53	1.53	1.57	1.50	3%	
W-110	2	1.02	1.02	1.04	0.99	3%	-
Cs-137	2	1.17	1.17	1.17	1.16	1%	-
photon all	16	1.21	1.25	1.57	0.99	12%	-
All	18	1.19	1.18	1.57	0.55	22%	-

outliers: 1 of 18

fraction of outliers: 6%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 84 : Eye - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	20	2.20	4.82	2.19	(outlier)
photon	N-30	24	2.20	5.56	2.53	(outlier)
		10	4.50	4.11	0.91	OK
	N-30/60°	12	4.50	3.97	0.88	OK
		29	4.70	3.75	0.80	OK
	W-80-L	30	4.70	4.62	0.98	OK
		17	5.50	5.43	0.99	OK
		18	5.50	5.10	0.93	OK
		21	5.50	5.20	0.95	OK
	W-80-M	23	5.50	5.13	0.93	OK
		8	46.0	41.29	0.90	OK
	W-80-H	9	46.0	41.42	0.90	OK
		14	350	318.64	0.91	OK
	W-110	15	350	306.86	0.88	OK
		13	5.00	3.93	0.79	OK
	Cs-137	16	5.00	3.92	0.78	OK
		26	4.80	3.56	0.74	OK
		27	4.80	3.31	0.69	OK
					Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated	

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	2.36	2.36	2.53	2.19	10%	-
N-30	2	0.90	0.90	0.91	0.88	2%	N-30: 60°/0°
N-30/60°	2	0.89	0.89	0.98	0.80	15%	
W-80-L	4	0.94	0.95	0.99	0.93	3%	W-80: H/L
W-80-M	2	0.90	0.90	0.90	0.90	0%	
W-80-H	2	0.89	0.89	0.91	0.88	3%	
W-110	2	0.79	0.79	0.79	0.78	0%	-
Cs-137	2	0.72	0.72	0.74	0.69	5%	-
photon all	16	0.90	0.87	0.99	0.69	10%	-
All	18	0.91	1.04	2.53	0.69	47%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

2 values out of diagramme range (>2)!

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 85 : Eye - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)		
beta	Sr-90	9	2.20	6.30	2.86	(outlier)	
photon	N-30	27	2.20	6.52	2.96	(outlier)	
		28	4.50	5.32	1.18	OK	
	N-30/60°	29	4.50	5.31	1.18	OK	
		6	4.50	5.43	1.21	OK	
	W-80-L	8	4.50	6.35	1.41	OK	
		11	4.60	5.12	1.11	OK	
		12	4.60	5.29	1.15	OK	
		22	4.60	5.59	1.22	OK	
	W-80-M	23	4.60	4.97	1.08	OK	
		19	46.0	50.70	1.10	OK	
	W-80-H	20	46.0	51.90	1.13	OK	
		25	353	392.70	1.11	OK	
	W-110	26	353	406.70	1.15	OK	
		7	4.70	4.52	0.96	OK	
	Cs-137	31	4.70	4.65	0.99	OK	
		17	5.30	5.15	0.97	OK	
		18	5.30	4.99	0.94	OK	
						Legend for Quality	
						L low dose	
						M medium dose	
						H high dose	
						NIR not irradiated	
						WIR wrongly irradiated	

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	2.91	2.91	2.96	2.86	2%	-
N-30	2	1.18	1.18	1.18	1.18	0%	N-30: 60°/0°
N-30/60°	2	1.31	1.31	1.41	1.21	11%	1.11
W-80-L	4	1.13	1.14	1.22	1.08	5%	W-80: H/L
W-80-M	2	1.12	1.12	1.13	1.10	2%	0.98
W-80-H	2	1.13	1.13	1.15	1.11	2%	
W-110	2	0.98	0.98	0.99	0.96	2%	-
Cs-137	2	0.96	0.96	0.97	0.94	2%	-
photon all	16	1.12	1.12	1.41	0.94	11%	-
All	18	1.14	1.32	2.96	0.94	45%	-

outliers: 0 of 16

fraction of outliers: 0%



Reporting number 86 : Eye - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	10	2.20	5.70	2.59	(outlier)
photon	N-30	13	2.20	5.40	2.45	(outlier)
		1	4.50	4.70	1.04	OK
	N-30/60°	2	4.50	5.00	1.11	OK
		11	4.80	4.35	0.91	OK
	W-80-L	19	4.80	4.75	0.99	OK
		21	5.70	6.05	1.06	OK
		24	5.70	6.65	1.17	OK
		31	5.70	5.90	1.04	OK
	W-80-M	32	5.70	6.95	1.22	OK
		25	48.0	56.00	1.17	OK
	W-80-H	26	48.0	56.00	1.17	OK
		6	331	520.00	1.57	OK
	W-110	8	331	520.00	1.57	OK
		27	4.50	3.90	0.87	OK
	Cs-137	28	4.50	4.75	1.06	OK
		4	5.40	6.15	1.14	OK
		5	5.40	5.10	0.94	OK
	NIR	3				
	NIR	7				
	NIR	9				
	NIR	12				
	NIR	14				
	NIR	15				
	NIR	16				
	NIR	17				
	NIR	18				
	NIR	20				
	NIR	22				
	NIR	23				
	NIR	29				
	NIR	30				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	2.52	2.52	2.59	2.45	4%	-
N-30	2	1.08	1.08	1.11	1.04	4%	
N-30/60°	2	0.95	0.95	0.99	0.91	6%	
W-80-L	4	1.11	1.12	1.22	1.04	8%	
W-80-M	2	1.17	1.17	1.17	1.17	0%	
W-80-H	2	1.57	1.57	1.57	1.57	0%	
W-110	2	0.96	0.96	1.06	0.87	14%	-
Cs-137	2	1.04	1.04	1.14	0.94	13%	-
photon all	16	1.09	1.13	1.57	0.87	18%	-
All	18	1.13	1.28	2.59	0.87	38%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

2 values out of diagramme range (>2)!

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 87 : Eye - photon dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	5 20	2.20 2.20	6.41 6.73	2.91 3.06	(outlier) (outlier)
photon	N-30	7 8	4.40 4.40	5.67 5.90	1.29 1.34	OK OK
		21 23	4.30 4.30	5.69 5.46	1.32 1.27	OK OK
	W-80-L	2 3 18 19	4.40 4.40 4.40 4.40	5.41 5.53 5.51 5.33	1.23 1.26 1.25 1.21	OK OK OK OK
		16 17	47.0 47.0	59.22 58.82	1.26 1.25	OK OK
	W-80-H	11 13	319 319	363.35 353.81	1.14 1.11	OK OK
		4 6	5.20 5.20	5.38 5.60	1.03 1.08	OK OK
	Cs-137	14 15	4.50 4.50	5.31 5.18	1.18 1.15	OK OK
	NIR	1				
	NIR	9				
	NIR	10				
	NIR	12				
	NIR	22				
	NIR	24				
	NIR	25				
	NIR	26				
	NIR	27				
	NIR	28				
	NIR	29				
	NIR	30				
	NIR	31				
	NIR	32				


beta results (xx) are greyed and were not considered

Legend for Quality
 L low dose
 M medium dose
 H high dose
 NIR not irradiated
 WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	2.99	2.99	3.06	2.91	3%	-
N-30	2	1.31	1.31	1.34	1.29	3%	N-30: 60°/0°
N-30/60°	2	1.30	1.30	1.32	1.27	3%	0.99
W-80-L	4	1.24	1.24	1.26	1.21	2%	W-80: H/L
W-80-M	2	1.26	1.26	1.26	1.25	0%	1.01
W-80-H	2	1.12	1.12	1.14	1.11	2%	
W-110	2	1.06	1.06	1.08	1.03	3%	-
Cs-137	2	1.17	1.17	1.18	1.15	2%	-
photon all	16	1.24	1.21	1.34	1.03	7%	-
All	18	1.25	1.41	3.06	1.03	41%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

2 values out of diagramme range (>2)!

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv


Reporting number 88 : Eye - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
	radiation type	radiation quality	dosemeter number		dose $H_p(3)$ mSv	response R (reported/true)
beta	Sr-90	15	2.20	1.60	0.73	OK
photon		32	2.20	1.50	0.68	OK
		23	5.10	4.70	0.92	OK
		25	5.10	4.80	0.94	OK
		2	5.30	5.00	0.94	OK
		6	5.30	4.80	0.91	OK
		4	5.20	5.70	1.10	OK
		5	5.20	5.90	1.13	OK
		17	5.20	5.80	1.12	OK
		19	5.20	5.90	1.13	OK
		13	54.0	60.40	1.12	OK
		14	54.0	63.60	1.18	OK
		11	358	386.30	1.08	OK
		12	358	396.50	1.11	OK
		18	5.30	5.30	1.00	OK
		20	5.30	5.40	1.02	OK
	Cs-137	3	4.80	4.60	0.96	OK
		8	4.80	4.60	0.96	OK
	NIR	1				
	NIR	7				
	NIR	9				
	NIR	10				
	NIR	16				
	NIR	21				
	NIR	22				
	NIR	24				
	NIR	26				
	NIR	27				
	NIR	28				
	NIR	29				
	NIR	30				
	NIR	31				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.70	0.70	0.73	0.68	5%	-
N-30	2	0.93	0.93	0.94	0.92	1%	N-30: 60°/0°
N-30/60°	2	0.92	0.92	0.94	0.91	3%	
W-80-L	4	1.13	1.12	1.13	1.10	2%	W-80: H/L
W-80-M	2	1.15	1.15	1.18	1.12	4%	
W-80-H	2	1.09	1.09	1.11	1.08	2%	
W-110	2	1.01	1.01	1.02	1.00	1%	-
Cs-137	2	0.96	0.96	0.96	0.96	0%	-
photon all	16	1.05	1.04	1.18	0.91	9%	-
All	18	1.01	1.00	1.18	0.68	14%	-

outliers: 0 of 18

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 89 : Eye - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	18	2.20	1.49	0.68	(OK)
photon	N-30	23	2.20	1.68	0.76	(OK)
		22	5.60	7.00	1.25	OK
	N-30/60°	24	5.60	7.84	1.40	OK
		14	4.80	6.99	1.46	OK
	W-80-L	16	4.80	6.76	1.41	OK
		11	4.30	5.71	1.33	OK
		12	4.30	5.39	1.25	OK
		26	4.30	4.86	1.13	OK
	W-80-M	27	4.30	5.49	1.28	OK
		15	47.0	57.63	1.23	OK
	W-80-H	17	47.0	64.49	1.37	OK
		3	332	462.37	1.39	OK
	W-110	5	332	458.48	1.38	OK
		25	5.40	6.07	1.12	OK
	Cs-137	28	5.40	5.94	1.10	OK
		6	5.60	6.92	1.24	OK
		7	5.60	7.23	1.29	OK
						Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.72	0.72	0.76	0.68	8%	-
N-30	2	1.33	1.33	1.40	1.25	8%	N-30: 60°/0° 1.08
N-30/60°	2	1.43	1.43	1.46	1.41	2%	
W-80-L	4	1.27	1.25	1.33	1.13	7%	W-80: H/L 1.04
W-80-M	2	1.30	1.30	1.37	1.23	8%	
W-80-H	2	1.39	1.39	1.39	1.38	1%	
W-110	2	1.11	1.11	1.12	1.10	2%	-
Cs-137	2	1.26	1.26	1.29	1.24	3%	-
photon all	16	1.28	1.29	1.46	1.10	8%	-
All	18	1.27	1.23	1.46	0.68	17%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 90 : Eye - photon dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	12 31	2.20 2.20	1.25 1.25	0.57 0.57	(outlier) (outlier)
photon	N-30	23 27	4.30 4.30	4.49 4.80	1.04 1.12	OK OK
	N-30/60°	20 21	5.40 5.40	5.84 5.87	1.08 1.09	OK OK
	W-80-L	9	4.60	5.11	1.11	OK
		10	4.60	5.13	1.12	OK
		18	4.60	4.92	1.07	OK
		19	4.60	4.93	1.07	OK
	W-80-M	2 5	53.0 53.0	56.05 58.17	1.06 1.10	OK OK
	W-80-H	14 16	332 332	319.00 331.20	0.96 1.00	OK OK
	W-110	29 30	4.90 4.90	4.90 4.97	1.00 1.01	OK OK
	Cs-137	7 8	4.70 4.70	4.72 4.73	1.00 1.01	OK OK
	NIR	1				
	NIR	3				
	NIR	4				
	NIR	6				
	NIR	11				
	NIR	13				
	NIR	15				
	NIR	17				
	NIR	22				
	NIR	24				
	NIR	25				
	NIR	26				
	NIR	28				
	NIR	32				


beta results (xx) are greyed and were not considered

Legend for Quality
 L low dose
 M medium dose
 H high dose
 NIR not irradiated
 WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.57	0.57	0.57	0.57	0%	-
N-30	2	1.08	1.08	1.12	1.04	5%	N-30: 60°/0°
N-30/60°	2	1.08	1.08	1.09	1.08	0%	
W-80-L	4	1.09	1.09	1.12	1.07	2%	W-80: H/L
W-80-M	2	1.08	1.08	1.10	1.06	3%	
W-80-H	2	0.98	0.98	1.00	0.96	3%	
W-110	2	1.01	1.01	1.01	1.00	1%	-
Cs-137	2	1.01	1.01	1.01	1.00	0%	-
photon all	16	1.06	1.05	1.12	0.96	5%	-
All	18	1.05	1.00	1.12	0.57	16%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv



Reporting number 91: Eye - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	19	2.20	1.15	0.52	outlier
		27	2.20	1.69	0.77	OK
	N-30	29	4.60	6.53	1.42	OK
		31	4.60	9.49	2.06	outlier
	N-30/60°	10	4.90	11.61	2.37	outlier
		11	4.90	7.42	1.51	OK
	W-80-L	6	5.30	8.89	1.68	OK
		7	5.30	10.32	1.95	outlier
		14	5.30	8.17	1.54	OK
		15	5.30	9.70	1.83	outlier
	W-80-M	18	53.0	80.28	1.51	OK
		22	53.0	82.06	1.55	OK
	W-80-H	28	328	497.33	1.52	OK
		30	328	609.14	1.86	outlier
photon	W-110	16	5.60	7.09	1.27	OK
		17	5.60	8.81	1.57	OK
	Cs-137	4	5.50	5.85	1.06	OK
		5	5.50	6.31	1.15	OK
						Legend for Quality
						L low dose
						M medium dose
						H high dose
						NIR not irradiated
						WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.65	0.65	0.77	0.52	27%	-
N-30	2	1.74	1.74	2.06	1.42	26%	N-30: 60°/0°
N-30/60°	2	1.94	1.94	2.37	1.51	31%	
W-80-L	4	1.75	1.75	1.95	1.54	10%	W-80: H/L
W-80-M	2	1.53	1.53	1.55	1.51	2%	
W-80-H	2	1.69	1.69	1.86	1.52	14%	
W-110	2	1.42	1.42	1.57	1.27	15%	-
Cs-137	2	1.11	1.11	1.15	1.06	5%	-
photon all	16	1.54	1.62	2.37	1.06	21%	-
All	18	1.53	1.51	2.37	0.52	30%	-

outliers: 6 of 18

fraction of outliers: 33%



Results: IC2019_{ext eye}

2 values out of diagramme range (>2)!

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 92 : Eye - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	2 27	2.20 2.20	0.97 0.88	0.44 0.40	(outlier) (outlier)
photon	N-30	3 5	4.50 4.50	5.97 6.09	1.33 1.35	OK OK
	N-30/60°	6 7	4.30 4.30	6.70 6.81	1.56 1.58	OK OK
	W-80-L	9 12 20 22	4.30 4.30 4.30 4.30	4.69 4.44 4.55 4.79	1.09 1.03 1.06 1.11	OK OK OK OK
		24 25	53.0 53.0	58.74 57.84	1.11 1.09	OK OK
		30 32	364 364	391.29 390.67	1.07 1.07	OK OK
		28 29	4.80 4.80	4.46 4.77	0.93 0.99	OK OK
	Cs-137	21 23	4.90 4.90	4.10 4.12	0.84 0.84	OK OK
	NIR	1				
	NIR	4				
	NIR	8				
	NIR	10				
	NIR	11				
	NIR	13				
	NIR	14				
	NIR	15				
	NIR	16				
	NIR	17				
	NIR	18				
	NIR	19				
	NIR	26				
	NIR	31				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.42	0.42	0.44	0.40	7%	-
N-30	2	1.34	1.34	1.35	1.33	1%	N-30: 60°/0°
N-30/60°	2	1.57	1.57	1.58	1.56	1%	1.17
W-80-L	4	1.07	1.07	1.11	1.03	3%	W-80: H/L
W-80-M	2	1.10	1.10	1.11	1.09	1%	1.02
W-80-H	2	1.07	1.07	1.07	1.07	0%	
W-110	2	0.96	0.96	0.99	0.93	5%	-
Cs-137	2	0.84	0.84	0.84	0.84	0%	-
photon all	16	1.08	1.13	1.58	0.84	20%	-
All	18	1.07	1.05	1.58	0.40	29%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv


Reporting number 93 : Eye - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	20	2.20	2.66	1.21	OK
photon	N-30	30	2.20	2.67	1.21	OK
		23	4.60	6.61	1.44	OK
	N-30/60°	24	4.60	6.37	1.38	OK
		1	4.90	7.59	1.55	OK
	W-80-L	3	4.90	6.90	1.41	OK
		2	4.60	5.37	1.17	OK
		4	4.60	5.46	1.19	OK
		26	4.60	5.57	1.21	OK
	W-80-M	27	4.60	5.37	1.17	OK
		5	48.0	56.29	1.17	OK
	W-80-H	8	48.0	48.30	1.01	OK
		15	328	395.94	1.21	OK
	W-110	17	328	384.42	1.17	OK
		6	5.60	6.99	1.25	OK
	Cs-137	7	5.60	5.43	0.97	OK
		18	4.40	3.74	0.85	OK
		19	4.40	4.13	0.94	OK
						Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	1.21	1.21	1.21	1.21	0%	-
N-30	2	1.41	1.41	1.44	1.38	3%	N-30: 60°/0°
N-30/60°	2	1.48	1.48	1.55	1.41	7%	
W-80-L	4	1.18	1.18	1.21	1.17	2%	W-80: H/L
W-80-M	2	1.09	1.09	1.17	1.01	11%	
W-80-H	2	1.19	1.19	1.21	1.17	2%	
W-110	2	1.11	1.11	1.25	0.97	18%	-
Cs-137	2	0.89	0.89	0.94	0.85	7%	-
photon all	16	1.18	1.19	1.55	0.85	16%	-
All	18	1.20	1.19	1.55	0.85	15%	-

outliers: 0 of 18

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv



Reporting number 94 : Eye - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	18 30	2.20 2.20	4.37 4.34	1.99 1.97	outlier outlier
photon	N-30 WIR					WIR WIR
	N-30/60°	31 32	5.00 5.00	7.87 7.38	1.57 1.48	OK OK
	W-80-L	10 14 25 26	4.70 4.70 4.70 4.70	5.30 5.23 5.18 5.18	1.13 1.11 1.10 1.10	OK OK OK OK
		6 7	51.0 51.0	57.00 56.50	1.12 1.11	OK OK
		1 2	320 320	358.90 353.90	1.12 1.11	OK OK
		23 24	5.50 5.50	5.55 5.45	1.01 0.99	OK OK
	Cs-137	12 13	4.40 4.40	3.68 3.65	0.84 0.83	OK OK
	NIR	3				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	4				
	NIR	5				
	NIR	8				
	NIR	9				
	NIR	11				
	NIR	15				
	NIR	16				
	NIR	17				
	NIR	19				
	NIR	20				
	NIR	21				
	NIR	22				
	NIR	28				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	1.98	1.98	1.99	1.97	0%	-
N-30 WIR	0						
N-30/60°	2	1.53	1.53	1.57	1.48	5%	
W-80-L	4	1.11	1.11	1.13	1.10	1%	W-80: H/L 1.00
W-80-M	2	1.11	1.11	1.12	1.11	1%	
W-80-H	2	1.11	1.11	1.12	1.11	1%	
W-110	2	1.00	1.00	1.01	0.99	1%	-
Cs-137	2	0.83	0.83	0.84	0.83	1%	-
photon all	14	1.11	1.12	1.57	0.83	18%	-
All	16	1.11	1.22	1.99	0.83	29%	-

outliers: 2 of 16

fraction of outliers: 13%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 95 : Eye - photon dosimeter



	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)		
beta	Sr-90	30	2.20	4.63	2.10	(outlier)	
photon	N-30	32	2.20	4.77	2.17	(outlier)	
		18	4.30	5.47	1.27	OK	
	N-30/60°	19	4.30	5.47	1.27	OK	
		12	4.50	6.54	1.45	OK	
	W-80-L	14	4.50	6.53	1.45	OK	
		16	5.20	5.68	1.09	OK	
		17	5.20	5.53	1.06	OK	
		23	5.20	5.65	1.09	OK	
	W-80-M	25	5.20	5.26	1.01	OK	
		21	56.0	65.73	1.17	OK	
	W-80-H	24	56.0	59.22	1.06	OK	
		4	342	372.65	1.09	OK	
	W-110	5	342	385.92	1.13	OK	
		8	4.40	4.40	1.00	OK	
	Cs-137	10	4.40	4.61	1.05	OK	
		28	5.50	4.30	0.78	OK	
		29	5.50	4.24	0.77	OK	
	NIR	1				Legend for Quality	
	NIR	2				L low dose	
	NIR	3				M medium dose	
	NIR	6				H high dose	
	NIR	7				NIR not irradiated	
	NIR	9				WIR wrongly irradiated	
	NIR	11					
	NIR	13					
	NIR	15					
	NIR	20					
	NIR	22					
	NIR	26					
	NIR	27					
	NIR	31					

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)	
Sr-90	2	2.14	2.14	2.17	2.10	2%	-	
N-30	2	1.27	1.27	1.27	1.27	0%	N-30: 60°/0°	
N-30/60°	2	1.45	1.45	1.45	1.45	0%	1.14	
W-80-L	4	1.07	1.06	1.09	1.01	3%	W-80: H/L	
W-80-M	2	1.12	1.12	1.17	1.06	7%	1.05	
W-80-H	2	1.11	1.11	1.13	1.09	2%		
W-110	2	1.02	1.02	1.05	1.00	3%		
Cs-137	2	0.78	0.78	0.78	0.77	1%		
photon all	16	1.09	1.11	1.45	0.77	17%		
All	18	1.09	1.22	2.17	0.77	31%		

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

2 values out of diagramme range (>2)!

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv


Reporting number 96 : Eye - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	10	2.20	1.69	0.77	OK
		16	2.20	1.76	0.80	OK
	N-30	31	5.20	5.86	1.13	OK
		32	5.20	5.71	1.10	OK
	N-30/60°	4	5.30	4.34	0.82	OK
		5	5.30	4.78	0.90	OK
	W-80-L	7	5.60	5.36	0.96	OK
		9	5.60	5.42	0.97	OK
		21	5.60	5.13	0.92	OK
		22	5.60	5.30	0.95	OK
	W-80-M	12	43.0	42.35	0.98	OK
		13	43.0	42.29	0.98	OK
	W-80-H	17	359	367.32	1.02	OK
		18	359	359.30	1.00	OK
	W-110	14	5.60	5.06	0.90	OK
		15	5.60	5.10	0.91	OK
	Cs-137	1	4.40	3.37	0.77	OK
		2	4.40	3.34	0.76	OK
						Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.78	0.78	0.80	0.77	3%	-
N-30	2	1.11	1.11	1.13	1.10	2%	N-30: 60°/0° 0.77
N-30/60°	2	0.86	0.86	0.90	0.82	7%	
W-80-L	4	0.95	0.95	0.97	0.92	2%	W-80: H/L 1.04
W-80-M	2	0.98	0.98	0.98	0.98	0%	
W-80-H	2	1.01	1.01	1.02	1.00	2%	
W-110	2	0.91	0.91	0.91	0.90	1%	-
Cs-137	2	0.76	0.76	0.77	0.76	1%	-
photon all	16	0.95	0.94	1.13	0.76	11%	-
All	18	0.93	0.92	1.13	0.76	12%	-

outliers: 0 of 18

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 97 : Eye - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	13	2.20	2.66	1.21	(OK)
		25	2.20	2.13	0.97	(OK)
	N-30	16	5.70	7.30	1.28	OK
		19	5.70	7.42	1.30	OK
	N-30/60°	15	4.70	6.87	1.46	OK
		18	4.70	6.82	1.45	OK
	W-80-L	10	4.90	5.72	1.17	OK
		11	4.90	5.54	1.13	OK
		26	4.90	5.36	1.09	OK
		28	4.90	5.26	1.07	OK
	W-80-M	3	51.0	57.47	1.13	OK
		6	51.0	55.53	1.09	OK
	W-80-H	30	324	365.18	1.13	OK
		31	324	361.89	1.12	OK
	W-110	7	5.80	5.79	1.00	OK
		8	5.80	5.62	0.97	OK
	Cs-137	17	5.60	4.87	0.87	OK
		20	5.60	4.76	0.85	OK
photon	NIR	1				
		2				
		4				
		5				
		9				
		12				
		14				
		21				
		22				
		23				
		24				
		27				
		29				
		32				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	1.09	1.09	1.21	0.97	16%	-
N-30	2	1.29	1.29	1.30	1.28	1%	N-30: 60°/0°
N-30/60°	2	1.46	1.46	1.46	1.45	1%	1.13
W-80-L	4	1.11	1.12	1.17	1.07	4%	W-80: H/L
W-80-M	2	1.11	1.11	1.13	1.09	2%	0.99
W-80-H	2	1.12	1.12	1.13	1.12	1%	
W-110	2	0.98	0.98	1.00	0.97	2%	-
Cs-137	2	0.86	0.86	0.87	0.85	2%	-
photon all	16	1.12	1.13	1.46	0.85	16%	-
All	18	1.12	1.13	1.46	0.85	15%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv


Reporting number 98 : Eye - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)		
beta	Sr-90	24	2.20	1.11	0.50	outlier	
photon	N-30	26	2.20	0.76	0.35	outlier	
		17	4.80	8.54	1.78	outlier	
	N-30/60°	18	4.80	6.19	1.29	OK	
		5	5.00	10.18	2.04	outlier	
	W-80-L	9	5.00	9.45	1.89	outlier	
		1	4.80	8.47	1.76	outlier	
		2	4.80	8.36	1.74	OK	
		3	4.80	5.43	1.13	OK	
	W-80-M	4	4.80	6.21	1.29	OK	
		8	54.0	84.77	1.57	OK	
	W-80-H	10	54.0	94.72	1.75	outlier	
		19	326	618.39	1.90	outlier	
	W-110	20	326	527.50	1.62	OK	
		14	4.70	4.69	1.00	OK	
	Cs-137	15	4.70	7.31	1.56	OK	
		11	5.50	6.63	1.21	OK	
		13	5.50	4.85	0.88	OK	
					Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated		

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.43	0.43	0.50	0.35	26%	-
N-30	2	1.53	1.53	1.78	1.29	23%	N-30: 60°/0° 1.28
N-30/60°	2	1.96	1.96	2.04	1.89	5%	
W-80-L	4	1.52	1.48	1.76	1.13	22%	W-80: H/L 1.12
W-80-M	2	1.66	1.66	1.75	1.57	8%	
W-80-H	2	1.76	1.76	1.90	1.62	11%	
W-110	2	1.28	1.28	1.56	1.00	31%	-
Cs-137	2	1.04	1.04	1.21	0.88	22%	-
photon all	16	1.59	1.53	2.04	0.88	23%	-
All	18	1.56	1.40	2.04	0.35	35%	-

outliers: 8 of 18

fraction of outliers: 44%



Results: IC2019_{ext eye}

1 values out of diagramme range (>2)!

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 99 : Eye - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	14 29	2.20 2.20	0.70 0.80	0.32 0.36	(outlier) (outlier)
photon	N-30	22 23	5.50 5.50	5.10 5.20	0.93 0.95	OK OK
	N-30/60°	20 21	4.80 4.80	3.40 3.80	0.71 0.79	OK OK
	W-80-L	4	5.10	5.40	1.06	OK
		7	5.10	5.40	1.06	OK
		16	5.10	5.40	1.06	OK
		18	5.10	5.30	1.04	OK
	W-80-M	1 2	48.0 48.0	51.00 51.90	1.06 1.08	OK OK
	W-80-H	17 19	331 331	355.50 355.80	1.07 1.07	OK OK
	W-110	8 9	4.90 4.90	4.80 4.90	0.98 1.00	OK OK
	Cs-137	10 11	5.00 5.00	3.90 4.10	0.78 0.82	OK OK
	NIR	3				
	NIR	5				
	NIR	6				
	NIR	12				
	NIR	13				
	NIR	15				
	NIR	24				
	NIR	25				
	NIR	26				
	NIR	27				
	NIR	28				
	NIR	30				
	NIR	31				
	NIR	32				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.34	0.34	0.36	0.32	9%	-
N-30	2	0.94	0.94	0.95	0.93	1%	N-30: 60°/0° 0.80
N-30/60°	2	0.75	0.75	0.79	0.71	8%	
W-80-L	4	1.06	1.05	1.06	1.04	1%	W-80: H/L 1.02
W-80-M	2	1.07	1.07	1.08	1.06	1%	
W-80-H	2	1.07	1.07	1.07	1.07	0%	
W-110	2	0.99	0.99	1.00	0.98	1%	-
Cs-137	2	0.80	0.80	0.82	0.78	4%	-
photon all	16	1.02	0.97	1.08	0.71	13%	-
All	18	0.99	0.90	1.08	0.32	26%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv


Reporting number 100 : Eye - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	6	2.20	1.79	0.81	OK
		19	2.20	1.86	0.85	OK
	N-30	23	4.40	4.71	1.07	OK
		27	4.40	5.19	1.18	OK
	N-30/60°	11	4.30	5.29	1.23	OK
		13	4.30	5.65	1.31	OK
	W-80-L	3	4.70	4.29	0.91	OK
		4	4.70	4.13	0.88	OK
		25	4.70	3.91	0.83	OK
		26	4.70	4.18	0.89	OK
	W-80-M	28	48.0	44.60	0.93	OK
		29	48.0	43.18	0.90	OK
	W-80-H	8	354	311.98	0.88	OK
		9	354	324.62	0.92	OK
	W-110	5	4.90	3.86	0.79	OK
		7	4.90	3.71	0.76	OK
	Cs-137	1	5.00	3.33	0.67	OK
		2	5.00	3.49	0.70	OK
photon	NIR	10				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
		12				
	NIR	14				
	NIR	15				
	NIR	16				
	NIR	17				
	NIR	18				
	NIR	20				
	NIR	21				
	NIR	22				
	NIR	24				
	NIR	30				
	NIR	31				
	NIR	32				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.83	0.83	0.85	0.81	3%	-
N-30	2	1.13	1.13	1.18	1.07	7%	N-30: 60°/0° 1.13
N-30/60°	2	1.27	1.27	1.31	1.23	5%	
W-80-L	4	0.88	0.88	0.91	0.83	4%	W-80: H/L 1.04
W-80-M	2	0.91	0.91	0.93	0.90	2%	
W-80-H	2	0.90	0.90	0.92	0.88	3%	
W-110	2	0.77	0.77	0.79	0.76	3%	-
Cs-137	2	0.68	0.68	0.70	0.67	3%	-
photon all	16	0.89	0.93	1.31	0.67	20%	-
All	18	0.89	0.92	1.31	0.67	19%	-

outliers: 0 of 18

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv


Reporting number 101: Eye - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Sr-90	1	2.20	5.02	2.28	outlier
		7	2.20	4.88	2.22	outlier
	N-30	2	4.80	4.23	0.88	OK
		4	4.80	4.44	0.92	OK
	N-30/60°	21	5.70	3.75	0.66	OK
		22	5.70	3.70	0.65	outlier
	W-80-L	14	4.50	4.17	0.93	OK
		16	4.50	6.60	1.47	OK
		24	4.50	4.37	0.97	OK
		28	4.50	6.06	1.35	OK
	W-80-M	10	52.0	75.85	1.46	OK
		13	52.0	64.50	1.24	OK
	W-80-H	8	384	347.35	0.90	OK
		9	384	376.35	0.98	OK
	W-110	17	4.30	4.16	0.97	OK
		18	4.30	6.39	1.49	OK
photon	Cs-137	25	4.40	3.13	0.71	OK
		26	4.40	3.04	0.69	OK
						Legend for Quality
						L low dose
						M medium dose
						H high dose
						NIR not irradiated
						WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	2.25	2.25	2.28	2.22	2%	-
N-30	2	0.90	0.90	0.92	0.88	3%	N-30: 60°/0° 0.72
N-30/60°	2	0.65	0.65	0.66	0.65	1%	
W-80-L	4	1.16	1.18	1.47	0.93	23%	W-80: H/L 1.15
W-80-M	2	1.35	1.35	1.46	1.24	11%	
W-80-H	2	0.94	0.94	0.98	0.90	6%	
W-110	2	1.23	1.23	1.49	0.97	30%	-
Cs-137	2	0.70	0.70	0.71	0.69	2%	-
photon all	16	0.95	1.02	1.49	0.65	29%	-
All	18	0.97	1.15	2.28	0.65	42%	-

outliers: 3 of 18

fraction of outliers: 17%



Results: IC2019_{ext eye}

2 values out of diagramme range (>2)!

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv


Reporting number 102 : Eye - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)		
beta	Sr-90	3	2.20	2.02	0.92	OK	
photon	N-30	27	2.20	2.11	0.96	OK	
		12	5.20	6.48	1.25	OK	
	N-30/60°	17	5.20	6.42	1.23	OK	
		1	5.30	6.33	1.19	OK	
	W-80-L	2	5.30	7.18	1.35	OK	
		14	4.50	4.43	0.98	OK	
		18	4.50	4.50	1.00	OK	
		29	4.50	4.47	0.99	OK	
	W-80-M	31	4.50	4.49	1.00	OK	
		19	48.0	48.30	1.01	OK	
	W-80-H	20	48.0	47.33	0.99	OK	
		22	319	325.67	1.02	OK	
	W-110	24	319	311.51	0.98	OK	
		11	5.30	4.76	0.90	OK	
	Cs-137	13	5.30	4.75	0.90	OK	
		15	5.50	4.22	0.77	OK	
		16	5.50	4.32	0.79	OK	
	NIR	4				Legend for Quality	
	NIR	5				L low dose	
	NIR	6				M medium dose	
	NIR	7				H high dose	
	NIR	8				NIR not irradiated	
	NIR	9				WIR wrongly irradiated	
	NIR	10					
	NIR	21					
	NIR	23					
	NIR	25					
	NIR	26					
	NIR	28					
	NIR	30					
	NIR	32					

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)	
Sr-90	2	0.94	0.94	0.96	0.92	3%	-	
N-30	2	1.24	1.24	1.25	1.23	1%	N-30: 60°/0°	
N-30/60°	2	1.27	1.27	1.35	1.19	9%	1.03	
W-80-L	4	1.00	0.99	1.00	0.98	1%	W-80: H/L	
W-80-M	2	1.00	1.00	1.01	0.99	1%	1.00	
W-80-H	2	1.00	1.00	1.02	0.98	3%		
W-110	2	0.90	0.90	0.90	0.90	0%	-	
Cs-137	2	0.78	0.78	0.79	0.77	2%	-	
photon all	16	1.00	1.02	1.35	0.77	16%	-	
All	18	0.99	1.01	1.35	0.77	15%	-	

outliers: 0 of 18

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv


Reporting number 103 : Eye - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	16 21	2.20 2.20	4.32 4.03	1.96 1.83	outlier outlier
photon	N-30	30 31	4.70 4.70	5.36 6.17	1.14 1.31	OK OK
	N-30/60°	27 29	5.20 5.20	7.46 6.85	1.43 1.32	OK OK
	W-80-L	2 7 15 17	5.50 5.50 5.50 5.50	5.19 5.49 6.27 5.45	0.94 1.00 1.14 0.99	OK OK OK OK
		1 4	54.0 54.0	54.81 55.66	1.02 1.03	OK OK
		9 11	333 333	341.52 363.28	1.03 1.09	OK OK
		5 6	4.40 4.40	4.03 3.66	0.92 0.83	OK OK
	Cs-137	8 10	4.90 4.90	3.70 3.54	0.76 0.72	OK OK
	NIR	3				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	12				
	NIR	13				
	NIR	14				
	NIR	18				
	NIR	19				
	NIR	20				
	NIR	22				
	NIR	23				
	NIR	24				
	NIR	25				
	NIR	26				
	NIR	28				
	NIR	32				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	1.90	1.90	1.96	1.83	5%	-
N-30	2	1.23	1.23	1.31	1.14	10%	N-30: 60°/0°
N-30/60°	2	1.38	1.38	1.43	1.32	6%	
W-80-L	4	0.99	1.02	1.14	0.94	8%	W-80: H/L
W-80-M	2	1.02	1.02	1.03	1.02	1%	
W-80-H	2	1.06	1.06	1.09	1.03	4%	
W-110	2	0.87	0.87	0.92	0.83	7%	-
Cs-137	2	0.74	0.74	0.76	0.72	3%	-
photon all	16	1.02	1.04	1.43	0.72	19%	-
All	18	1.03	1.14	1.96	0.72	29%	-

outliers: 2 of 18

fraction of outliers: 11%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 104 : Eye - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	8 23	2.20 2.20	2.43 2.33	1.10 1.06	(OK) (OK)
photon	N-30	10 12	4.70 4.70	4.74 4.72	1.01 1.00	OK OK
	N-30/60°	28 30	5.30 5.30	5.81 5.86	1.10 1.11	OK OK
	W-80-L	6 7 16 17	4.90 4.90 4.90 4.90	5.46 5.35 5.49 5.45	1.11 1.09 1.12 1.11	OK OK OK OK
		25 29	46.0 46.0	51.36 50.85	1.12 1.11	OK OK
		31 32	344 344	385.81 372.81	1.12 1.08	OK OK
		20 21	5.50 5.50	5.41 5.43	0.98 0.99	OK OK
	Cs-137	24 27	4.80 4.80	3.95 3.93	0.82 0.82	OK OK
	NIR	1				
	NIR	2				
	NIR	3				
	NIR	4				
	NIR	5				
	NIR	9				
	NIR	11				
	NIR	13				
	NIR	14				
	NIR	15				
	NIR	18				
	NIR	19				
	NIR	22				
	NIR	26				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	1.08	1.08	1.10	1.06	3%	-
N-30	2	1.01	1.01	1.01	1.00	0%	N-30: 60°/0°
N-30/60°	2	1.10	1.10	1.11	1.10	1%	1.09
W-80-L	4	1.11	1.11	1.12	1.09	1%	W-80: H/L
W-80-M	2	1.11	1.11	1.12	1.11	1%	1.00
W-80-H	2	1.10	1.10	1.12	1.08	2%	
W-110	2	0.99	0.99	0.99	0.98	0%	-
Cs-137	2	0.82	0.82	0.82	0.82	0%	-
photon all	16	1.09	1.04	1.12	0.82	10%	-
All	18	1.09	1.05	1.12	0.82	9%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 105 : Eye - photon dosimeter



	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	15 24	2.20 2.20	3.26 3.03	1.48 1.38	(OK) (OK)
photon	N-30 WIR					WIR WIR
	N-30/60°	5 6	5.20 5.20	4.69 4.43	0.90 0.85	OK OK
	W-80-L	3 7 25 26	5.30 5.30 5.30 5.30	3.92 3.59 3.94 4.02	0.74 0.68 0.74 0.76	OK OK OK OK
		14 16	46.0 46.0	35.64 33.67	0.77 0.73	OK OK
		11 13	342 342	252.98 257.49	0.74 0.75	OK OK
		30 31	5.00 5.00	3.28 3.43	0.66 0.69	OK OK
	Cs-137	18 19	4.80 4.80	2.67 2.76	0.56 0.58	outlier outlier
	NIR	1				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	2				
	NIR	4				
	NIR	10				
	NIR	12				
	NIR	17				
	NIR	20				
	NIR	21				
	NIR	22				
	NIR	23				
	NIR	27				
	NIR	28				
	NIR	29				
	NIR	32				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	1.43	1.43	1.48	1.38	5%	-
N-30 WIR	0						
N-30/60°	2	0.88	0.88	0.90	0.85	4%	
W-80-L	4	0.74	0.73	0.76	0.68	5%	W-80: H/L 1.03
W-80-M	2	0.75	0.75	0.77	0.73	4%	
W-80-H	2	0.75	0.75	0.75	0.74	1%	
W-110	2	0.67	0.67	0.69	0.66	3%	-
Cs-137	2	0.57	0.57	0.58	0.56	2%	-
photon all	14	0.74	0.72	0.90	0.56	13%	-
All	16	0.74	0.81	1.48	0.56	32%	-

outliers: 2 of 14

fraction of outliers: 14%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv


Reporting number 106 : Eye - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	12 31	2.20 2.20	1.29 1.28	0.59 0.58	OK OK
photon	N-30	1 3	5.10 5.10	3.85 4.00	0.75 0.78	OK OK
	N-30/60°	29 30	5.00 5.00	3.23 3.13	0.65 0.63	outlier outlier
	W-80-L	4 5 20 21	4.40 4.40 4.40 4.40	3.99 3.85 4.05 4.28	0.91 0.88 0.92 0.97	OK OK OK OK
		23 24	56.0 56.0	53.52 55.32	0.96 0.99	OK OK
		10 11	354 354	317.02 332.02	0.90 0.94	OK OK
		13 14	4.70 4.70	4.44 4.26	0.94 0.91	OK OK
	Cs-137	7 8	5.20 5.20	7.01 6.82	1.35 1.31	OK OK
	NIR	2				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	6				
	NIR	9				
	NIR	15				
	NIR	16				
	NIR	17				
	NIR	18				
	NIR	19				
	NIR	22				
	NIR	25				
	NIR	26				
	NIR	27				
	NIR	28				
	NIR	32				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.58	0.58	0.59	0.58	1%	-
N-30	2	0.77	0.77	0.78	0.75	3%	N-30: 60°/0°
N-30/60°	2	0.64	0.64	0.65	0.63	2%	
W-80-L	4	0.91	0.92	0.97	0.88	4%	W-80: H/L
W-80-M	2	0.97	0.97	0.99	0.96	2%	
W-80-H	2	0.92	0.92	0.94	0.90	3%	
W-110	2	0.93	0.93	0.94	0.91	3%	-
Cs-137	2	1.33	1.33	1.35	1.31	2%	-
photon all	16	0.91	0.92	1.35	0.63	21%	-
All	18	0.91	0.89	1.35	0.58	24%	-

outliers: 2 of 18

fraction of outliers: 11%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 107 : Eye - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)		
beta	Sr-90	22	2.20	7.01	3.19	(outlier)	
photon	N-30	24	2.20	6.67	3.03	(outlier)	
		4	5.30	5.38	1.02	OK	
	N-30/60°	6	5.30	5.47	1.03	OK	
		11	4.70	4.90	1.04	OK	
	W-80-L	12	4.70	5.14	1.09	OK	
		16	5.30	4.94	0.93	OK	
		18	5.30	5.44	1.03	OK	
		28	5.30	4.74	0.89	OK	
	W-80-M	31	5.30	4.58	0.86	OK	
		19	43.0	37.67	0.88	OK	
	W-80-H	21	43.0	39.12	0.91	OK	
		2	319	261.40	0.82	OK	
	W-110	3	319	234.89	0.74	OK	
		20	4.60	4.73	1.03	OK	
	Cs-137	23	4.60	4.20	0.91	OK	
		13	5.70	7.52	1.32	OK	
		14	5.70	7.38	1.29	OK	
					Legend for Quality		
					L low dose		
					M medium dose		
					H high dose		
					NIR not irradiated		
					WIR wrongly irradiated		

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	3.11	3.11	3.19	3.03	4%	-
N-30	2	1.02	1.02	1.03	1.02	1%	N-30: 60°/0°
N-30/60°	2	1.07	1.07	1.09	1.04	3%	1.04
W-80-L	4	0.91	0.93	1.03	0.86	8%	W-80: H/L
W-80-M	2	0.89	0.89	0.91	0.88	3%	0.96
W-80-H	2	0.78	0.78	0.82	0.74	8%	
W-110	2	0.97	0.97	1.03	0.91	8%	-
Cs-137	2	1.31	1.31	1.32	1.29	1%	-
photon all	16	0.97	0.99	1.32	0.74	16%	-
All	18	1.02	1.22	3.19	0.74	57%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

2 values out of diagramme range (>2)!

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 108 : Eye - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	10 31	2.20 2.20	5.48 5.62	2.49 2.55	(outlier) (outlier)
photon	N-30	22	4.90	5.38	1.10	OK
		32	4.90	5.12	1.05	OK
	N-30/60°	16	5.70	6.81	1.19	OK
		17	5.70	6.81	1.19	OK
	W-80-L	23	4.40	4.38	1.00	OK
		24	4.40	4.36	0.99	OK
		28	4.40	4.31	0.98	OK
		30	4.40	4.11	0.93	OK
	W-80-M	26	50.0	47.21	0.94	OK
		27	50.0	50.56	1.01	OK
	W-80-H	1	335	315.37	0.94	OK
		2	335	322.79	0.96	OK
	W-110	20	4.60	4.44	0.97	OK
		21	4.60	4.47	0.97	OK
	Cs-137	6 7	5.60 5.60	7.19 7.09	1.28 1.27	OK OK
	NIR	3				
	NIR	4				
	NIR	5				
	NIR	8				
	NIR	9				
	NIR	11				
	NIR	12				
	NIR	13				
	NIR	14				
	NIR	15				
	NIR	18				
	NIR	19				
	NIR	25				
	NIR	29				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	2.52	2.52	2.55	2.49	2%	-
N-30	2	1.07	1.07	1.10	1.05	4%	N-30: 60°/0° 1.11
N-30/60°	2	1.19	1.19	1.19	1.19	0%	
W-80-L	4	0.98	0.97	1.00	0.93	3%	W-80: H/L 1.00
W-80-M	2	0.98	0.98	1.01	0.94	5%	
W-80-H	2	0.95	0.95	0.96	0.94	2%	
W-110	2	0.97	0.97	0.97	0.97	0%	-
Cs-137	2	1.28	1.28	1.28	1.27	1%	-
photon all	16	0.99	1.05	1.28	0.93	11%	-
All	18	1.00	1.21	2.55	0.93	40%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

2 values out of diagramme range (>2)!

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 109 : Eye - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	16 24	2.20 2.20	4.45 4.47	2.02 2.03	(outlier) (outlier)
photon	N-30	23 28	4.90 4.90	4.17 4.31	0.85 0.88	OK OK
	N-30/60°	15 17	5.70 5.70	5.11 5.21	0.90 0.91	OK OK
	W-80-L	5 9 21 22	5.50 5.50 5.50 5.50	4.21 4.13 4.27 4.38	0.76 0.75 0.78 0.80	OK OK OK OK
		18 20	50.0 50.0	38.20 38.00	0.76 0.76	OK OK
		3 4	335 335	253.70 241.70	0.76 0.72	OK OK
		1 2	4.60 4.60	3.44 3.46	0.75 0.75	OK OK
	Cs-137	30 32	5.00 5.00	5.11 5.02	1.02 1.00	OK OK
	NIR	6				
	NIR	7				
	NIR	8				
	NIR	10				
	NIR	11				
	NIR	12				
	NIR	13				
	NIR	14				
	NIR	19				
	NIR	25				
	NIR	26				
	NIR	27				
	NIR	29				
	NIR	31				

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	2.03	2.03	2.03	2.02	0%	-
N-30	2	0.86	0.86	0.88	0.85	2%	N-30: 60°/0°
N-30/60°	2	0.90	0.90	0.91	0.90	1%	1.05
W-80-L	4	0.77	0.77	0.80	0.75	2%	W-80: H/L
W-80-M	2	0.76	0.76	0.76	0.76	0%	0.99
W-80-H	2	0.74	0.74	0.76	0.72	3%	
W-110	2	0.75	0.75	0.75	0.75	0%	-
Cs-137	2	1.01	1.01	1.02	1.00	1%	-
photon all	16	0.77	0.82	1.02	0.72	11%	-
All	18	0.79	0.96	2.03	0.72	42%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

2 values out of diagramme range (>2)!

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 110 : Eye - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)		
beta	Sr-90	24	2.20	3.10	1.41	(OK)	
photon	N-30	25	2.20	3.10	1.41	(OK)	
		17	4.50	2.50	0.56	outlier	
	N-30/60°	18	4.50	2.50	0.56	outlier	
		3	4.80	3.20	0.67	OK	
	W-80-L	4	4.80	3.10	0.65	outlier	
		7	4.80	10.40	2.17	outlier	
		8	4.80	9.60	2.00	outlier	
		31	4.80	10.90	2.27	outlier	
	W-80-M	32	4.80	9.90	2.06	outlier	
		15	55.0	103.00	1.87	outlier	
	W-80-H	16	55.0	106.00	1.93	outlier	
		12	319	543.00	1.70	outlier	
	W-110	13	319	602.00	1.89	outlier	
		10	4.50	7.70	1.71	OK	
	Cs-137	11	4.50	8.00	1.78	outlier	
		5	5.40	3.10	0.57	outlier	
		6	5.40	2.80	0.52	outlier	
	NIR	1				Legend for Quality	
	NIR	2				L low dose	
	NIR	9				M medium dose	
	NIR	14				H high dose	
	NIR	19				NIR not irradiated	
	NIR	20				WIR wrongly irradiated	
	NIR	21					
	NIR	22					
	NIR	23					
	NIR	26					
	NIR	27					
	NIR	28					
	NIR	29					
	NIR	30					

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)	
Sr-90	2	1.41	1.41	1.41	1.41	0%	-	
N-30	2	0.56	0.56	0.56	0.56	0%	N-30: 60°/0°	
N-30/60°	2	0.66	0.66	0.67	0.65	2%	1.18	
W-80-L	4	2.11	2.13	2.27	2.00	6%	W-80: H/L	
W-80-M	2	1.90	1.90	1.93	1.87	2%	0.89	
W-80-H	2	1.79	1.79	1.89	1.70	7%		
W-110	2	1.74	1.74	1.78	1.71	3%	-	
Cs-137	2	0.55	0.55	0.57	0.52	7%	-	
photon all	16	1.74	1.43	2.27	0.52	48%	-	
All	18	1.71	1.43	2.27	0.52	46%	-	

outliers: 14 of 16

fraction of outliers: 88%



Results: IC2019_{ext eye}

3 values out of diagramme range (>2)!

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv


Reporting number 111 : Eye - photon/beta dosimeter

	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)	
beta	Sr-90	11 15	2.20 2.20	6.86 6.65	3.12 3.02	outlier outlier
photon	N-30	31 32	5.10 5.10	3.49 3.52	0.68 0.69	OK OK
	N-30/60°	3 6	5.00 5.00	3.65 3.57	0.73 0.71	OK OK
	W-80-L	4 5 8 9	4.30 4.30 4.30 4.30	4.57 4.45 4.40 4.44	1.06 1.03 1.02 1.03	OK OK OK OK
		W-80-M	1 2	43.0 43.0	41.88 45.29	0.97 1.05
		W-80-H	21 25	381 381	61.03 60.74	0.16 0.16
		W-110	12 13	5.70 5.70	7.20 6.54	1.26 1.15
	Cs-137	22 23	4.90 4.90	5.48 5.25	1.12 1.07	OK OK
	NIR	7				Legend for Quality L low dose M medium dose H high dose NIR not irradiated WIR wrongly irradiated
	NIR	10				
	NIR	14				
	NIR	16				
	NIR	17				
	NIR	18				
	NIR	19				
	NIR	20				
	NIR	24				
	NIR	26				
	NIR	27				
	NIR	28				
	NIR	29				
	NIR	30				

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	3.07	3.07	3.12	3.02	2%	-
N-30	2	0.69	0.69	0.69	0.68	1%	N-30: 60°/0°
N-30/60°	2	0.72	0.72	0.73	0.71	2%	
W-80-L	4	1.03	1.04	1.06	1.02	2%	W-80: H/L
W-80-M	2	1.01	1.01	1.05	0.97	6%	
W-80-H	2	0.16	0.16	0.16	0.16	0%	
W-110	2	1.21	1.21	1.26	1.15	7%	-
Cs-137	2	1.09	1.09	1.12	1.07	3%	-
photon all	16	1.03	0.87	1.26	0.16	38%	-
All	18	1.03	1.11	3.12	0.16	70%	-

outliers: 4 of 18

fraction of outliers: 22%



Results: IC2019_{ext eye}

2 values out of diagramme range (>2)!

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 112 : Eye - photon/beta dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results	
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(0.07)$ mSv	response R (reported/true)	
beta	Sr-90	5	2.20	1.32	0.60	OK
		6	2.20	1.35	0.61	OK
	N-30	23	5.30	5.78	1.09	OK
		24	5.30	6.00	1.13	OK
	N-30/60°	1	4.40	5.22	1.19	OK
		2	4.40	5.06	1.15	OK
	W-80-L	20	5.80	5.63	0.97	OK
		22	5.80	6.36	1.10	OK
		31	5.80	6.02	1.04	OK
		32	5.80	5.67	0.98	OK
	W-80-M	26	48.0	51.13	1.07	OK
		28	48.0	51.44	1.07	OK
photon	W-80-H	16	350	327.97	0.94	OK
		18	350	340.67	0.97	OK
	W-110	12	5.30	5.54	1.05	OK
		13	5.30	5.70	1.08	OK
	Cs-137	14	5.60	5.27	0.94	OK
		15	5.60	5.49	0.98	OK
	NIR	3				
	NIR	4				
	NIR	7				
	NIR	8				
	NIR	9				
	NIR	10				
	NIR	11				
	NIR	17				
	NIR	19				
	NIR	21				
	NIR	25				
	NIR	27				
	NIR	29				
	NIR	30				

Legend for Quality
 L low dose
 M medium dose
 H high dose
 NIR not irradiated
 WIR wrongly irradiated

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.61	0.61	0.61	0.60	2%	-
N-30	2	1.11	1.11	1.13	1.09	3%	
N-30/60°	2	1.17	1.17	1.19	1.15	2%	1.05
W-80-L	4	1.01	1.02	1.10	0.97	6%	
W-80-M	2	1.07	1.07	1.07	1.07	0%	1.05
W-80-H	2	0.96	0.96	0.97	0.94	3%	
W-110	2	1.06	1.06	1.08	1.05	2%	-
Cs-137	2	0.96	0.96	0.98	0.94	3%	-
photon all	16	1.06	1.05	1.19	0.94	7%	-
All	18	1.04	1.00	1.19	0.60	16%	-

outliers: 0 of 18

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv

Reporting number 113 : Eye - photon dosimeter


	true values reported by the irradiating laboratory			values reported by participant	results		
radiation type	radiation quality	dosemeter number	dose $H_p(3)$ mSv	dose $H_p(3)$ mSv	response R (reported/true)		
beta	Sr-90	9	2.20	1.42	0.65	(OK)	
photon	N-30	18	2.20	1.41	0.64	(OK)	
		21	4.70	4.83	1.03	OK	
	N-30/60°	22	4.70	4.89	1.04	OK	
		5	5.10	4.48	0.88	OK	
	W-80-L	6	5.10	4.34	0.85	OK	
		14	5.20	6.01	1.16	OK	
		15	5.20	6.28	1.21	OK	
		26	5.20	6.08	1.17	OK	
	W-80-M	28	5.20	6.15	1.18	OK	
		10	46.0	49.87	1.08	OK	
	W-80-H	11	46.0	49.41	1.07	OK	
		17	344	347.59	1.01	OK	
	W-110	19	344	367.25	1.07	OK	
		24	5.00	5.57	1.11	OK	
	Cs-137	25	5.00	5.60	1.12	OK	
		12	4.40	4.90	1.11	OK	
		13	4.40	4.92	1.12	OK	
						Legend for Quality	
						L low dose	
						M medium dose	
						H high dose	
						NIR not irradiated	
						WIR wrongly irradiated	

beta results (xx) are greyed and were not considered

radiation quality	number of values	median (R)	mean (R)	maximum (R)	minimum (R)	coefficient of variation (R)	mean response ratios (R/R)
Sr-90	2	0.64	0.64	0.65	0.64	0%	-
N-30	2	1.03	1.03	1.04	1.03	1%	N-30: 60°/0°
N-30/60°	2	0.86	0.86	0.88	0.85	2%	0.84
W-80-L	4	1.18	1.18	1.21	1.16	2%	W-80: H/L
W-80-M	2	1.08	1.08	1.08	1.07	1%	0.92
W-80-H	2	1.04	1.04	1.07	1.01	4%	
W-110	2	1.12	1.12	1.12	1.11	0%	-
Cs-137	2	1.12	1.12	1.12	1.11	0%	-
photon all	16	1.10	1.08	1.21	0.85	9%	-
All	18	1.08	1.03	1.21	0.64	16%	-

outliers: 0 of 16

fraction of outliers: 0%



Results: IC2019_{ext eye}

ISO 14146:2018 trumpet curve, lower dose limit (H_0): 0.3 mSv